

HSV1 IgG

**Enzyme ImmunoAssay (ELISA) for the
quantitative/qualitative determination
of IgG antibodies to
Herpes Simplex Virus type 1
in human serum and plasma**

- for “in vitro” diagnostic use only -



DIA.PRO

**Diagnostic Bioprobes Srl
Via G. Carducci n° 27
20099 Sesto San Giovanni
(Milano) - Italy**

Phone +39 02 27007161

Fax +39 02 44386771

e-mail: info@diapro.it

HSV1 IgG

A. INTENDED USE

Enzyme ImmunoAssay (ELISA) for the quantitative/qualitative determination of IgG antibodies to Herpes Simplex Virus type 1 in human plasma and sera.
For "in vitro" diagnostic use only.

B. INTRODUCTION

Herpes Simplex Virus type 1 (HSV1) and type 2 (HSV2) are large complex DNA-containing viruses which have been shown to induce the synthesis of several proteins during infection, possessing an high number of crossreactive determinants and just a few of type-specific sequences.

The majority of primary and recurrent genital herpetic infections are caused by HSV2; while non genital infections, such as common cold sores, are caused primarily by HSV1.

The detection of virus specific IgG and IgM antibodies are important in the diagnosis of acute/primary virus infections or reactivations of a latent one, in the absence of evident clinical symptoms.

Asymptomatic infections may happen for HSV in apparently healthy individuals and during pregnancy. Severe herpetic infections may happen in immunocompromised and suppressed patients in which the disease may evolve toward critical pathologies.

The determination of HSV specific antibodies has then become important in the monitoring of "risk" patients and in the follow up of acute and severe infections.

C. PRINCIPLE OF THE TEST

Microplates are coated with native inactivated HSV1.

The solid phase is first treated with the diluted sample and IgG to HSV are captured, if present, by the antigens.

After washing out all the other components of the sample, in the 2nd incubation bound anti HSV1 IgG are detected by the addition of polyclonal specific anti hIgG antibodies, labelled with peroxidase (HRP).

The enzyme captured on the solid phase, acting on the substrate/chromogen mixture, generates an optical signal that is proportional to the amount of anti HSV1 IgG antibodies present in the sample. A Calibration Curve, calibrated against an internal Gold Standard, makes possible a quantitative determination of the IgG antibody in the patient.

D. COMPONENTS

Each kit contains sufficient reagents to perform 96 tests.

1. Microplate: MICROPLATE

n° 1. 12 strips x 8 microwells coated with native UV inactivated HSV1 in presence of bovine proteins.

Plates are sealed into a bag with desiccant. Allow the microplate to reach room temperature before opening; reseal unused strips in the bag with desiccant and store at 2..8°C.

2. Calibration Curve: CAL N° ...

Ready to use and color coded standard curve derived from human plasma positive for HSV1 IgG ranging:

4ml CAL1 = 0 arbU/ml
4ml CAL2 = 5 arbU/ml
2ml CAL3 = 10 arbU/ml
2ml CAL4 = 20 arbU/ml
2ml CAL5 = 50 arbU/ml
4ml CAL6 = 100 arbU/ml.

Standards are calibrated in arbitrary units against an internal Gold Standard (or IGS).

It contains human serum proteins, 2% casein, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.1% Tween 20, 0.09% Na-azide and 0.045% ProClin 300 as preservatives. Standards are blue colored.

3. Control Serum: CONTROL ...ml

1 vial. Lyophilized. It contains fetal bovine serum proteins, human IgG antibodies to HSV1 at about 20 arbU/ml ±20%, 0.2 mg/ml gentamicine sulphate and 0.045% ProClin 300 as preservatives.

Note: The volume necessary to dissolve the content of the vial may vary from lot to lot. Please use the right volume reported on the label.

4. Wash buffer concentrate: WASHBUF 20X

1x60ml/bottle20x concentrated solution. Once diluted, the wash solution contains 10 mM phosphate buffer pH 7.0+/-0.2, 0.05% Tween 20 and 0.045% ProClin 300.

5. Enzyme conjugate : CONJ

2x8ml/vial. Ready to use and red colour coded. It contains Horseradish peroxidase conjugated polyclonal antibodies to human IgG, 5% BSA, 10 mM Tris buffer pH 6.8+/-0.1, 0.045% ProClin 300, 0.02% gentamicine sulphate as preservatives and 0.01% red alimentary dye.

6. Chromogen/Substrate: SUBS TMB

1x16ml/vial. It contains 50 mM citrate-phosphate buffer pH 3.5-3.8, 4% dimethylsulphoxide, 0.03% tetra-methyl-benzidine (or TMB) and 0.02% hydrogen peroxide (or H₂O₂).

Note: To be stored protected from light as sensitive to strong illumination.

7. Sulphuric Acid: H₂SO₄ 0.3 M

1x15ml/vial. It contains 0.3 M H₂SO₄ solution.

Attention: Irritant (H315, H319; P280, P302+P352, 332+P313, P305+ P351+P338, P337+P313, P362+P363)

8. Specimen Diluent: DILSPE

2x60ml/vial. It contains 2% casein, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.1% Tween 20, 0.09% Na-azide 0.1% and 0.045% ProClin 300 as preservatives. The reagent is blue colour coded.

9. Plate sealing foils n°2

10. Package insert n°1

E. MATERIALS REQUIRED BUT NOT PROVIDED

1. Calibrated Micropipettes (1000 ul, 100 ul and 10 ul) and disposable plastic tips.
2. EIA grade water (double distilled or deionised, charcoal treated to remove oxidizing chemicals used as disinfectants).
3. Timer with 60 minute range or higher.
4. Absorbent paper tissues.
5. Calibrated ELISA microplate thermostatic incubator (dry or wet), set at +37°C (+/-0.5°C tolerance)..
6. Calibrated ELISA microwell reader with 450nm (reading) and with 620-630nm (blanking) filters.
7. Calibrated ELISA microplate washer.
8. Vortex or similar mixing tools.

F. WARNINGS AND PRECAUTIONS

1. The kit has to be used by skilled and properly trained technical personnel only, under the supervision of a medical doctor responsible of the laboratory.
2. All the personnel involved in performing the assay have to wear protective laboratory clothes, talc-free gloves and glasses. The use of any sharp (needles) or cutting (blades) devices should be avoided. All the personnel involved should be trained in biosafety procedures, as recommended by the Center for Disease Control, Atlanta, U.S. and reported in the National Institute of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
3. All the personnel involved in sample handling should be vaccinated for HBV and HAV, for which vaccines are available, safe and effective.
4. The laboratory environment should be controlled so as to avoid contaminants such as dust or air-born microbial agents, when opening kit vials and microplates and when performing the test. Protect the Chromogen (TMB) from strong light and avoid vibration of the bench surface where the test is undertaken.
5. Upon receipt, store the kit at 2..8°C into a temperature controlled refrigerator or cold room.
6. Do not interchange components between different lots of the kits. It is recommended that components between two kits of the same lot should not be interchanged.
7. Check that the reagents are clear and do not contain visible heavy particles or aggregates. If not, advise the laboratory supervisor to initiate the necessary procedures for kit replacement.
8. Avoid cross-contamination between serum/plasma samples by using disposable tips and changing them after each sample. Do not reuse disposable tips.
9. Avoid cross-contamination between kit reagents by using disposable tips and changing them between the use of each one. Do not reuse disposable tips.
10. Do not use the kit after the expiration date stated on the external container and internal (vials) labels. A study conducted on an opened kit did not pointed out any relevant loss of activity up to six 6 uses of the device and up to 3 months.
11. Treat all specimens as potentially infective. All human serum specimens should be handled at Biosafety Level 2, as recommended by the Center for Disease Control, Atlanta, U.S. in compliance with what reported in the Institutes of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
12. The use of disposable plastic-ware is recommended in the preparation of the liquid components or in transferring components into automated workstations, in order to avoid cross contamination.
13. Waste produced during the use of the kit has to be discarded in compliance with national directives and laws concerning laboratory waste of chemical and biological substances. In particular, liquid waste generated from the washing procedure, from residuals of controls and from samples has to be treated as potentially infective material and inactivated before waste. Suggested procedures of inactivation are treatment with a 10% final concentration of household bleach for 16-18 hrs or heat inactivation by autoclave at 121°C for 20 min..
14. Accidental spills from samples and operations have to be adsorbed with paper tissues soaked with household bleach and then with water. Tissues should then be discarded in proper containers designated for laboratory/hospital waste.
15. The Sulphuric Acid is an irritant. In case of spills, wash the surface with plenty of water
16. Other waste materials generated from the use of the kit (example: tips used for samples and controls, used microplates) should be handled as potentially infective and disposed according to national directives and laws concerning laboratory wastes.

G. SPECIMEN: PREPARATION AND WARNINGS

1. Blood is drawn aseptically by venepuncture and plasma or serum is prepared using standard techniques of preparation of samples for clinical laboratory analysis. No influence has been observed in the preparation of the sample with citrate, EDTA and heparin.
2. Samples have to be clearly identified with codes or names in order to avoid misinterpretation of results. Bar code labeling and electronic reading is strongly recommended.
3. Haemolysed ("red") and visibly hyperlipemic ("milky") samples have to be discarded as they could generate false results. Samples containing residues of fibrin or heavy particles or microbial filaments and bodies should be discarded as they could give rise to false results.
4. Sera and plasma can be stored at +2°...+8°C in primary collection tubes for up to five days after collection. Do not freeze primary tubes of collection. For longer storage periods, sera and plasma samples, carefully removed from the primary collection tube, can be stored frozen at -20°C for several months. Any frozen samples should not be frozen/thawed more than once as this may generate particles that could affect the test result.
5. If particles are present, centrifuge at 2.000 rpm for 20 min or filter using 0.2-0.8u filters to clean up the sample for testing.

H. PREPARATION OF COMPONENTS AND WARNINGS

A study conducted on an opened kit has not pointed out any relevant loss of activity up to 6 re-uses of the device and up to 3 months.

Microplate:

Allow the microplate to reach room temperature (about 1 hr) before opening the container. Check that the desiccant has not turned dark green, indicating a defect in storing. In this case, call Dia.Pro's customer service. Unused strips have to be placed back into the aluminum pouch, with the desiccant supplied, firmly zipped and stored at +2°..8°C. After first opening, remaining strips are stable until the humidity indicator inside the desiccant bag turns from yellow to green.

Calibration Curve

Ready to use component. Mix carefully on vortex before use.

Control Serum

Add the volume of ELISA grade water, reported on the label, to the lyophilised powder; let fully dissolve and then gently mix on vortex.

Note: The control after dissolution is not stable. Store frozen in aliquots at -20°C.

Wash buffer concentrate:

The whole content of the concentrated solution has to be diluted 20x with bidistilled water and mixed gently end-over-end before use. During preparation avoid foaming as the presence of bubbles could impact on the efficiency of the washing cycles.

Note: Once diluted, the wash solution is stable for 1 week at +2..8° C.

Enzyme conjugate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

If this component has to be transferred use only plastic, possibly sterile disposable containers.

Chromogen/Substrate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

Do not expose to strong illumination, oxidizing agents and metallic surfaces.

If this component has to be transferred use only plastic, possible sterile disposable container

Sample Diluent

Ready to use component. Mix carefully on vortex before use.

Sulphuric Acid:

Ready to use. Mix well on vortex before use.

Attention: Irritant (H315, H319; P280, P302+P352, 332+P313, P305+ P351+P338, P337+P313, P362+P363).

Legenda:

Warning H statements:

H315 – Causes skin irritation.

H319 – Causes serious eye irritation.

Precautionary P statements:

P280 – Wear protective gloves/protective clothing/eye protection/face protection.

P302 + P352 – IF ON SKIN: Wash with plenty of soap and water.

P332 + P313 – If skin irritation occurs: Get medical advice/attention.

P305 + P351 + P338 – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P337 + P313 – If eye irritation persists: Get medical advice/attention.

P362 + P363 - Take off contaminated clothing and wash it before reuse.

I. INSTRUMENTS AND TOOLS USED IN COMBINATION WITH THE KIT

1. Micropipettes have to be calibrated to deliver the correct volume required by the assay and must be submitted to regular decontamination (household alcohol, 10% solution of bleach, hospital grade disinfectants) of those parts that could accidentally come in contact with the sample. They should also be regularly maintained in order to show a precision of 1% and a trueness of +/-2%. Decontamination of spills or residues of kit components should also be carried out regularly.
2. The ELISA incubator has to be set at +37°C (tolerance of +/- 0.5°C) and regularly checked to ensure the correct temperature is maintained. Both dry incubators and water baths are suitable for the incubations, provided that the instrument is validated for the incubation of ELISA tests.
3. The **ELISA washer** is extremely important to the overall performances of the assay. The washer must be carefully validated in advance, checked for the delivery of the right dispensation volume and regularly submitted to maintenance according to the manufacturer's instructions for use. In particular the washer, at the end of the daily workload, has to be extensively cleaned out of salts with deionized water. Before use, the washer has to be extensively primed with the diluted Washing Solution. The instrument weekly has to be submitted to decontamination according to its manual (NaOH 0.1 M decontamination suggested).
5 washing cycles (aspiration + dispensation of 350ul/well of washing solution + 20 sec soaking = 1 cycle) are sufficient to ensure the assay with the declared performances. If soaking is not possible add one more cycle of washing.
An incorrect washing cycle or salt-blocked needles are the major cause of false positive reactions.
4. Incubation times have a tolerance of $\pm 5\%$.
5. The ELISA microplate reader has to be equipped with a reading filter of 450nm and with a second filter of 620-630nm, mandatory for blanking purposes. Its standard performances should be (a) bandwidth ≤ 10 nm; (b) absorbance range from 0 to ≥ 2.0 ; (c) linearity to ≥ 2.0 ; repeatability $\geq 1\%$. Blanking is carried out on the well identified in the section "Assay Procedure". The optical system of the reader has to be calibrated regularly to ensure that the correct optical density is measured. It should be regularly maintained according to the manufacturer's instructions.

6. When using an ELISA automated work station, all critical steps (dispensation, incubation, washing, reading, data handling) have to be carefully set, calibrated, controlled and regularly serviced in order to match the values reported in the section "Internal Quality Control". The assay protocol has to be installed in the operating system of the unit and validated as for the washer and the reader. In addition, the liquid handling part of the station (dispensation and washing) has to be validated and correctly set. Particular attention must be paid to avoid carry over by the needles used for dispensing and for washing. This must be studied and controlled to minimize the possibility of contamination of adjacent wells. The use of ELISA automated work stations is recommended when the number of samples to be tested exceed 20-30 units per run.
7. Dia.Pro's customer service offers support to the user in the setting and checking of instruments used in combination with the kit, in order to assure compliance with the requirements described. Support is also provided for the installation of new instruments to be used with the kit.

L. PRE ASSAY CONTROLS AND OPERATIONS

1. Check the expiration date of the kit printed on the external label (primary container). Do not use if expired.
2. Check that the liquid components are not contaminated by visible particles or aggregates.
3. Check that the Chromogen (TMB) is colourless or pale blue by aspirating a small volume of it with a sterile plastic pipette.
4. Check that no breakage occurred in transportation and no spillage of liquid is present inside the box (primary container). Check that the aluminium pouch, containing the microplate, is not punctured or damaged.
5. Dissolve the content of the lyophilised Control Serum as reported in the proper section.
6. Dilute all the content of the 20x concentrated Wash Solution as described above.
7. Allow all the other components to reach room temperature (about 1 hr) and then mix gently on vortex all liquid reagents.
8. Set the ELISA incubator at +37°C and prepare the ELISA washer by priming with the diluted washing solution, according to the manufacturers instructions. Set the right number of washing cycles as reported in the specific section.
9. Check that the ELISA reader is turned on or ensure it will be turned on at least 20 minutes before reading.
10. If using an automated work station, turn on, check settings and be sure to use the right assay protocol.
11. Check that the micropipettes are set to the required volume.
12. Check that all the other equipment is available and ready to use.
13. In case of problems, do not proceed further with the test and advise the supervisor.

M. ASSAY PROCEDURE

The assay has to be carried out according to what reported below, taking care to maintain the same incubation time for all the samples in testing.

The kit may be used for quantitative and qualitative determinations as well.

M1. QUANTITATIVE DETERMINATION:

Automated assay:

In case the test is carried out automatically with an ELISA system, we suggest to make the instrument aspirate 1000 μ l Sample Diluent and then 10 μ l sample (1:101 dilution factor). The whole content is then dispensed into a properly defined dilution tube. Before the next sample is aspirated, needles have

to be duly washed to avoid any cross-contamination among samples. When all the samples have been diluted make the instrument dispense 100 µl samples into the proper wells of the microplate.

This procedure may be carried out also in two steps of dilutions of 1:10 each (90 µl Sample Diluent + 10 µl sample) into a second dilution platform. Make then the instrument aspirate first 100 µl Sample Diluent, then 10 µl liquid from the first dilution in the platform and finally dispense the whole content in the proper well of the assay microplate.

Do not dilute Calibrators and the dissolved Control Serum as they are ready to use.

Dispense 100 µl calibrators/control in the appropriate calibration/control wells.

For the next operations follow the operative instructions reported below for the Manual Assay.

It is strongly recommended to check that the time lap between the dispensation of the first and the last sample will be calculated by the instrument and taken into consideration by delaying the first washing operation accordingly.

Manual assay:

1. Dilute samples 1:101 into a properly defined dilution tube (example: 1000 µl Sample Diluent + 10 µl sample). Do not dilute the Calibration Set as calibrators are ready to use. Mix carefully all the liquid components on vortex and then proceed as described below.
2. Place the required number of microwells in the microwell holder. Leave the A1 and B1 empty for the operation of blanking.
3. Dispense 100 µl of Calibrators and 100 µl Control Serum in duplicate. Then dispense 100 µl of diluted samples in each properly identified well.
4. Incubate the microplate for **60 min at +37°C**.

Important note: Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

5. Wash the microplate with an automatic washer as reported previously (section I.3).
6. Pipette 100 µl Enzyme Conjugate into each well, except A1+B1 blanking wells, and cover with the sealer. Check that this red coloured component has been dispensed in all the wells, except A1 and B1.

Important note: Be careful not to touch the plastic inner surface of the well with the tip filled with the Enzyme Conjugate. Contamination might occur.

7. Incubate the microplate for **60 min at +37°C**.
8. Wash microwells as in step 5.
9. Pipette 100 µl Chromogen/Substrate mixture into each well, the blank wells A1 and B1 included. Then incubate the microplate at **room temperature (18-24°C) for 20 minutes**.

Important note: Do not expose to strong direct illumination. High background might be generated.

10. Pipette 100 µl Sulphuric Acid to stop the enzymatic reaction into all the wells using the same pipetting sequence as in step 9. Addition of acid will turn the positive calibrators, the control serum and the positive samples from blue to yellow.
11. Measure the colour intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and at 620-630nm (background subtraction, mandatory), blanking the instrument on A1 or B1 or both.

M2. QUALITATIVE DETERMINATION

If only a qualitative determination is required, proceed as described below:

Automated assay:

Proceed as described in section M1.

Manual assay:

1. Dilute samples 1:101 into a properly defined dilution tube (example: 1000 µl Sample Diluent + 10 µl sample). Do not dilute the Calibration Set as calibrators are ready to use. Mix carefully all the liquid components on vortex and then proceed as described below.
2. Place the required number of Microwells in the microwell holder. Leave A1 well empty for the operation of blanking.
3. Dispense 100 µl of Calibrator 0 arbU/ml and Calibrator 5 arbU/ml in duplicate and Calibrator 100 arbU/ml in single. Then dispense 100 µl of diluted samples in each properly identified well.
4. Incubate the microplate for **60 min at +37°C**.

Important note: Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

5. Wash the microplate with an automatic washer as reported previously (section I.3).
6. Pipette 100 µl Enzyme Conjugate into each well, except the A1 well, and cover with the sealer. Check that this red coloured component has been dispensed in all the wells, except A1.

Important note: Be careful not to touch the plastic inner surface of the well with the tip filled with the Enzyme Conjugate. Contamination might occur.

7. Incubate the microplate for **60 min at +37°C**.
8. Wash microwells as in step 5.
9. Pipette 100 µl Chromogen/Substrate mixture into each well, the blank well included. Then incubate the microplate at **room temperature (18-24°C) for 20 minutes**.

Important note: Do not expose to strong direct illumination. High background might be generated.

10. Pipette 100 µl Sulphuric Acid into all the wells using the same pipetting sequence as in step 9. Addition of acid will turn the positive calibrators, the control serum and the positive samples from blue to yellow.
11. Measure the colour intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and at 620-630nm (background subtraction, mandatory), blanking the instrument on A1.

General Important notes:

1. Ensure that no finger prints are present on the bottom of the microwell before reading. Finger prints could generate false positive results on reading.
2. Reading has to be carried out just after the addition of the Stop Solution and anyway not any longer than 20 minutes after its addition. Some self oxidation of the chromogen can occur leading to high background.

N. ASSAY SCHEME

Method	Operations
Calibrators & Control (*)	100 µl
Samples diluted 1:101	100 µl
1 st incubation	60 min
Temperature	+37°C
Wash step	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
Enzyme conjugate	100 µl
2 nd incubation	60 min
Temperature	+37°C
Wash step	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
TMB/H ₂ O ₂	100 µl
3 rd incubation	20 min
Temperature	r.t.
Sulphuric Acid	100 µl
Reading OD	450nm / 620-630nm

(*) Important Notes:

- The Control Serum (CS) it does not affect the test's results calculation.
- The Control Serum (CS) used only if a laboratory internal quality control is required by the Management.

An example of dispensation scheme for Quantitative Analysis is reported below:

Microplate

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	CAL4	S 1									
B	BLK	CAL4	S 2									
C	CAL1	CAL5	S 3									
D	CAL1	CAL5	S 4									
E	CAL2	CAL6	S 5									
F	CAL2	CAL6	S 6									
G	CAL3	CS(*)	S 7									
H	CAL3	CS(*)	S 8									

Legenda: BLK = Blank CAL = Calibrator
CS(*) = Control Serum - Not mandatory S = Sample

An example of dispensation scheme in qualitative assays is reported below:

Microplate

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	S 3	S 11									
B	CAL1	S 4	S 12									
C	CAL1	S 5	S 13									
D	CAL2	S 6	S 14									
E	CAL2	S 7	S 15									
F	CAL6	S 8	S 16									
G	S 1	S 9	S 17									
H	S 2	S 10	S 18									

Legenda: BLK = Blank CAL = Calibrators
S = Sample

O. INTERNAL QUALITY CONTROL

A validation check is carried out on the calibrators any time the kit is used in order to verify whether the performances of the assay are as qualified.

Control that the following data are matched:

Check	Requirements
Blank well	< 0.050 OD450nm value
CAL 1 0 arbU/ml	< 0.150 mean OD450nm value after blanking coefficient of variation < 30%
CAL 2 5 arbU/ml	OD450nm > OD450nm CAL1 + 0.100
CAL 6 100 arbU/ml	OD450nm > 1.000

If the results of the test match the requirements stated above, proceed to the next section.

If they do not, do not proceed any further and operate as follows:

Problem	Check
Blank well > 0.050 OD450nm	1. that the Chromogen/Substrate solution has not got contaminated during the assay
CAL 1 0 arbU/ml > 0.150 OD450nm after blanking coefficient of variation > 30%	1. that the washing procedure and the washer settings are as validated in the pre qualification study; 2. that the proper washing solution has been used and the washer has been primed with it before use; 3. that no mistake has been done in the assay procedure (dispensation of a positive calibrator instead of the negative one); 4. that no contamination of the negative calibrator or of their wells has occurred due spills of positive samples or the enzyme conjugate; 5. that micropipettes haven't got contaminated with positive samples or with the enzyme conjugate 6. that the washer needles are not blocked or partially obstructed.
CAL 2 5 arbU/ml OD450nm < OD450nm CAL1 + 0.100	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (ex.: dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the calibrator has occurred.
CAL 6 100 arbU/ml < 1.000 OD450nm	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the positive control has occurred.

Should one of these problems have happened, after checking, report to the supervisor for further actions.

** Note:

If Control Serum has used, verify the following data:

Check	Requirements
Control Serum	Mean OD450nm CAL 4 $\pm$ 20%

If the results of the test doesn't match the requirements stated above, operate as follows:

Problem	Check
Control Serum Different from expected value	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the control serum has occurred.

Anyway, if all other parameters (Blank, CAL1, CAL2, CAL6), match the established requirements, the test may be considered valid.

P. RESULTS


P.1 Quantitative method

If the test turns out to be valid, use for the quantitative method an approved curve fitting program to draw the calibration curve from the values obtained by reading at 450nm (4-parameters interpolation is suggested).

Then on the calibration curve calculate the concentration of anti Herpes Simplex Virus type 1 IgG antibody in samples.

An example of Calibration curve is reported in the next page.

Example of Calibration Curve :



Important Note:

Do not use the calibration curve above to make calculations.

P.2 Qualitative method

In the qualitative method, calculate the mean OD450nm values for the Calibrators 0 and 5 arbU/ml and then check that the assay is valid.

Example of calculation:

The following data must not be used instead of real figures obtained by the user.

Calibrator 0 arbU/ml: 0.020 – 0.024 OD450nm
 Mean Value: 0.022 OD450nm
 Lower than 0.150 – Accepted

Calibrator 5 arbU/ml: 0.350 – 0.370 OD450nm
 Mean Value: 0.360 OD450nm
 Higher than Cal 0 + 0.100 – Accepted

Calibrator 100 arbU/ml: 2.245 OD450nm
 Higher than 1.000 – Accepted

Q. INTERPRETATION OF RESULTS

Samples with a concentration lower than 5 arbU/ml are considered negative for anti HSV1 IgG antibody.

Samples with a concentration higher than 5 arbU/ml are considered positive for anti HSV1 IgG antibody.

Particular attention in the interpretation of results has to be used in the follow-up of pregnancy for a primary infection of HSV due to the risk of neonatal malformations.

Important notes:

1. Interpretation of results should be done under the supervision of the laboratory supervisor to reduce the risk of judgment errors and misinterpretations.
2. When test results are transmitted from the laboratory to another facility, attention must be paid to avoid erroneous data transfer.
3. In the follow-up of pregnancy for HSV infection a positive result (presence of IgG antibody > 5 arbU/ml) should be confirmed to ruled out the risk of a false positive result and a false definition of protection.

R. PERFORMANCES

1. Limit of detection

The limit of detection of the assay has been calculated by means of an internal Gold Standard in absence of an international preparation to refer to.

The limit of detection has been calculated as mean OD450nm Calibrator 0 arbU/ml + 5 SD.

The table below reports the mean OD450nm values of this standard when diluted in negative plasma and then examined in the assay for three lots.

Mean OD450nm values (n = 2)

IgG arbU/ml	HSV1G.PU Lot # 0703	HSV1G.PU Lot # 1203	HSV1G.PU Lot # 0204/2
0	0.077	0.034	0.043
5	0.355	0.404	0.318
10	0.742	0.713	0.516
20	1.254	1.216	0.944
50	1.952	1.928	1.728
100	2.623	2.261	2.072

The assay shows a limit of detection far better than 5 arbU/ml.

In addition the preparation code Accurun n° 150, produced by Boston Biomedica Inc., BBI, USA, was tested in dilutions to determine the limit of its detection and provide a further value of analytical sensitivity

Mean OD450nm values (n = 2)

Dilution	HSV1G.CE Lot # 1004	HSV1G.PU Lot # 1203	HSV1G.PU Lot # 0204/2
1 X	1.248	1.218	1.300
2 X	0.860	0.848	0.876
4 X	0.545	0.526	0.583
8 X	0.315	0.300	0.329
16 X	0.164	0.152	0.148
32 X	0.082	0.064	0.072
0 arbU/ml	0.057	0.050	0.047
5 arbU/ml	0.288	0.355	0.318

2. Diagnostic sensitivity:

The diagnostic sensitivity has been tested in a performance evaluation study on panels of samples classified positive by a kit US FDA approved. Positive samples from different stage of HSV infection were tested. The value, obtained from the analysis of more than 300 specimens, has been > 98%.

3. Diagnostic specificity:

The diagnostic specificity has been determined on panels of negative samples from not infected individuals, classified negative with a kit US FDA approved.

Both plasma, derived with different standard techniques of preparation (citrate, EDTA and heparin), and sera have been used to determine the value of specificity.

Frozen specimens have been tested, as well, to check for interferences due to collection and storage.

No interference was observed.

Potentially interfering samples derived from patients with different pathologies (mostly ANA, AMA and RF positive) and from pregnant women were tested.

No crossreaction was observed.

An overall value > 98% of specificity was found when examined on more than 100 specimens.

3. Precision:

It has been calculated on the Calibrator 5 arbU/ml, considered the cut-off of the assay, examined in 16 replicates in three separate runs for three lots.

Results are reported as follows:

HSV1G.CE Lot # 1004

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.292	0.290	0.285	0.289
Std.Deviation	0.024	0.024	0.027	0.025
CV %	8.24	8.28	9.42	8.65

HSV1G.PU: lot 1203

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.365	0.382	0.378	0.375
Std.Deviation	0.022	0.029	0.018	0.023
CV %	6.02	7.59	4.76	6.12

HSV1G.PU: Lot 0204/2

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.322	0.298	0.304	0.308
Std.Deviation	0.018	0.019	0.016	0.018
CV %	5.59	6.38	5.26	5.74

The variability shown in the tables above did not result in sample misclassification.

S. LIMITATIONS OF THE PROCEDURE

Bacterial contamination or heat inactivation of the specimen may affect the absorbance values of the samples with consequent alteration of the level of the analyte.

Frozen samples containing fibrin particles or aggregates after thawing may generate some false results.

This test is suitable only for testing single samples and not pooled ones.

Diagnosis of an infectious disease should not be established on the basis of a single test result. The patient's clinical history, symptomatology, as well as other diagnostic data should be considered.

REFERENCES

- Engvall E. and Perlmann P.. J.Immunochemistry 8: 871-874, 1971
- Engvall E. and Perlmann P.. J.Immunol.. 109: 129-135, 1971
- Remington J.S. and Klein J.O.. (1996) In "Infectious diseases of fetus and newborn infant". Sanders, Philadelphia, London, Toronto.
- Volk W.A. (1982) In "essential of Medical Microbiology". 2nd ed., pp 729, G.B. Lippincott Co. Philadelphia, New York, S.Josè, Toronto.
- Leinikki P.O. et al.. J.Clin.Microbiol.. 8:418, 1978
- Piroid E. et al.. Révue Méd.Vet.. 131:25, 1980.
- Vaheri A. et al.. J.Med.Virol.. 5:171, 1980.
- Vejtorp M. et al.. Acta Path.Microbiol.Scand.. 88:349, 1980.
- Voller A. et al.. Brit.J.Exp.Pathol.. 56:338, 1975

All the IVD Products manufactured by the company are under the control of a certified Quality Management System in compliance with ISO 13485 rule. Each lot is submitted to a quality control and released into the market only if conforming with the EC technical specifications and acceptance criteria.

Manufacturer:

Dia.Pro Diagnostic Bioprobes Srl
Via G. Carducci n° 27 – Sesto San Giovanni (MI) – Italy





Dia.Pro
Diagnostic
Bio**Probes**

EC DECLARATION OF CONFORMITY

MANUFACTURER	DIA.PRO DIAGNOSTIC BIOPROBES S.R.L. VIA G. CARDUCCI N° 27 – 20099 SESTO SAN GIOVANNI (MILANO) – ITALY
PRODUCT	HSV1 IgG CODE: HSV1G.CE (96 tests)
CLASSIFICATION	GENERAL IVD
CONFORMITY ASSESSMENT ROUTE	SELF CERTIFICATION

**WE HEREBY DECLARE THAT THE ABOVE MENTIONED PRODUCT MEETS
THE PROVISIONS OF THE COUNCIL DIRECTIVE 98/79/EC
FOR IN VITRO DIAGNOSTIC DEVICES.**

ISO CERTIFICATE	UNE EN ISO 13485 N° 2013 11 0039 EN, RELEASED BY AEMPS (AGENCIA ESPAÑOLA DE MEDICAMENTOS Y PRODUCTOS SANITARIOS)
------------------------	---

PLACE & DATE OF FIRST ISSUE	MILANO – OCTOBER 2004
PLACE & DATE OF CURRENT ISSUE	SESTO SAN GIOVANNI (MI) – MARCH 2019
SIGNATURE Legal Representative Dr.ssa Fiorenza Scozzesi	

Rev: 05/2018

HSV1 IgM

**“Capture” Enzyme Immuno Assay
(ELISA) for the determination
of IgM antibodies to
Herpes Simplex Virus type 1
in human plasma and sera**

- for “in vitro” diagnostic use only -



DIA.PRO

**Diagnostic Bioprobes Srl
Via G. Carducci n° 27
20099 Sesto San Giovanni
(Milano) - Italy**

Phone +39 02 27007161

Fax +39 02 26007726

e-mail: info@diapro.it

REF HSV1M.CE
96 tests

HSV1 IgM

A. INTENDED USE

Enzyme ImmunoAssay (ELISA) for the determination of IgM antibodies to Herpes Simplex Virus types 1 in human plasma and sera with the "capture" system. The device is intended for the follow-up of HSV1 infected patients and for the monitoring of risk of neonatal defects due to HSV infection during pregnancy.

For "in vitro" diagnostic use only.

B. INTRODUCTION

Herpes Simplex Virus type 1 (HSV1) and type 2 (HSV2) are large complex DNA-containing viruses which have been shown to induce the synthesis of several proteins during infection, possessing an high number of cross-reactive determinants and just a few of type-specific sequences.

The majority of primary and recurrent genital herpetic infections are caused by HSV2; while non genital infections, such as common cold sores, are caused primarily by HSV1.

The detection of virus specific IgG and IgM antibodies are important in the diagnosis of acute/primary virus infections or reactivations of a latent one, in the absence of evident clinical symptoms.

A-symptomatic infections may happen for HSV in apparently healthy individuals and during pregnancy. Severe herpetic infections may happen in immuno-compromised and suppressed patients in which the disease may evolve toward critical pathologies.

The determination of HSV specific antibodies has then become important in the monitoring of "risk" patients and in the follow up of acute and severe infections.

C. PRINCIPLE OF THE TEST

The assay is based on the principle of "IgM capture" where IgM class antibodies in the sample are first captured by the solid phase coated with anti hIgM antibody.

After washing out all the other components of the sample and in particular IgG antibodies, the specific IgM captured on the solid phase are detected by the addition of a preparation of inactivated HSV1, labeled with a HSV1 specific antibody conjugated with peroxidase (HRP).

After incubation, microwells are washed to remove unbound conjugate and then the chromogen/substrate is added.

In the presence of bound conjugate the colorless substrate is hydrolyzed to a colored end-product, whose optical density may be detected and is proportional to the amount of IgM antibodies to HSV1 present in the sample.

A system is described how to control whether the positivity shown by a sample is true or not (Confirmation Test), helpful for the clinician to make a correct interpretation of results.

D. COMPONENTS

The kit contains reagents for 96 tests.

1. Microplate: MICROPLATE

12 strips x 8 microwells coated with anti human IgM affinity purified goat antibody, in presence of bovine proteins. Plates are sealed into a bag with desiccant. Allow the microplate to reach room temperature before opening; reseal unused strips in the bag with desiccant and store at 2..8°C.

2. Negative Control: CONTROL -

1x4.0 ml/vial. Ready to use control. It contains 1% human serum proteins, 2% casein, 10 mM tris buffer pH 6.0+/-0.1, 0.1%

Tween 20, 0.09% sodium azide and 0.1% Kathon GC as preservatives.

The negative control is cpale yellow color coded..

3. Positive Control: CONTROL +

1x4.0 ml/vial. Ready to use control. It contains 1% human serum positive for HSV1 IgM, 2% casein, 10 mM tris buffer pH 6.0+/-0.1, 0.1% Tween 20, 0.09% sodium azide and 0.1% Kathon GC as preservatives.

The positive control is green colour coded.

4. Calibrator: CAL ...ml

N° 1 lyophilized vial. To be dissolved with EIA grade water as reported in the label. It contains anti HSV1 IgM, fetal bovine serum, 0.2 mg/ml gentamicine sulphate and 0.1% Kathon GC as preservatives.

Note: The volume necessary to dissolve the content of the vial may vary from lot to lot. Please use the right volume reported on the label.

5. Lyophilized HSV1 Ag: AG HSV1

N° 6 lyophilized vials. The vials contain gamma-ray inactivated HSV1 in protein buffer. The solution contains 2% bovine proteins, 10 mM Tris HCl buffer pH 6.8+/-0.1, 0.2 mg/ml gentamicine sulphate and 0.1% Kathon GC. To be dissolved with 1.9 ml of Antigen Diluent as reported in the specific section.

6. Wash buffer concentrate: WASHBUF 20X

1x60ml/bottle. 20x concentrated solution. Once diluted, the wash solution contains 10 mM phosphate buffer pH 7.0+/-0.2, 0.05% Tween 20 and 0.1% Kathon GC.

7. Enzyme conjugate: CONJ 20X

1x0.8 ml/vial. 20x concentrated solution of a HSV1-specific antibody, labeled with HRP and diluted in a protein buffer containing 10 mM Tris buffer pH 6.8+/-0.1, 2% BSA, 0.1% Kathon GC and 0.2 mg/ml gentamicine sulphate as preservatives.

8. Antigen Diluent : AG DIL

n° 1 vial of 16 ml. Protein buffer solution for the preparation of the Immunocomplex. The solution contains 10 mM Tris buffer pH 6.8+/-0.1, 2% BSA, 0.1% Kathon GC and 0.2 mg/ml gentamicine sulphate as preservatives. The reagent is code coloured with 0.01% red alimentary dye

9. Specimen Diluent : DILSPE

2x60.0 ml/vial. Proteic buffered solution for the dilution of samples. It contains 2% casein, 10 mM tris buffer pH 6.0+/-0.1, 0.1% Tween 20, 0.09% sodium azide and 0.1% Kathon GC as preservatives.

The reagent is color coded with 0.01% blue alimentary dye.

10. Chromogen/Substrate : SUBS TMB

1x16ml/vial. It contains a 50 mM citrate-phosphate buffered solution at pH 3.5-3.8, 0.03% tetra-methyl-benzidine (TMB), 0.02% hydrogen peroxide (H₂O₂) and 4% dimethylsulphoxide.

Note: To be stored protected from light as sensitive to strong illumination.

11. Sulphuric Acid: H₂SO₄ 0.3 M

1x15ml/vial. It contains 0.3 M H₂SO₄ solution.

Attention: Irritant (H315, H319; P280, P302+P352, 332+P313, P305+ P351+P338, P337+P313, P362+P363)

12. Plate sealing foils n° 2

13. Package insert n° 1

E. MATERIALS REQUIRED BUT NOT PROVIDED

1. Calibrated Micropipettes (1000 ul, 100 ul and 10 ul) and disposable plastic tips.
2. EIA grade water (double distilled or deionised, charcoal treated to remove oxidizing chemicals used as disinfectants).
3. Timer with 60 minute range or higher.
4. Absorbent paper tissues.
5. Calibrated ELISA microplate thermostatic incubator (dry or wet), set at +37°C (+/-0.5°C tolerance)..
6. Calibrated ELISA microwell reader with 450nm (reading) and with 620-630nm (blanking) filters.
7. Calibrated ELISA microplate washer.
8. Vortex or similar mixing tools.

F. WARNINGS AND PRECAUTIONS

1. The kit has to be used by skilled and properly trained technical personnel only, under the supervision of a medical doctor responsible of the laboratory.
2. All the personnel involved in performing the assay have to wear protective laboratory clothes, talc-free gloves and glasses. The use of any sharp (needles) or cutting (blades) devices should be avoided. All the personnel involved should be trained in biosafety procedures, as recommended by the Center for Disease Control, Atlanta, U.S. and reported in the National Institute of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
3. All the personnel involved in sample handling should be vaccinated for HBV and HAV, for which vaccines are available, safe and effective.
4. The laboratory environment should be controlled so as to avoid contaminants such as dust or air-born microbial agents, when opening kit vials and microplates and when performing the test. Protect the Chromogen (TMB) from strong light and avoid vibration of the bench surface where the test is undertaken.
5. Upon receipt, store the kit at 2..8°C into a temperature controlled refrigerator or cold room.
6. Do not interchange components between different lots of the kits. It is recommended that components between two kits of the same lot should not be interchanged.
7. Check that the reagents are clear and do not contain visible heavy particles or aggregates. If not, advise the laboratory supervisor to initiate the necessary procedures for kit replacement.
8. Avoid cross-contamination between serum/plasma samples by using disposable tips and changing them after each sample.
9. Avoid cross-contamination between kit reagents by using disposable tips and changing them between the use of each one.
10. Do not use the kit after the expiration date stated on the external container and internal (vials) labels. A study conducted on an opened kit did not pointed out any relevant loss of activity up to six 6 uses of the device and up to 3 months.
11. Treat all specimens as potentially infective. All human serum specimens should be handled at Biosafety Level 2, as recommended by the Center for Disease Control, Atlanta, U.S. in compliance with what reported in the Institutes of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
12. The use of disposable plastic-ware is recommended in the preparation of the liquid components or in transferring components into automated workstations, in order to avoid cross contamination.
13. Waste produced during the use of the kit has to be discarded in compliance with national directives and laws concerning laboratory waste of chemical and biological substances. In particular, liquid waste generated from the washing procedure, from residuals of controls and from samples has to be treated as potentially infective material and inactivated before waste. Suggested procedures of inactivation are treatment with a 10% final concentration of household bleach for 16-18 hrs or heat inactivation by autoclave at 121°C for 20 min..

14. Accidental spills from samples and operations have to be adsorbed with paper tissues soaked with household bleach and then with water. Tissues should then be discarded in proper containers designated for laboratory/hospital waste.

15. The Sulphuric Acid is an irritant. In case of spills, wash the surface with plenty of water

16. Other waste materials generated from the use of the kit (example: tips used for samples and controls, used microplates) should be handled as potentially infective and disposed according to national directives and laws concerning laboratory wastes.

G. SPECIMEN: PREPARATION AND WARNINGS

1. Blood is drawn aseptically by venepuncture and plasma or serum is prepared using standard techniques of preparation of samples for clinical laboratory analysis. No influence has been observed in the preparation of the sample with citrate, EDTA and heparin.
2. Samples have to be clearly identified with codes or names in order to avoid misinterpretation of results. Bar code labeling and electronic reading is strongly recommended.
3. Haemolysed ("red") and visibly hyperlipemic ("milky") samples have to be discarded as they could generate false results. Samples containing residues of fibrin or heavy particles or microbial filaments and bodies should be discarded as they could give rise to false results.
4. Sera and plasma can be stored at +2..8°C for up to five days after collection. For longer storage periods, samples can be stored frozen at -20°C for several months. Any frozen samples should not be frozen/thawed more than once as this may generate particles that could affect the test result.
5. If particles are present, centrifuge at 2.000 rpm for 20 min or filter using 0.2-0.8u filters to clean up the sample for testing.

H. PREPARATION OF COMPONENTS AND WARNINGS

A study conducted on an opened kit has not pointed out any relevant loss of activity up to 6 re-uses of the device and up to 3 months.

Microplate:

Allow the microplate to reach room temperature (about 1 hr) before opening the container. Check that the desiccant has not turned dark green, indicating a defect in storing. In this case, call Dia.Pro's customer service.

Unused strips have to be placed back into the aluminum pouch, with the desiccant supplied, firmly zipped and stored at +2..8°C. After first opening, remaining strips are stable until the humidity indicator inside the desiccant bag turns from yellow to green.

Negative Control:

Ready to use. Mix well on vortex before use.

Positive Control:

Ready to use. Mix well on vortex before use.

Calibrator:

Add the volume of ELISA grade water reported on the label to the lyophilized powder. Let fully dissolve and then gently mix on vortex.

Important Note: The solution is not stable. Store the Calibrator frozen in aliquots at -20°C.

Wash buffer concentrate:

The whole content of the concentrated solution has to be diluted 20x with bidistilled water and mixed gently end-over-end before use. During preparation avoid foaming as the presence of bubbles could impact on the efficiency of the washing cycles.

Note: Once diluted, the wash solution is stable for 1 week at +2..8° C.

Ag/Ab Immunocomplex:

Proceed carefully as follows:

1. Dissolve the content of a lyophilized vial with 1.9 ml of Conjugate/Antigen Diluent. Let fully dissolved the lyophilized content and then gently mix on vortex.
2. Gently mix the concentrated Enzyme Conjugate on vortex. Then add 0.1 ml of it to the vial of the dissolved HSV1 Ag and mix gently on vortex.

Important Notes:

1. *Dissolve and prepare only the number of vials necessary to the test. The Immunocomplex obtained is not stable. Store any residual solution frozen in aliquots at -20°C.*
2. *The preparation of the Immucplex has to be done **right before** the dispensation of samples and controls into the plate. Mix again on vortex gently just before its use.*

Specimen Diluent:

Ready to use. Mix well on vortex before use

Chromogen/Substrate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

Do not expose to strong illumination, oxidizing agents and metallic surfaces.

If this component has to be transferred use only plastic, possible sterile disposable container

Sulphuric Acid:

Ready to use. Mix well on vortex before use.

Attention: Irritant (H315, H319; P280, P302+P352, 332+P313, P305+ P351+P338, P337+P313, P362+P363).

Legenda:

Warning H statements:

H315 – Causes skin irritation.

H319 – Causes serious eye irritation.

Precautionary P statements:

P280 – Wear protective gloves/protective clothing/eye protection/face protection.

P302 + P352 – IF ON SKIN: Wash with plenty of soap and water.

P332 + P313 – If skin irritation occurs: Get medical advice/attention.

P305 + P351 + P338 – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P337 + P313 – If eye irritation persists: Get medical advice/attention.

P362 + P363 - Take off contaminated clothing and wash it before reuse.

I. INSTRUMENTS AND TOOLS USED IN COMBINATION WITH THE KIT

1. Micropipettes have to be calibrated to deliver the correct volume required by the assay and must be submitted to regular decontamination (household alcohol, 10% solution of bleach, hospital grade disinfectants) of those parts that could accidentally come in contact with the sample. They should also be regularly maintained in order to show a precision of 1% and a trueness of +/-2%. Decontamination of spills or residues of kit components should also be carried out regularly.
2. The ELISA incubator has to be set at +37°C (tolerance of +/- 0.5°C) and regularly checked to ensure the correct temperature is maintained. Both dry incubators and water baths are suitable for the incubations, provided that the instrument is validated for the incubation of ELISA tests.
3. The ELISA washer is extremely important to the overall performances of the assay. The washer must be carefully validated and correctly optimised using the kit controls and reference panels, before using the kit for routine laboratory tests. Usually 4-5 washing cycles (aspiration + dispensation

of 350ul/well of washing solution = 1 cycle) are sufficient to ensure that the assay performs as expected. A soaking time of 20-30 seconds between cycles is suggested. In order to set correctly their number, it is recommended to run an assay with the kit controls and well characterized negative and positive reference samples, and check to match the values reported below in the section "Internal Quality Control". Regular calibration of the volumes delivered by, and maintenance (decontamination and cleaning of needles) of the washer has to be carried out according to the instructions of the manufacturer.

4. Incubation times have a tolerance of $\pm 5\%$.
5. The ELISA microplate reader has to be equipped with a reading filter of 450nm and with a second filter (620-630nm, strongly recommended) for blanking purposes. Its standard performances should be (a) bandwidth ≤ 10 nm; (b) absorbance range from 0 to ≥ 2.0 ; (c) linearity to ≥ 2.0 ; repeatability $\geq 1\%$. Blanking is carried out on the well identified in the section "Assay Procedure". The optical system of the reader has to be calibrated regularly to ensure that the correct optical density is measured. It should be regularly maintained according to the manufacturer 's instructions.
6. When using an ELISA automated work station, all critical steps (dispensation, incubation, washing, reading, data handling) have to be carefully set, calibrated, controlled and regularly serviced in order to match the values reported in the section "Internal Quality Control". The assay protocol has to be installed in the operating system of the unit and validated as for the washer and the reader. In addition, the liquid handling part of the station (dispensation and washing) has to be validated and correctly set. Particular attention must be paid to avoid carry over by the needles used for dispensing and for washing. This must be studied and controlled to minimize the possibility of contamination of adjacent wells. The use of ELISA automated work stations is recommended when the number of samples to be tested exceed 20-30 units per run.
7. Dia.Pro's customer service offers support to the user in the setting and checking of instruments used in combination with the kit, in order to assure compliance with the requirements described. Support is also provided for the installation of new instruments to be used with the kit.

L. PRE ASSAY CONTROLS AND OPERATIONS

1. Check the expiration date of the kit printed on the external label (primary container). Do not use the device if expired.
2. Check that the liquid components are not contaminated by visible particles or aggregates. Check that the Chromogen/Substrate is colorless or pale blue by aspirating a small volume of it with a sterile plastic pipette. Check that no breakage occurred in transportation and no spillage of liquid is present inside the box (primary container). Check that the aluminum pouch, containing the microplate, is not punctured or damaged.
3. Dilute all the content of the 20x concentrated Wash Solution as described above.
4. Dissolve the Calibrator as described above and gently mix.
5. Allow all the other components to reach room temperature (about 1 hr) and then mix gently on vortex all liquid reagents.
6. Set the ELISA incubator at +37°C and prepare the ELISA washer by priming with the diluted washing solution, according to the manufacturers instructions. Set the right number of washing cycles as found in the validation of the instrument for its use with the kit.
7. Check that the ELISA reader is turned on or ensure it will be turned on at least 20 minutes before reading.
8. If using an automated work station, turn on, check settings and be sure to use the right assay protocol.
9. Check that the micropipettes are set to the required volume.

- Check that all the other equipment is available and ready to use.
- In case of problems, do not proceed further with the test and advise the supervisor.

M. ASSAY PROCEDURE

The assay has to be carried out according to what reported below, taking care to maintain the same incubation time for all the samples in testing.

M.1 Automated assay:

In case the test is carried out automatically with an ELISA system, we suggest to make the instrument aspirate 1000 µl Specimen Diluent and then 10 µl sample (1:101 dilution factor). The whole content is then dispensed into a properly defined dilution tube. Before the next sample is aspirated, needles have to be duly washed to avoid any cross-contamination among samples. When all the samples have been diluted make the instrument dispense 100 µl diluted samples into the proper wells of the microplate.

This procedure may be carried out also in two steps of dilutions of 1:10 each (90 µl Specimen Diluent + 10 µl sample) into a second dilution platform. Make then the instrument aspirate first 100 µl Specimen Diluent, then 10 µl liquid from the first dilution in the platform and finally dispense the whole content in the proper well of the assay microplate.

Do not dilute controls/calibrator as they are ready to use.

Dispense 100 µl calibrators/controls in the appropriate calibration/control wells.

For the next operations follow the operative instructions reported below for the Manual Assay.

It is strongly recommended to check that the time lap between the dispensation of the first and the last sample will be calculated by the instrument and taken into consideration by delaying the first washing operation accordingly.

M. 2 Manual assay:

- Dilute samples 1:101 by dispensing first 10 µl sample and then 1 ml Specimen Diluent into a dilution tube; mix gently on vortex.
- Place the required number of Microwells in the microwell holder. Leave the well in position A1 empty for the operation of blanking.
- Dispense 100 µl of Negative Control and 100 µl of Calibrator in the proper wells in duplicate. Dispense 100 µl of Positive Control in single into the proper well. Do not dilute controls and the calibrator as they are ready to use !
- Dispense 100 µl diluted samples in the proper sample wells and then check that all the samples wells are blue colored and that controls and calibrator have been dispensed.
- Incubate the microplate for **60 min at +37°C**.

Important note: Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

- Wash the microplate with an automatic washer as reported previously (section I.3).
- Pipette 100 µl of the **Ag/Ab Immunocomplex** into each well, except the blanking well A1, and cover with the sealer. Check that all wells are red colored, except A1.

Important note: Be careful not to touch the plastic inner surface of the well with the tip filled with the **Ag/Ab Immunocomplex**. Contamination might occur.

- Incubate the microplate for **60 min at +37°C**.
- Wash microwells as in step 6.
- Pipette 100 µl Chromogen/Substrate mixture into each well, the blank well included. Then incubate the microplate at **room temperature (18-24°C) for 20 minutes**.

Important note: Do not expose to strong direct illumination. High background might be generated.

- Pipette 100 µl Sulphuric Acid into all the wells using the same pipetting sequence as in step 10. Addition of acid will turn the positive control and positive samples from blue to yellow.
- Measure the color intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and at 620-630nm (background subtraction, strongly recommended), blanking the instrument on A1.

Important notes:

- If the second filter is not available ensure that no finger prints are present on the bottom of the microwell before reading at 450nm. Finger prints could generate false positive results on reading.
- Reading has to be carried out just after the addition of the Stop Solution and anyway not any longer than 20 minutes after its addition. Some self oxidation of the chromogen can occur leading to high background.

N. ASSAY SCHEME

Controls&calibrator(*)	100 ul
Samples diluted 1:101	100 ul
1st incubation	60 min
Temperature	+37°C
Washing	4-5 cycles
Immunocomplex	100 ul
2nd incubation	60 min
Temperature	+37°C
Washing	4-5 cycles
TMB/H2O2 mix	100 ul
3rd incubation	20 min
Temperature	r.t.
Sulphuric Acid	100 ul
Reading OD	450nm

(*) Important Notes:

- The Calibrator (CAL) does not affect the Cut Off calculation, therefore it does not affect the test's results calculation.
- The Calibrator (CAL) used only if a laboratory internal quality control is required by the Management.

An example of dispensation scheme is reported below:

Microplate												
	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	S3										
B	NC	S4										
C	NC	S5										
D	CAL(*)	S6										
E	CAL(*)	S7										
F	PC	S8										
G	S1	S9										
H	S2	S10										

Legenda: BLK = Blank NC = Negative Control
CAL(*) = Calibrator-Not mandatory PC = Positive Control S = Sample

O. INTERNAL QUALITY CONTROL

A validation check is carried out on the controls any time the kit is used in order to verify whether the performances of the assay are as qualified.

Control that the following data are matched:

Parameter	Requirements
Blank well	< 0.05 OD450nm value
Negative Control mean value (NC)	< 0.200 OD450nm value after blanking coefficient of variation < 30%
Positive Control	> 1.000 OD450nm

If the results of the test match the requirements stated above, proceed to the next section.

If they do not, do not proceed any further and perform the following checks:

Problem	Check
Blank well > 0.05 OD450nm	1. that the Chromogen/Substrate solution has not become contaminated during the assay
Negative Control (NC) > 0.200 OD450nm after blanking coefficient of variation > 30%	1. that the washing procedure and the washer settings are as validated in the pre qualification study; 2. that the proper washing solution has been used and the washer has been primed with it before use; 3. that no mistake has been done in the assay procedure (dispensation of positive control instead of negative control); 4. that no contamination of the negative control or of the wells where the control was dispensed has occurred due to positive samples, to spills or to the enzyme conjugate; 5. that micropipettes have not become contaminated with positive samples or with the enzyme conjugate 6. that the washer needles are not blocked or partially obstructed.
Positive Control < 1.000 OD450nm	1. that the procedure has been correctly performed; 2. that no mistake has occurred during the distribution of the control (dispensation of negative control instead of positive control). 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the positive control has occurred.

If any of the above problems have occurred, report the problem to the supervisor for further actions.

** Important Note:

If the Calibrator has used, verify the following data:

Check	Requirements
Calibrator	S/Co > 1.2

If the results of the test doesn't match the requirements stated above, operate as follows:

Problem	Check
Calibrator S/Co < 1.2	1. that the procedure has been correctly performed; 2. that no mistake has occurred during its distribution (e.g.: dispensation of negative control instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the calibrator has occurred.

Anyway, if all other parameters (Blank, Negative Control, Positive Control), match the established requirements, the test may be considered valid.

P. CALCULATION OF THE CUT-OFF

The test results are calculated by means of the mean OD450nm value of the Negative Control (NC) and a mathematical calculation, in order to define the following cut-off formulation:

$$\text{Cut-Off} = \text{NC} + 0.250$$

The value found for the test is used for the interpretation of results as described in the next paragraph.

Important note: When the calculation of results is performed by the operating system of an ELISA automated work station, ensure that the proper formulation is used to calculate the cut-off value and generate the correct interpretation of results.

Q. INTERPRETATION OF RESULTS

Test results are interpreted as a ratio of the sample OD450nm and the Cut-Off value (or S/Co) according to the following table:

S/Co	Interpretation
< 1.0	Negative
1.0 - 1.2	Equivocal
> 1.2	Positive

A negative result indicates that the patient is not undergoing an acute infection of Herpes Simplex Virus type 1.

Any patient showing an equivocal result, should be re-tested by examining a second sample taken from the patient after 1-2 weeks from first testing.

A positive result is indicative of a Herpes Simplex Virus type 1 infection.

An example of calculation is reported below:

Important Note: The following data must not be used instead or real figures obtained by the user.

Negative Control: 0.100 – 0.120 – 0.080 OD450nm

Mean Value: 0.100 OD450nm

Lower than 0.150 – Accepted

Positive Control: 1.850 OD450nm

Higher than 1.000 – Accepted

$$\text{Cut-Off} = 0.110 + 0.250 = 0.360$$

Calibrator: 1.000 - 0.900 OD450nm

Mean value: 0.950 OD450nm

S/Co higher than 1.2 – Accepted S/Co = 2.6

Sample 1: 0.075 OD450nm

Sample 2: 1.580 OD450nm

Sample 1 S/Co < 1 = negative

Sample 2 S/Co > 1.2 = positive

Important notes:

1. Interpretation of results should be done under the supervision of the laboratory supervisor to reduce the risk of judgment errors and misinterpretations.
2. Particular attention in the interpretation of results has to be used in the follow-up of pregnancy for an infection of HSV due to the risk of severe neonatal malformations.
3. In pregnancy monitoring, it is strongly recommended that any positive result is confirmed first with the procedure described below and secondly with a different device for HSV IgM detection, before taking any preventive medical action.
4. Any positive sample should be submitted to the Confirmation Test reported in section T before giving a result of positivity. By carrying out this test, false reactions, leading to a misinterpretation of the analytical result, can be revealed and then ruled out.
5. When test results are transmitted from the laboratory to another facility, attention must be paid to avoid erroneous data transfer.
6. Diagnosis of infection has to be taken and released to the patient by a suitably qualified medical doctor.

False positive reactions may be anyway pointed out and then ruled out in the interpretation of results with the procedure reported in section T, able to verify whether or not a positive result is real.

4. Precision:

Results are reported as follows:

HSV1M.CE: lot # RD1

Negative (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.083	0.107	0.116	0.102
Std.Deviation	0.004	0.017	0.013	0.011
CV %	5.12	15.82	11.59	10.84

Low reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.393	0.436	0.421	0.417
Std.Deviation	0.031	0.019	0.007	0.019
CV %	7.93	4.38	1.68	4.66

High reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	1.469	1.530	1.541	1.513
Std.Deviation	0.034	0.055	0.037	0.042
CV %	2.31	3.60	2.39	2.77

HSV1M.CE: lot # RD2

Negative (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.101	0.099	0.097	0.099
Std.Deviation	0.009	0.011	0.013	0.011
CV %	8.91	11.11	13.40	11.14

Low reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.412	0.395	0.420	0.409
Std.Deviation	0.015	0.009	0.012	0.012
CV %	3.64	2.27	2.86	2.92

High reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	1.512	1.498	1.534	1.515
Std.Deviation	0.042	0.035	0.028	0.035
CV %	2.78	2.34	1.83	2.31

HSV1M.CE: lot # RD3

Negative (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.095	0.112	0.092	0.100
Std.Deviation	0.012	0.009	0.010	0.011
CV %	12.6	8.04	10.86	10.50

Low reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.405	0.398	0.412	0.405
Std.Deviation	0.012	0.015	0.014	0.014
CV %	2.96	3.77	3.40	3.37

R. PERFORMANCE CHARACTERISTICS

1. Limit of detection

No international standard for HSV1&2 IgM Antibody detection has been defined so far by the European Community.

In its absence, an Internal Gold Standard (or IGS), calibrated on the preparation named "Accurun – Anti HSV2 IgM plasma" produced by Boston Biomedica Inc., USA, code 9106072, has been defined in order to provide the device with a constant and excellent sensitivity.

The limit of detection of the assay has been therefore calculated on the IGS. A limiting dilution curve was prepared in the Negative Control (NC).

Results of Quality Control are given in the following table:

OD450nm values

IGS	HSV1M.CE Lot # RD1	HSV1M.CE Lot # RD2	HSV1M.CE Lot # RD3
1X	0.450	0.460	0.455
2X	0.277	0.300	0.288
4X	0.216	0.198	0.185
NC	0.115	0.085	0.086

2. Diagnostic sensitivity:

The diagnostic sensitivity has been tested in a performance evaluation study on panels of 40 samples classified positive by a CE marked kit. The value obtained from the analysis was > 98%.

3. Diagnostic specificity:

The diagnostic specificity has been determined in the performance evaluation on panels of more than 300 specimens, negative with the reference kit, derived from normal individuals of European origin.

Both plasma, derived with different standard techniques of preparation (citrate, EDTA and heparin), and sera have been used to determine the specificity. No false reactivity due to the method of specimen preparation has been observed.

Frozen specimens have also been tested to check whether this interferes with the performance of the test. No interference was observed on clean and particle free samples.

A study conducted on more than 60 potentially cross-reactive samples has not revealed any interference in the system.

No cross reaction were observed.

The Performance Evaluation has provided a value > 98%.

High reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	1.489	1.475	1.518	1.494
Std.Deviation	0.025	0.032	0.028	0.028
CV %	1.68	2.17	1.84	1.90

S. LIMITATIONS

Frozen samples containing fibrin particles or aggregates may generate false positive results.

Bacterial contamination or heat inactivation of the specimen may affect the absorbance values of the samples with consequent alteration of the level of the analyte.

This test is suitable only for testing single samples and not pooled ones.

Diagnosis of an infectious disease should not be established on the basis of a single test result. The patient's clinical history, symptomatology, as well as other diagnostic data should be considered.

T. CONFIRMATION TEST

In order to provide the medical doctor with the best accuracy in the follow-up of pregnancy, where a false positive result could lead to an operation of abortion, a confirmation test is reported. The confirmation test has to be carried out on any positive sample before a diagnosis of primary infection of HSV is released to the doctor.

Proceed for confirmation as follows:

1. Prepare the Antigen/Conjugate Complex as described in the proper section. This reagent is called Solution A.
2. Then 25 ul concentrated Enzymatic Conjugate are diluted in 500 ul Antigen Diluent and mixed gently on vortex. Do not use any lyophilized antigen vial for this procedure ! This solution is called Solution B.
3. The well A1 of the strip is left empty for blanking.
4. The Negative Control is dispensed in the strip in positions B1+C1. This is used for the calculation of the cut-off and S/Co values.
5. The positive sample to be confirmed, diluted 1:101, is dispensed in the strip in position D1+E1.
6. The strip is incubated for 60 min at +37°C.
7. After washing, the blank well A1 is left empty.
8. 100 µl of Solution A are dispensed in wells B1+C1+D1.
9. Then 100 µl of Solution B are added to well E1.
10. The strip is incubated for 60 min at +37°C.
11. After washing, 100 µl Chromogen/Substrate are added to all the wells and the strip is incubated for 20 min at r.t.
12. 100 µl Sulphuric Acid are added to all the wells and then their color intensity is measured at 450nm (reading filter) and at 620-630nm (background subtraction), blanking the instrument on A1.

Interpretation of results is carried out as follows:

1. If the sample in position D1 shows a S/Co value lower than 1.0 a problem of dispensation or contamination in the first test is likely to be occurred. The Assay Procedure in Section M has to be repeated to double check the analysis.
2. If the sample in position D1 shows a S/Co value higher than 1.2 and in position E1 shows a S/Co value still higher than 1.2 the sample is considered a **false positive**. The reactivity of the sample is in fact not dependent on the specific presence of HSV1 and a crossreaction with enzymatic conjugate has occurred.
3. If the sample in position D1 shows a S/Co value higher than 1.2 and in position E1 shows a S/Co value lower than 1.0 the sample is considered a **true positive**. The reactivity of the sample is in fact dependent on the specific presence of HSV and not due to any crossreaction.

The following table is reported for the interpretation of results

Well	S/Co		
D1	< 1.0	> 1.2	> 1.2
E1	< 1.0	> 1.2	< 1.0
Interpretation	Problem of contam.	False positive	True positive

REFERENCES

1. Engvall E. and Perlmann P.. J.Immunochemistry 8: 871-874, 1971
2. Engvall E. and Perlmann P.. J.Immunol.. 109: 129-135, 1971
3. Remington J.S. and Klein J.O.. (1996) In "Infectious diseases of fetus and newborn infant". Sanders, Philadelphia, London, Toronto.
4. Volk W.A. (1982) In "essential of Medical Microbiology". 2nd ed., pp 729, G.B. Lippincott Co. Philadelphia, New York, S.José, Toronto.
5. Leinikki P.O. et al.. J.Clin.Microbiol.. 8:418, 1978
6. Piroid E. et al.. Révue Méd.Vet.. 131:25, 1980.
7. Vaheri A. et al.. J.Med.Virol.. 5:171, 1980.
8. Vejtorp M. et al.. Acta Path.Microbiol.Scand.. 88:349, 1980.
1. Voller A. et al.. Brit.J.Exp.Pathol.. 56:338, 1975

All the IVD Products manufactured by the company are under the control of a certified Quality Management System in compliance with ISO 13485 rule. Each lot is submitted to a quality control and released into the market only if conforming with the EC technical specifications and acceptance criteria.

Manufacturer:
Dia.Pro Diagnostic Bioprobes Srl
Via G. Carducci n° 27 – Sesto San Giovanni (MI) – Italy





Dia.Pro
Diagnostic
Bio**Probes**

EC DECLARATION OF CONFORMITY

MANUFACTURER	DIA.PRO DIAGNOSTIC BIOPROBES S.R.L. VIA G. CARDUCCI N° 27 – 20099 SESTO SAN GIOVANNI (MILANO) – ITALY
PRODUCT	HSV1 IgM CODE: HSV1M.CE (96 tests)
CLASSIFICATION	GENERAL IVD
CONFORMITY ASSESSMENT ROUTE	SELF CERTIFICATION

**WE HEREBY DECLARE THAT THE ABOVE MENTIONED PRODUCT MEETS
THE PROVISIONS OF THE COUNCIL DIRECTIVE 98/79/EC
FOR IN VITRO DIAGNOSTIC DEVICES.**

ISO CERTIFICATE	UNE EN ISO 13485 N° 2013 11 0039 EN, RELEASED BY AEMPS (AGENCIA ESPAÑOLA DE MEDICAMENTOS Y PRODUCTOS SANITARIOS)
------------------------	---

PLACE & DATE OF FIRST ISSUE	MILANO – OCTOBER 2004
PLACE & DATE OF CURRENT ISSUE	SESTO SAN GIOVANNI (MI) – MARCH 2019
SIGNATURE Legal Representative Dr.ssa Fiorenza Scozzesi	

Rev: 05/2018

HSV2 IgG

**Enzyme ImmunoAssay (ELISA) for the
quantitative/qualitative determination
of IgG antibodies to
Herpes Simplex Virus type 2
in human serum and plasma**

- for “in vitro” diagnostic use only -



DIA.PRO

**Diagnostic Bioprobes Srl
Via G. Carducci n° 27
20099 Sesto San Giovanni
(Milano) - Italy**

Phone +39 02 27007161

Fax +39 02 44386771

e-mail: info@diapro.it

HSV2 IgG

A. INTENDED USE

Enzyme ImmunoAssay (ELISA) for the quantitative/qualitative determination of IgG antibodies to Herpes Simplex Virus type 2 in human plasma and sera.

For "in vitro" diagnostic use only.

B. INTRODUCTION

Herpes Simplex Virus type 1 (HSV1) and type 2 (HSV2) are large complex DNA-containing viruses which have been shown to induce the synthesis of several proteins during infection, possessing an high number of crossreactive determinants and just a few of type-specific sequences.

The majority of primary and recurrent genital herpetic infections are caused by HSV2; while non genital infections, such as common cold sores, are caused primarily by HSV1.

The detection of virus specific IgG and IgM antibodies are important in the diagnosis of acute/primary virus infections or reactivations of a latent one, in the absence of evident clinical symptoms.

Asymptomatic infections may happen for HSV in apparently healthy individuals and during pregnancy. Severe herpetic infections may happen in immunocompromised and suppressed patients in which the disease may evolve toward critical pathologies.

The determination of HSV specific antibodies has then become important in the monitoring of "risk" patients and in the follow up of acute and severe infections.

C. PRINCIPLE OF THE TEST

Microplates are coated with synthetic HSV2 specific glycoprotein G or gG.

The solid phase is first treated with the diluted sample and IgG to HSV2 are captured, if present, by the antigens.

After washing out all the other components of the sample, in the 2nd incubation bound anti HSV2 IgG are detected by the addition of polyclonal specific anti hIgG antibodies, labelled with peroxidase (HRP).

The enzyme captured on the solid phase, acting on the substrate/chromogen mixture, generates an optical signal that is proportional to the amount of anti HSV2 IgG antibodies present in the sample. A Calibration Curve, calibrated against an internal Gold Standard, makes possible a quantitative determination of the IgG antibody in the patient.

D. COMPONENTS

Each kit contains sufficient reagents to perform 96 tests.

1. Microplate: MICROPLATE

n° 1. 12 strips x 8 microwells coated with synthetic HSV2-specific gG in presence of bovine proteins.

Plates are sealed into a bag with desiccant. Allow the microplate to reach room temperature before opening; reseal unused strips in the bag with desiccant and store at 2..8°C.

2. Calibration Curve: CAL N° ...

Ready to use and color coded standard curve derived from human plasma positive for HSV2 IgG ranging:

4ml CAL1 = 0 arbU/ml
4ml CAL2 = 5 arbU/ml
2ml CAL3 = 10 arbU/ml
2ml CAL4 = 20 arbU/ml
2ml CAL5 = 50 arbU/ml
4ml CAL6 = 100 arbU/ml.

Standards are calibrated in arbitrary units against an internal Gold Standard (or IGS).

It contains human serum proteins, 2% casein, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.1% Tween 20, 0.09% Na-azide and ProClin 300 0.045% as preservatives. Standards are blue colored.

3. Control Serum: CONTROL ...ml

1 vial. Lyophilized. It contains fetal bovine serum proteins, human IgG antibodies to HSV2 at about 20 arbU/ml ± 20%, 0.2 mg/ml gentamicine sulphate and 0.045% ProClin 300 as preservatives.

Note: The volume necessary to dissolve the content of the vial may vary from lot to lot. Please use the right volume reported on the label.

4. Wash buffer concentrate: WASHBUF 20X

1x60ml/bottle20x concentrated solution. Once diluted, the wash solution contains 10 mM phosphate buffer pH 7.0+/-0.2, 0.05% Tween 20 and 0.045% ProClin 300.

5. Enzyme conjugate : CONJ

2x8ml/vial. Ready to use and red colour coded. It contains Horseradish peroxidase conjugated polyclonal antibodies to human IgG, 5% BSA, 10 mM Tris buffer pH 6.8+/-0.1, 0.045% ProClin 300, 0.02% gentamicine sulphate as preservatives and 0.01% red alimentary dye.

6. Chromogen/Substrate: SUBS TMB

1x16ml/vial. It contains 50 mM citrate-phosphate buffer pH 3.5-3.8, 4% dimethylsulphoxide, 0.03% tetra-methyl-benzidine (or TMB) and 0.02% hydrogen peroxide (or H₂O₂).

Note: To be stored protected from light as sensitive to strong illumination.

7. Sulphuric Acid: H₂SO₄ 0.3 M

1x15ml/vial. It contains 0.3 M H₂SO₄ solution.

Attention: Irritant (H315, H319; P280, P302+P352, 332+P313, P305+ P351+P338, P337+P313, P362+P363)

8. Specimen Diluent: DILSPE

2x60ml/vial. It contains 2% casein, 10 mM Na-citrate buffer pH 6.0 +/-0.1, 0.1% Tween 20, 0.09% Na-azide 0.1% and 0.045% ProClin 300 as preservatives. The reagent is blue colour coded.

9. Plate sealing foils n°2

10. Package insert n°1

E. MATERIALS REQUIRED BUT NOT PROVIDED

1. Calibrated Micropipettes (1000 ul, 100 ul and 10 ul) and disposable plastic tips.
2. EIA grade water (double distilled or deionised, charcoal treated to remove oxidizing chemicals used as disinfectants).
3. Timer with 60 minute range or higher.
4. Absorbent paper tissues.
5. Calibrated ELISA microplate thermostatic incubator (dry or wet), set at +37°C (+/-0.5°C tolerance)..
6. Calibrated ELISA microwell reader with 450nm (reading) and with 620-630nm (blanking) filters.
7. Calibrated ELISA microplate washer.
8. Vortex or similar mixing tools.

F. WARNINGS AND PRECAUTIONS

1. The kit has to be used by skilled and properly trained technical personnel only, under the supervision of a medical doctor responsible of the laboratory.
2. All the personnel involved in performing the assay have to wear protective laboratory clothes, talc-free gloves and glasses.

The use of any sharp (needles) or cutting (blades) devices should be avoided. All the personnel involved should be trained in biosafety procedures, as recommended by the Center for Disease Control, Atlanta, U.S. and reported in the National Institute of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.

3. All the personnel involved in sample handling should be vaccinated for HBV and HAV, for which vaccines are available, safe and effective.

4. The laboratory environment should be controlled so as to avoid contaminants such as dust or air-borne microbial agents, when opening kit vials and microplates and when performing the test. Protect the Chromogen (TMB) from strong light and avoid vibration of the bench surface where the test is undertaken.

5. Upon receipt, store the kit at 2..8°C into a temperature controlled refrigerator or cold room.

6. Do not interchange components between different lots of the kits. It is recommended that components between two kits of the same lot should not be interchanged.

7. Check that the reagents are clear and do not contain visible heavy particles or aggregates. If not, advise the laboratory supervisor to initiate the necessary procedures for kit replacement.

8. Avoid cross-contamination between serum/plasma samples by using disposable tips and changing them after each sample. Do not reuse disposable tips.

9. Avoid cross-contamination between kit reagents by using disposable tips and changing them between the use of each one. Do not reuse disposable tips.

10. Do not use the kit after the expiration date stated on the external container and internal (vials) labels. A study conducted on an opened kit did not pointed out any relevant loss of activity up to six 6 uses of the device and up to 3 months.

11. Treat all specimens as potentially infective. All human serum specimens should be handled at Biosafety Level 2, as recommended by the Center for Disease Control, Atlanta, U.S. in compliance with what reported in the Institutes of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.

12. The use of disposable plastic-ware is recommended in the preparation of the liquid components or in transferring components into automated workstations, in order to avoid cross contamination.

13. Waste produced during the use of the kit has to be discarded in compliance with national directives and laws concerning laboratory waste of chemical and biological substances. In particular, liquid waste generated from the washing procedure, from residuals of controls and from samples has to be treated as potentially infective material and inactivated before waste. Suggested procedures of inactivation are treatment with a 10% final concentration of household bleach for 16-18 hrs or heat inactivation by autoclave at 121°C for 20 min..

14. Accidental spills from samples and operations have to be adsorbed with paper tissues soaked with household bleach and then with water. Tissues should then be discarded in proper containers designated for laboratory/hospital waste.

15. The Sulphuric Acid is an irritant. In case of spills, wash the surface with plenty of water

16. Other waste materials generated from the use of the kit (example: tips used for samples and controls, used microplates) should be handled as potentially infective and disposed according to national directives and laws concerning laboratory wastes.

G. SPECIMEN: PREPARATION AND WARNINGS

1. Blood is drawn aseptically by venepuncture and plasma or serum is prepared using standard techniques of preparation of samples for clinical laboratory analysis. No influence has been observed in the preparation of the sample with citrate, EDTA and heparin.

2. Samples have to be clearly identified with codes or names in order to avoid misinterpretation of results. Bar code labeling and electronic reading is strongly recommended.

3. Haemolysed ("red") and visibly hyperlipemic ("milky") samples have to be discarded as they could generate false results. Samples containing residues of fibrin or heavy particles or microbial filaments and bodies should be discarded as they could give rise to false results.

4. Sera and plasma can be stored at +2°...+8°C in primary collection tubes for up to five days after collection.

Do not freeze primary tubes of collection. For longer storage periods, sera and plasma samples, carefully removed from the primary collection tube, can be stored frozen at -20°C for several months. Any frozen samples should not be frozen/thawed more than once as this may generate particles that could affect the test result.

5. If particles are present, centrifuge at 2.000 rpm for 20 min or filter using 0.2-0.8u filters to clean up the sample for testing.

H. PREPARATION OF COMPONENTS AND WARNINGS

A study conducted on an opened kit has not pointed out any relevant loss of activity up to 6 re-uses of the device and up to 3 months.

Microplate:

Allow the microplate to reach room temperature (about 1 hr) before opening the container. Check that the desiccant has not turned dark green, indicating a defect in storage.

In this case, call Dia.Pro's customer service.

Unused strips have to be placed back into the aluminum pouch, with the desiccant supplied, firmly zipped and stored at +2°..8°C.

After first opening, remaining strips are stable until the humidity indicator inside the desiccant bag turns from yellow to green.

Calibration Curve

Ready to use component. Mix carefully on vortex before use.

Control Serum

Add the volume of ELISA grade water, reported on the label, to the lyophilised powder; let fully dissolve and then gently mix on vortex.

Note: The control after dissolution is not stable. Store frozen in aliquots at -20°C.

Wash buffer concentrate:

The whole content of the concentrated solution has to be diluted 20x with bidistilled water and mixed gently end-over-end before use. During preparation avoid foaming as the presence of bubbles could impact on the efficiency of the washing cycles.

Note: Once diluted, the wash solution is stable for 1 week at +2..8° C.

Enzyme conjugate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

If this component has to be transferred use only plastic, possibly sterile disposable containers.

Chromogen/Substrate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

Do not expose to strong illumination, oxidizing agents and metallic surfaces.

If this component has to be transferred use only plastic, possible sterile disposable container

Sample Diluent

Ready to use component. Mix carefully on vortex before use.

Sulphuric Acid:

Ready to use. Mix well on vortex before use.

Attention: Irritant (H315, H319; P280, P302+P352, 332+P313, P305+ P351+P338, P337+P313, P362+P363).

Legenda:

Warning **H** statements:

H315 – Causes skin irritation.

H319 – Causes serious eye irritation.

Precautionary **P** statements:

P280 – Wear protective gloves/protective clothing/eye protection/face protection.

P302 + P352 – IF ON SKIN: Wash with plenty of soap and water.

P332 + P313 – If skin irritation occurs: Get medical advice/attention.

P305 + P351 + P338 – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P337 + P313 – If eye irritation persists: Get medical advice/attention.

P362 + P363 - Take off contaminated clothing and wash it before reuse.

I. INSTRUMENTS AND TOOLS USED IN COMBINATION WITH THE KIT

1. Micropipettes have to be calibrated to deliver the correct volume required by the assay and must be submitted to regular decontamination (household alcohol, 10% solution of bleach, hospital grade disinfectants) of those parts that could accidentally come in contact with the sample. They should also be regularly maintained in order to show a precision of 1% and a trueness of +/-2%. Decontamination of spills or residues of kit components should also be carried out regularly.
2. The ELISA incubator has to be set at +37°C (tolerance of +/- 0.5°C) and regularly checked to ensure the correct temperature is maintained. Both dry incubators and water baths are suitable for the incubations, provided that the instrument is validated for the incubation of ELISA tests.
3. The **ELISA washer** is extremely important to the overall performances of the assay. The washer must be carefully validated in advance, checked for the delivery of the right dispensation volume and regularly submitted to maintenance according to the manufacturer's instructions for use. In particular the washer, at the end of the daily workload, has to be extensively cleaned out of salts with deionized water. Before use, the washer has to be extensively primed with the diluted Washing Solution. The instrument weekly has to be submitted to decontamination according to its manual (NaOH 0.1 M decontamination suggested).
5 washing cycles (aspiration + dispensation of 350ul/well of washing solution + 20 sec soaking = 1 cycle) are sufficient to ensure the assay with the declared performances. If soaking is not possible add one more cycle of washing.
An incorrect washing cycle or salt-blocked needles are the major cause of false positive reactions.
4. Incubation times have a tolerance of ±5%.
5. The ELISA microplate reader has to be equipped with a reading filter of 450nm and with a second filter (620-630nm, mandatory) for blanking purposes. Its standard performances should be (a) bandwidth ≤ 10 nm; (b) absorbance range from 0 to ≥ 2.0; (c) linearity to ≥ 2.0; repeatability ≥ 1%. Blanking is carried out on the well identified in the section "Assay Procedure". The optical system of the reader has to be calibrated regularly to ensure that the correct optical density is measured. It should be regularly maintained according to the manufacturer 's instructions.
6. When using an ELISA automated work station, all critical steps (dispensation, incubation, washing, reading, data handling) have to be carefully set, calibrated, controlled and regularly serviced in order to match the values reported in the section "Internal Quality Control". The assay protocol has to be installed in the operating system of the unit and

validated as for the washer and the reader. In addition, the liquid handling part of the station (dispensation and washing) has to be validated and correctly set. Particular attention must be paid to avoid carry over by the needles used for dispensing and for washing. This must be studied and controlled to minimize the possibility of contamination of adjacent wells. The use of ELISA automated work stations is recommended when the number of samples to be tested exceed 20-30 units per run.

7. Dia.Pro's customer service offers support to the user in the setting and checking of instruments used in combination with the kit, in order to assure compliance with the requirements described. Support is also provided for the installation of new instruments to be used with the kit.

L. PRE ASSAY CONTROLS AND OPERATIONS

1. Check the expiration date of the kit printed on the external label (primary container). Do not use if expired.
2. Check that the liquid components are not contaminated by visible particles or aggregates.
3. Check that the Chromogen (TMB) is colourless or pale blue by aspirating a small volume of it with a sterile plastic pipette.
4. Check that no breakage occurred in transportation and no spillage of liquid is present inside the box (primary container). Check that the aluminium pouch, containing the microplate, is not punctured or damaged.
5. Dissolve the content of the lyophilised Control Serum as reported in the proper section.
6. Dilute all the content of the 20x concentrated Wash Solution as described above.
7. Allow all the other components to reach room temperature (about 1 hr) and then mix gently on vortex all liquid reagents.
8. Set the ELISA incubator at +37°C and prepare the ELISA washer by priming with the diluted washing solution, according to the manufacturers instructions. Set the right number of washing cycles as reported in the specific section.
9. Check that the ELISA reader is turned on or ensure it will be turned on at least 20 minutes before reading.
10. If using an automated work station, turn on, check settings and be sure to use the right assay protocol.
11. Check that the micropipettes are set to the required volume.
12. Check that all the other equipment is available and ready to use.
13. In case of problems, do not proceed further with the test and advise the supervisor.

M. ASSAY PROCEDURE

The assay has to be carried out according to what reported below, taking care to maintain the same incubation time for all the samples in testing.

The kit may be used for quantitative and qualitative determinations as well.

M1. QUANTITATIVE DETERMINATION:

Automated assay:

In case the test is carried out automatically with an ELISA system, we suggest to make the instrument aspirate 1000 µl Sample Diluent and then 10 µl sample (1:101 dilution factor). The whole content is then dispensed into a properly defined dilution tube. Before the next sample is aspirated, needles have to be duly washed to avoid any cross-contamination among samples. When all the samples have been diluted make the instrument dispense 100 µl samples into the proper wells of the microplate.

This procedure may be carried out also in two steps of dilutions of 1:10 each (90 µl Sample Diluent + 10 µl sample) into a second dilution platform. Make then the instrument aspirate first

100 µl Sample Diluent, then 10 µl liquid from the first dilution in the platform and finally dispense the whole content in the proper well of the assay microplate.

Do not dilute Calibrators and the dissolved Control Serum as they are ready to use.

Dispense 100 µl calibrators/control in the appropriate calibration/control wells.

For the next operations follow the operative instructions reported below for the Manual Assay.

It is strongly recommended to check that the time lap between the dispensation of the first and the last sample will be calculated by the instrument and taken into consideration by delaying the first washing operation accordingly.

Manual assay:

1. Dilute samples 1:101 into a properly defined dilution tube (example: 1000 µl Sample Diluent + 10 µl sample). Do not dilute the Calibration Set as calibrators are ready to use. Mix carefully all the liquid components on vortex and then proceed as described below.
2. Place the required number of microwells in the microwell holder. Leave the A1 and B1 empty for the operation of blanking.
3. Dispense 100 µl of Calibrators and 100 µl Control Serum in duplicate. Then dispense 100 µl of diluted samples in each properly identified well.
4. Incubate the microplate for **60 min at +37°C**.

Important note: Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

5. Wash the microplate with an automatic washer as reported previously (section I.3).
6. Pipette 100 µl Enzyme Conjugate into each well, except A1+B1 blanking wells, and cover with the sealer. Check that this red coloured component has been dispensed in all the wells, except A1 and B1.

Important note: Be careful not to touch the plastic inner surface of the well with the tip filled with the Enzyme Conjugate. Contamination might occur.

7. Incubate the microplate for **60 min at +37°C**.
8. Wash microwells as in step 5.
9. Pipette 100 µl Chromogen/Substrate mixture into each well, the blank wells A1 and B1 included. Then incubate the microplate at **room temperature (18-24°C) for 20 minutes**.

Important note: Do not expose to strong direct illumination. High background might be generated.

10. Pipette 100 µl Sulphuric Acid to stop the enzymatic reaction into all the wells using the same pipetting sequence as in step 9. Addition of acid will turn the positive calibrators, the control serum and the positive samples from blue to yellow.
11. Measure the colour intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and at 620-630nm (background subtraction, mandatory), blanking the instrument on A1 or B1 or both.

M2. QUALITATIVE DETERMINATION

If only a qualitative determination is required, proceed as described below:

Automated assay:

Proceed as described in section M1.

Manual assay:

1. Dilute samples 1:101 into a properly defined dilution tube (example: 1000 µl Sample Diluent + 10 µl sample). Do not dilute the Calibration Set as calibrators are ready to use. Mix carefully all the liquid components on vortex and then proceed as described below.
2. Place the required number of Microwells in the microwell holder. Leave A1 well empty for the operation of blanking.
3. Dispense 100 µl of Calibrator 0 arbU/ml and Calibrator 5 arbU/ml in duplicate and Calibrator 100 arbU/ml in single. Then dispense 100 µl of diluted samples in each properly identified well.
4. Incubate the microplate for **60 min at +37°C**.

Important note: Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

5. Wash the microplate with an automatic washer as reported previously (section I.3).
6. Pipette 100 µl Enzyme Conjugate into each well, except the A1 well, and cover with the sealer. Check that this red coloured component has been dispensed in all the wells, except A1.

Important note: Be careful not to touch the plastic inner surface of the well with the tip filled with the Enzyme Conjugate. Contamination might occur.

7. Incubate the microplate for **60 min at +37°C**.
8. Wash microwells as in step 5.
9. Pipette 100 µl Chromogen/Substrate mixture into each well, the blank well included. Then incubate the microplate at **room temperature (18-24°C) for 20 minutes**.

Important note: Do not expose to strong direct illumination. High background might be generated.

10. Pipette 100 µl Sulphuric Acid into all the wells using the same pipetting sequence as in step 9. Addition of acid will turn the positive calibrators, the control serum and the positive samples from blue to yellow.
11. Measure the colour intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and at 620-630nm (background subtraction, mandatory), blanking the instrument on A1.

General Important notes:

1. Ensure that no finger prints are present on the bottom of the microwell before reading. Finger prints could generate false positive results on reading.
2. Reading has to be carried out just after the addition of the Stop Solution and anyway not any longer than 20 minutes after its addition. Some self oxidation of the chromogen can occur leading to high background.

N. ASSAY SCHEME

Method	Operations
Calibrators & Control (*)	100 µl
Samples diluted 1:101	100 µl
1 st incubation	60 min
Temperature	+37°C
Wash step	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
Enzyme conjugate	100 µl
2 nd incubation	60 min
Temperature	+37°C
Wash step	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
TMB/H ₂ O ₂	100 µl
3 rd incubation	20 min
Temperature	r.t.
Sulphuric Acid	100 µl
Reading OD	450nm / 620-630nm

(*) Important Notes:

- The Control Serum (CS) it does not affect the test's results calculation.
- The Control Serum (CS) used only if a laboratory internal quality control is required by the Management.

An example of dispensation scheme for Quantitative Analysis is reported below:

Microplate

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	CAL4	S 1									
B	BLK	CAL4	S 2									
C	CAL1	CAL5	S 3									
D	CAL1	CAL5	S 4									
E	CAL2	CAL6	S 5									
F	CAL2	CAL6	S 6									
G	CAL3	CS(*)	S 7									
H	CAL3	CS(*)	S 8									

Legenda: BLK = Blank CAL = Calibrator
CS(*) = Control Serum - Not mandatory S = Sample

An example of dispensation scheme in qualitative assays is reported below:

Microplate

	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	S 2	S 10									
B	CAL1	S 3	S 11									
C	CAL1	S 4	S 12									
D	CAL2	S 5	S 13									
E	CAL2	S 6	S 14									
F	CAL6	S 7	S 15									
G	S 1	S 8	S 16									
H	S 2	S 9	S 17									

Legenda: BLK = Blank CAL = Calibrators
S = Sample

O. INTERNAL QUALITY CONTROL

A validation check is carried out on the calibrators any time the kit is used in order to verify whether the performances of the assay are as qualified.

Control that the following data are matched:

Check	Requirements
Blank well	< 0.050 OD450nm value
CAL 1 0 arbU/ml	< 0.150 mean OD450nm value after blanking coefficient of variation < 30%
CAL 2 5 arbU/ml	OD450nm ≥ OD450nm CAL1 + 0.100
CAL 6 100 arbU/ml	OD450nm ≥ 1.000

If the results of the test match the requirements stated above, proceed to the next section.

If they do not, do not proceed any further and operate as follows:

Problem	Check
Blank well > 0.050 OD450nm	1. that the Chromogen/Sustrate solution has not got contaminated during the assay
CAL 1 0 arbU/ml > 0.150 OD450nm after blanking coefficient of variation > 30%	1. that the washing procedure and the washer settings are as validated in the pre qualification study; 2. that the proper washing solution has been used and the washer has been primed with it before use; 3. that no mistake has been done in the assay procedure (dispensation of a positive calibrator instead of the negative one); 4. that no contamination of the negative calibrator or of their wells has occurred due spills of positive samples or the enzyme conjugate; 5. that micropipettes haven't got contaminated with positive samples or with the enzyme conjugate 6. that the washer needles are not blocked or partially obstructed.
CAL 2 5 arbU/ml OD450nm ≤ OD450nm CAL1 + 0.100	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (ex.: dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the calibrator has occurred.
CAL 6 100 arbU/ml ≤ 1.000 OD450nm	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the positive control has occurred.

Should one of these problems have happened, after checking, report to the supervisor for further actions.

** Note:

If Control Serum has used, verify the following data:

Check	Requirements
Control Serum	Mean OD450nm CAL 4 $\pm$ 20%

If the results of the test doesn't match the requirements stated above, operate as follows:

Problem	Check
Control Serum Different from expected value	1. that the procedure has been correctly executed; 2. that no mistake has been done in its distribution (dispensation of a wrong calibrator instead); 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the control serum has occurred.

Anyway, if all other parameters (Blank, CAL1, CAL2, CAL6), match the established requirements, the test may be considered valid.

P. RESULTS


P.1 Quantitative method

If the test turns out to be valid, use for the quantitative method an approved curve fitting program to draw the calibration curve from the values obtained by reading at 450nm (4-parameters interpolation is suggested).

Then on the calibration curve calculate the concentration of anti Herpes Simplex Virus type 2 IgG antibody in samples.

An example of Calibration curve is reported in the next page.

Example of Calibration Curve :



Important Note:

Do not use the calibration curve above to make calculations.

P.2 Qualitative method

In the qualitative method, calculate the mean OD450nm values for the Calibrators 0 and 5 arbU/ml and then check that the assay is valid.

Example of calculation:

The following data must not be used instead of real figures obtained by the user.

Calibrator 0 arbU/ml: 0.020 – 0.024 OD450nm
Mean Value: 0.022 OD450nm
Lower than 0.150 – Accepted

Calibrator 5 arbU/ml: 0.350 – 0.370 OD450nm
Mean Value: 0.360 OD450nm
Higher than Cal 0 + 0.100 – Accepted

Calibrator 100 arbU/ml: 2.245 OD450nm
Higher than 1.000 – Accepted

Q. INTERPRETATION OF RESULTS

Samples with a concentration lower than 5 arbU/ml are considered negative for anti HSV2 IgG antibody.

Samples with a concentration higher than 5 arbU/ml are considered positive for anti HSV2 IgG antibody.

Particular attention in the interpretation of results has to be used in the follow-up of pregnancy for a primary infection of HSV due to the risk of neonatal malformations.

Important notes:

1. Interpretation of results should be done under the supervision of the laboratory supervisor to reduce the risk of judgment errors and misinterpretations.
2. When test results are transmitted from the laboratory to another facility, attention must be paid to avoid erroneous data transfer.
3. In the follow-up of pregnancy for HSV infection a positive result (presence of IgG antibody > 5 arbU/ml) should be confirmed to ruled out the risk of a false positive result and a false definition of protection.

R. PERFORMANCES

1. Limit of detection

The limit of detection of the assay has been calculated by means of an internal Gold Standard in absence of an international preparation to refer to.

The limit of detection has been calculated as mean OD450nm Calibrator 0 arbU/ml + 5 SD.

The table below reports the mean OD450nm values of this standard when diluted in negative plasma and then examined in the assay for three lots.

Mean OD450nm values (n = 2)

IgG arbU/ml	HSV2G.PU Lot # 1203	HSV2G.PU Lot # 1103	HSV2G Lot # 0304/2
0	0.022	0.030	0.014
5	0.353	0.384	0.269
10	0.596	0.606	0.557
20	1.169	1.471	0.895
50	2.030	2.276	1.776
100	3.102	3.353	2.893

The assay shows a limit of detection far better than 5 arbU/ml.

2. Diagnostic sensitivity:

The diagnostic sensitivity has been tested in a performance evaluation study on panels of samples classified positive by a kit US FDA approved. Positive samples from different stage of HSV infection were tested. The value, obtained from the analysis of more than 300 specimens, has been $\geq 98\%$.

3. Diagnostic specificity:

The diagnostic specificity has been determined on panels of negative samples from not infected individuals, classified negative with a kit US FDA approved.

Both plasma, derived with different standard techniques of preparation (citrate, EDTA and heparin), and sera have been used to determine the value of specificity.

Frozen specimens have been tested, as well, to check for interferences due to collection and storage.

No interference was observed.

Potentially interfering samples derived from patients with different pathologies (mostly ANA, AMA and RF positive) and from pregnant women were tested.

No crossreaction was observed.

An overall value $> 98\%$ of specificity was found when examined on more than 100 specimens.

3. Precision:

It has been calculated on the Calibrator 5 arbU/ml, considered the cut-off of the assay, examined in 16 replicates in three separate runs for three lots.

Results are reported as follows:

HSV2G.CE: lot 1004

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.286	0.303	0.256	0.282
Std.Deviation	0.022	0.037	0.020	0.026
CV %	7.7	12.4	7.74	9.28

HSV2G.PU: lot 1103

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.375	0.384	0.394	0.384
Std.Deviation	0.019	0.022	0.015	0.019
CV %	5.07	5.73	3.81	4.87

HSV2G.PU: lot 1203

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.352	0.345	0.332	0.343
Std.Deviation	0.017	0.020	0.024	0.020
CV %	4.83	5.78	7.23	5.95

The variability shown in the tables above did not result in sample misclassification.

S. LIMITATIONS OF THE PROCEDURE

Bacterial contamination or heat inactivation of the specimen may affect the absorbance values of the samples with consequent alteration of the level of the analyte.

Frozen samples containing fibrin particles or aggregates after thawing may generate some false results.

This test is suitable only for testing single samples and not pooled ones.

Diagnosis of an infectious disease should not be established on the basis of a single test result. The patient's clinical history,

symptomatology, as well as other diagnostic data should be considered.

REFERENCES

1. Engvall E. and Perlmann P.. J.Immunochemistry 8: 871-874, 1971
2. Engvall E. and Perlmann P.. J.Immunol.. 109: 129-135, 1971
3. Remington J.S. and Klein J.O.. (1996) In "Infectious diseases of fetus and newborn infant". Sanders, Philadelphia, London, Toronto.
4. Volk W.A. (1982) In "essential of Medical Microbiology". 2nd ed., pp 729, G.B. Lippincott Co. Philadelphia, New York, S.José, Toronto.
5. Leinikki P.O. et al.. J.Clin.Microbiol.. 8:418, 1978
6. Piroid E. et al.. Révue Méd.Vet.. 131:25, 1980.
7. Vaheri A. et al.. J.Med.Virol.. 5:171, 1980.
8. Vejtorp M. et al.. Acta Path.Microbiol.Scand.. 88:349, 1980.
9. Voller A. et al.. Brit.J.Exp.Pathol.. 56:338, 1975.

All the IVD Products manufactured by the company are under the control of a certified Quality Management System in compliance with ISO 13485 rule. Each lot is submitted to a quality control and released into the market only if conforming with the EC technical specifications and acceptance criteria.

Manufacturer:

Dia.Pro Diagnostic Bioprobes S.r.l.
Via G. Carducci n° 27 – Sesto San Giovanni (MI) – Italy





Dia.Pro
Diagnostic
Bio**Probes**

EC DECLARATION OF CONFORMITY

MANUFACTURER	DIA.PRO DIAGNOSTIC BIOPROBES S.R.L. VIA G. CARDUCCI N° 27 – 20099 SESTO SAN GIOVANNI (MILANO) – ITALY
PRODUCT	HSV2 IgG CODE: HSV2G.CE (96 tests)
CLASSIFICATION	GENERAL IVD
CONFORMITY ASSESSMENT ROUTE	SELF CERTIFICATION

**WE HEREBY DECLARE THAT THE ABOVE MENTIONED PRODUCT MEETS
THE PROVISIONS OF THE COUNCIL DIRECTIVE 98/79/EC
FOR IN VITRO DIAGNOSTIC DEVICES.**

ISO CERTIFICATE	UNE EN ISO 13485 N° 2013 11 0039 EN, RELEASED BY AEMPS (AGENCIA ESPAÑOLA DE MEDICAMENTOS Y PRODUCTOS SANITARIOS)
------------------------	---

PLACE & DATE OF FIRST ISSUE	MILANO – OCTOBER 2004
PLACE & DATE OF CURRENT ISSUE	SESTO SAN GIOVANNI (MI) – MARCH 2019
SIGNATURE Legal Representative Dr.ssa Fiorenza Scozzesi	

Rev: 05/2018

HSV2 IgM

**“Capture” Enzyme Immuno Assay
(ELISA) for the determination
of IgM antibodies to
Herpes Simplex Virus type 2
in human plasma and sera**

- for “in vitro” diagnostic use only -



DIA.PRO

**Diagnostic Bioprobes Srl
Via G. Carducci n° 27
20099 Sesto San Giovanni
(Milano) - Italy**

Phone +39 02 27007161

Fax +39 02 44386771

e-mail: info@diapro.it

REF HSV2M.CE
96 tests

HSV2 IgM

A. INTENDED USE

Enzyme ImmunoAssay (ELISA) for the determination of IgM antibodies to Herpes Simplex Virus types 2 in human plasma and sera with the "capture" system. The device is intended for the follow-up of HSV2 infected patients and for the monitoring of risk of neonatal defects due to HSV infection during pregnancy.
For "in vitro" diagnostic use only.

B. INTRODUCTION

Herpes Simplex Virus type 1 (HSV1) and type 2 (HSV2) are large complex DNA-containing viruses which have been shown to induce the synthesis of several proteins during infection, possessing an high number of cross-reactive determinants and just a few of type-specific sequences.

The majority of primary and recurrent genital herpetic infections are caused by HSV2; while non genital infections, such as common cold sores, are caused primarily by HSV1.

The detection of virus specific IgG and IgM antibodies are important in the diagnosis of acute/primary virus infections or reactivations of a latent one, in the absence of evident clinical symptoms.

A-symptomatic infections may happen for HSV in apparently healthy individuals and during pregnancy. Severe herpetic infections may happen in immuno-compromised and suppressed patients in which the disease may evolve toward critical pathologies.

The determination of HSV specific antibodies has then become important in the monitoring of "risk" patients and in the follow up of acute and severe infections.

C. PRINCIPLE OF THE TEST

The assay is based on the principle of "IgM capture" where IgM class antibodies in the sample are first captured by the solid phase coated with anti hIgM antibody.

After washing out all the other components of the sample and in particular IgG antibodies, the specific IgM captured on the solid phase are detected by the addition of a preparation of inactivated HSV2, labeled with a HSV2 specific antibody conjugated with peroxidase (HRP).

After incubation, microwells are washed to remove unbound conjugate and then the chromogen/substrate is added.

In the presence of bound conjugate the colorless substrate is hydrolyzed to a colored end-product, whose optical density may be detected and is proportional to the amount of IgM antibodies to HSV2 present in the sample.

A system is described how to control whether the positivity shown by a sample is true or not (Confirmation Test), helpful for the clinician to make a correct interpretation of results.

D. COMPONENTS

The kit contains reagents for 96 tests.

1. Microplate: MICROPLATE

12 strips x 8 microwells coated with anti human IgM affinity purified goat antibody, in presence of bovine proteins. Plates are sealed into a bag with desiccant. Allow the microplate to reach room temperature before opening; reseal unused strips in the bag with desiccant and store at 2..8°C.

2. Negative Control: CONTROL -

1x4.0 ml/vial. Ready to use control. It contains 1% human serum proteins, 2% casein, 10 mM tris buffer pH 6.0+/-0.1, 0.1%

Tween 20, 0.09% sodium azide and 0.045% ProClin 300 as preservatives.

The negative control is pale yellow color coded.

3. Positive Control: CONTROL +

1x4.0 ml/vial. Ready to use control. It contains 1% human serum positive for HSV2 IgM, 2% casein, 10 mM tris buffer pH 6.0+/-0.1, 0.1% Tween 20, 0.09% sodium azide and 0.045% ProClin 300 as preservatives.

The positive control is green colour coded.

4. Calibrator: CAL ...ml

N° 1 lyophilized vial. To be dissolved with EIA grade water as reported in the label. It contains anti HSV2 IgM, fetal bovine serum, 0.2 mg/ml gentamicine sulphate and 0.045% ProClin 300 as preservatives.

Note: The volume necessary to dissolve the content of the vial may vary from lot to lot. Please use the right volume reported on the label.

5. Lyophilized HSV2 Ag: AG HSV2

N° 6 lyophilized vials. The vials contain lyophilized gamma-ray inactivated HSV2 in protein buffer. The solution contains 2% bovine proteins, 10 mM Tris HCl buffer pH 6.8+/-0.1, 0.2 mg/ml gentamicine sulphate and 0.045% ProClin 300. To be dissolved with 1.9 ml of Antigen Diluent as reported in the specific section.

6. Wash buffer concentrate: WASHBUF 20X

1x60ml/bottle. 20x concentrated solution. Once diluted, the wash solution contains 10 mM phosphate buffer pH 7.0+/-0.2, 0.05% Tween 20 and 0.045% ProClin 300.

7. Enzyme conjugate: CONJ 20X

1x0.8 ml/vial. 20x concentrated solution of a HSV2-specific antibody, labeled with HRP and diluted in a protein buffer containing 10 mM Tris buffer pH 6.8+/-0.1, 2% BSA, 0.045% ProClin 300 and 0.2 mg/ml gentamicine sulphate as preservatives.

8. Antigen Diluent : AG DIL

n° 1 vial of 16 ml. Protein buffer solution for the preparation of the Immunocomplex. The solution contains 10 mM Tris buffer pH 6.8+/-0.1, 2% BSA, 0.045% ProClin 300 and 0.2 mg/ml gentamicine sulphate as preservatives. The reagent is code coloured with 0.01% red alimentary dye

9. Specimen Diluent : DILSPE

2x60.0 ml/vial. Proteic buffered solution for the dilution of samples. It contains 2% casein, 10 mM tris buffer pH 6.0+/-0.1, 0.1% Tween 20, 0.09% sodium azide and 0.045% ProClin 300 as preservatives.

The reagent is color coded with 0.01% blue alimentary dye.

10. Chromogen/Substrate : SUBS TMB

1x16ml/vial. It contains a 50 mM citrate-phosphate buffered solution at pH 3.5-3.8, 0.03% tetra-methyl-benzidine (TMB), 0.02% hydrogen peroxide (H₂O₂) and 4% dimethylsulphoxide.

Note: To be stored protected from light as sensitive to strong illumination.

11. Sulphuric Acid: H₂SO₄ 0.3 M

1x15ml/vial. It contains 0.3 M H₂SO₄ solution.
Attention: Irritant (H315, H319; P280, P302+P352, 332+P313, P305+ P351+P338, P337+P313, P362+P363)

12. Plate sealing foils n° 2

13. Package insert n° 1

E. MATERIALS REQUIRED BUT NOT PROVIDED

1. Calibrated Micropipettes (1000 ul, 100 ul and 10 ul) and disposable plastic tips.
2. EIA grade water (double distilled or deionised, charcoal treated to remove oxidizing chemicals used as disinfectants).
3. Timer with 60 minute range or higher.
4. Absorbent paper tissues.
5. Calibrated ELISA microplate thermostatic incubator (dry or wet), set at +37°C (+/-0.5°C tolerance).
6. Calibrated ELISA microwell reader with 450nm (reading) and with 620-630nm (blanking) filters.
7. Calibrated ELISA microplate washer.
8. Vortex or similar mixing tools.

F. WARNINGS AND PRECAUTIONS

1. The kit has to be used by skilled and properly trained technical personnel only, under the supervision of a medical doctor responsible of the laboratory.
2. All the personnel involved in performing the assay have to wear protective laboratory clothes, talc-free gloves and glasses. The use of any sharp (needles) or cutting (blades) devices should be avoided. All the personnel involved should be trained in biosafety procedures, as recommended by the Center for Disease Control, Atlanta, U.S. and reported in the National Institute of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
3. All the personnel involved in sample handling should be vaccinated for HBV and HAV, for which vaccines are available, safe and effective.
4. The laboratory environment should be controlled so as to avoid contaminants such as dust or air-borne microbial agents, when opening kit vials and microplates and when performing the test. Protect the Chromogen (TMB) from strong light and avoid vibration of the bench surface where the test is undertaken.
5. Upon receipt, store the kit at 2..8°C into a temperature controlled refrigerator or cold room.
6. Do not interchange components between different lots of the kits. It is recommended that components between two kits of the same lot should not be interchanged.
7. Check that the reagents are clear and do not contain visible heavy particles or aggregates. If not, advise the laboratory supervisor to initiate the necessary procedures for kit replacement.
8. Avoid cross-contamination between serum/plasma samples by using disposable tips and changing them after each sample. Do not reuse disposable tips.
9. Avoid cross-contamination between kit reagents by using disposable tips and changing them between the use of each one. Do not reuse disposable tips.
10. Do not use the kit after the expiration date stated on the external container and internal (vials) labels. A study conducted on an opened kit did not pointed out any relevant loss of activity up to six 6 uses of the device and up to 3 months.
11. Treat all specimens as potentially infective. All human serum specimens should be handled at Biosafety Level 2, as recommended by the Center for Disease Control, Atlanta, U.S. in compliance with what reported in the Institutes of Health's publication: "Biosafety in Microbiological and Biomedical Laboratories", ed. 1984.
12. The use of disposable plastic-ware is recommended in the preparation of the liquid components or in transferring components into automated workstations, in order to avoid cross contamination.
13. Waste produced during the use of the kit has to be discarded in compliance with national directives and laws concerning laboratory waste of chemical and biological substances. In particular, liquid waste generated from the washing procedure, from residuals of controls and from samples has to be treated as potentially infective material and inactivated before waste. Suggested procedures of inactivation are

treatment with a 10% final concentration of household bleach for 16-18 hrs or heat inactivation by autoclave at 121°C for 20 min..

14. Accidental spills from samples and operations have to be adsorbed with paper tissues soaked with household bleach and then with water. Tissues should then be discarded in proper containers designated for laboratory/hospital waste.

15. The Sulphuric Acid is an irritant. In case of spills, wash the surface with plenty of water

16. Other waste materials generated from the use of the kit (example: tips used for samples and controls, used microplates) should be handled as potentially infective and disposed according to national directives and laws concerning laboratory wastes.

G. SPECIMEN: PREPARATION AND WARNINGS

1. Blood is drawn aseptically by venepuncture and plasma or serum is prepared using standard techniques of preparation of samples for clinical laboratory analysis. No influence has been observed in the preparation of the sample with citrate, EDTA and heparin.
2. Samples have to be clearly identified with codes or names in order to avoid misinterpretation of results. Bar code labeling and electronic reading is strongly recommended.
3. Haemolysed ("red") and visibly hyperlipemic ("milky") samples have to be discarded as they could generate false results. Samples containing residues of fibrin or heavy particles or microbial filaments and bodies should be discarded as they could give rise to false results.
4. Sera and plasma can be stored at +2°...+8°C in primary collection tubes for up to five days after collection. Do not freeze primary tubes of collection. For longer storage periods, sera and plasma samples, carefully removed from the primary collection tube, can be stored frozen at -20°C for at least 12 months. Any frozen samples should not be frozen/thawed more than once as this may generate particles that could affect the test result.
5. If particles are present, centrifuge at 2.000 rpm for 20 min or filter using 0.2-0.8u filters to clean up the sample for testing.

H. PREPARATION OF COMPONENTS AND WARNINGS

A study conducted on an opened kit has not pointed out any relevant loss of activity up to 6 re-uses of the device and up to 3 months.

Microplate:

Allow the microplate to reach room temperature (about 1 hr) before opening the container. Check that the desiccant has not turned dark green, indicating a defect in storing. In this case, call Dia.Pro's customer service.

Unused strips have to be placed back into the aluminum pouch, with the desiccant supplied, firmly zipped and stored at +2°..8°C. After first opening, remaining strips are stable until the humidity indicator inside the desiccant bag turns from yellow to green.

Negative Control:

Ready to use. Mix well on vortex before use.

Positive Control:

Ready to use. Mix well on vortex before use.

Calibrator:

Add the volume of ELISA grade water reported on the label to the lyophilized powder. Let fully dissolve and then gently mix on vortex.

Important Note: The solution is not stable. Store the Calibrator frozen in aliquots at -20°C.

Wash buffer concentrate:

The whole content of the concentrated solution has to be diluted 20x with bidistilled water and mixed gently end-over-end before

use. During preparation avoid foaming as the presence of bubbles could impact on the efficiency of the washing cycles.

Note: Once diluted, the wash solution is stable for 1 week at +2..8° C.

Ag/Ab Immunocomplex:

Proceed carefully as follows:

1. Dissolve the content of a lyophilized vial with 1.9 ml of Conjugate/Antigen Diluent. Let fully dissolved the lyophilized content and then gently mix on vortex.
2. Gently mix the concentrated Enzyme Conjugate on vortex. Then add 0.1 ml of it to the vial of the dissolved HSV2 Ag and mix gently on vortex.

Important Notes:

1. Dissolve and prepare only the number of vials necessary to the test. The Immunocomplex obtained is not stable. Store any residual solution frozen in aliquots at -20°C.
2. The preparation of the Immunocomplex has to be done **right before** the dispensation of samples and controls into the plate. Mix again on vortex gently just before its use.

Specimen Diluent:

Ready to use. Mix well on vortex before use

Chromogen/Substrate:

Ready to use. Mix well on vortex before use.

Be careful not to contaminate the liquid with oxidizing chemicals, air-driven dust or microbes.

Do not expose to strong illumination, oxidizing agents and metallic surfaces.

If this component has to be transferred use only plastic, possible sterile disposable container

Sulphuric Acid:

Ready to use. Mix well on vortex before use.

Attention: Irritant (H315, H319; P280, P302+P352, 332+P313, P305+ P351+P338, P337+P313, P362+P363).

Legenda:

Warning H statements:

H315 – Causes skin irritation.

H319 – Causes serious eye irritation.

Precautionary P statements:

P280 – Wear protective gloves/protective clothing/eye protection/face protection.

P302 + P352 – IF ON SKIN: Wash with plenty of soap and water.

P332 + P313 – If skin irritation occurs: Get medical advice/attention.

P305 + P351 + P338 – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P337 + P313 – If eye irritation persists: Get medical advice/attention.

P362 + P363 – Take off contaminated clothing and wash it before reuse.

I. INSTRUMENTS AND TOOLS USED IN COMBINATION WITH THE KIT

1. Micropipettes have to be calibrated to deliver the correct volume required by the assay and must be submitted to regular decontamination (household alcohol, 10% solution of bleach, hospital grade disinfectants) of those parts that could accidentally come in contact with the sample. They should also be regularly maintained in order to show a precision of 1% and a trueness of +/-2%. Decontamination

of spills or residues of kit components should also be carried out regularly.

2. The ELISA incubator has to be set at +37°C (tolerance of +/- 0.5°C) and regularly checked to ensure the correct temperature is maintained. Both dry incubators and water baths are suitable for the incubations, provided that the instrument is validated for the incubation of ELISA tests.
3. The **ELISA washer** is extremely important to the overall performances of the assay. The washer must be carefully validated in advance, checked for the delivery of the right dispensation volume and regularly submitted to maintenance according to the manufacturer's instructions for use. In particular the washer, at the end of the daily workload, has to be extensively cleaned out of salts with deionized water. Before use, the washer has to be extensively primed with the diluted Washing Solution. The instrument weekly has to be submitted to decontamination according to its manual (NaOH 0.1 M decontamination suggested). 5 washing cycles (aspiration + dispensation of 350ul/well of washing solution + 20 sec soaking = 1 cycle) are sufficient to ensure the assay with the declared performances. If soaking is not possible add one more cycle of washing. An incorrect washing cycle or salt-blocked needles are the major cause of false positive reactions.
4. Incubation times have a tolerance of ±5%.
5. The ELISA microplate reader has to be equipped with a reading filter of 450nm and with a second filter de 620-630nm, mandatory for blanking purposes. Its standard performances should be (a) bandwidth ≤ 10 nm; (b) absorbance range from 0 to ≥ 2.0; (c) linearity to ≥ 2.0; repeatability ≥ 1%. Blanking is carried out on the well identified in the section "Assay Procedure". The optical system of the reader has to be calibrated regularly to ensure that the correct optical density is measured. It should be regularly maintained according to the manufacturer 's instructions.
6. When using an ELISA automated work station, all critical steps (dispensation, incubation, washing, reading, data handling) have to be carefully set, calibrated, controlled and regularly serviced in order to match the values reported in the section "Internal Quality Control". The assay protocol has to be installed in the operating system of the unit and validated as for the washer and the reader. In addition, the liquid handling part of the station (dispensation and washing) has to be validated and correctly set. Particular attention must be paid to avoid carry over by the needles used for dispensing and for washing. This must be studied and controlled to minimize the possibility of contamination of adjacent wells. The use of ELISA automated work stations is recommended when the number of samples to be tested exceed 20-30 units per run.
7. Dia.Pro's customer service offers support to the user in the setting and checking of instruments used in combination with the kit, in order to assure compliance with the requirements described. Support is also provided for the installation of new instruments to be used with the kit.

L. PRE ASSAY CONTROLS AND OPERATIONS

1. Check the expiration date of the kit printed on the external label (primary container). Do not use the device if expired.
2. Check that the liquid components are not contaminated by visible particles or aggregates. Check that the Chromogen/Substrate is colorless or pale blue by aspirating a small volume of it with a sterile plastic pipette. Check that no breakage occurred in transportation and no spillage of liquid is present inside the box (primary container). Check that the aluminum pouch, containing the microplate, is not punctured or damaged.
3. Dilute all the content of the 20x concentrated Wash Solution as described above.

- Dissolve the Calibrator as described above and gently mix.
- Allow all the other components to reach room temperature (about 1 hr) and then mix gently on vortex all liquid reagents.
- Set the ELISA incubator at +37°C and prepare the ELISA washer by priming with the diluted washing solution, according to the manufacturers instructions. Set the right number of washing cycles as reported in the specific section.
- Check that the ELISA reader is turned on or ensure it will be turned on at least 20 minutes before reading.
- If using an automated work station, turn on, check settings and be sure to use the right assay protocol.
- Check that the micropipettes are set to the required volume.
- Check that all the other equipment is available and ready to use.
- In case of problems, do not proceed further with the test and advise the supervisor.

M. ASSAY PROCEDURE

The assay has to be carried out according to what reported below, taking care to maintain the same incubation time for all the samples in testing.

M.1 Automated assay:

In case the test is carried out automatically with an ELISA system, we suggest to make the instrument aspirate 1000 µl Specimen Diluent and then 10 µl sample (1:101 dilution factor). The whole content is then dispensed into a properly defined dilution tube. Before the next sample is aspirated, needles have to be duly washed to avoid any cross-contamination among samples. When all the samples have been diluted make the instrument dispense 100 µl diluted samples into the proper wells of the microplate.

This procedure may be carried out also in two steps of dilutions of 1:10 each (90 µl Specimen Diluent + 10 µl sample) into a second dilution platform. Make then the instrument aspirate first 100 µl Specimen Diluent, then 10 µl liquid from the first dilution in the platform and finally dispense the whole content in the proper well of the assay microplate.

Do not dilute controls/calibrator as they are ready to use.

Dispense 100 µl calibrators/control in the appropriate calibration/control wells.

For the next operations follow the operative instructions reported below for the Manual Assay.

It is strongly recommended to check that the time lap between the dispensation of the first and the last sample will be calculated by the instrument and taken into consideration by delaying the first washing operation accordingly.

M. 2 Manual assay:

- Dilute samples 1:101 by dispensing first 10 µl sample and then 1 ml Specimen Diluent into a dilution tube; mix gently on vortex.
- Place the required number of Microwells in the microwell holder. Leave the well in position A1 empty for the operation of blanking.
- Dispense 100 µl of Negative Control and 100 µl of Calibrator in the proper wells in duplicate. Dispense 100 µl of Positive Control in single into the proper well. Do not dilute controls and the calibrator as they are ready to use !
- Dispense 100 µl diluted samples in the proper sample wells and then check that all the samples wells are blue colored and that controls and calibrator have been dispensed.
- Incubate the microplate for **60 min at +37°C**.

Important note: Strips have to be sealed with the adhesive sealing foil, supplied, only when the test is carried out manually. Do not cover strips when using ELISA automatic instruments.

- Wash the microplate with an automatic as reported previously (section I.3).
- Pipette 100 µl of the **Ag/Ab Immunocomplex** into each well, except the blanking well A1, and cover with the sealer. Check that all wells are red colored, except A1.

Important note: Be careful not to touch the plastic inner surface of the well with the tip filled with the **Ag/Ab Immunocomplex**. Contamination might occur.

- Incubate the microplate for **60 min at +37°C**.
- Wash microwells as in step 6.
- Pipette 100 µl Chromogen/Substrate mixture into each well, the blank well included. Then incubate the microplate at **room temperature (18-24°C) for 20 minutes**.

Important note: Do not expose to strong direct illumination. High background might be generated.

- Pipette 100 µl Sulphuric Acid into all the wells using the same pipetting sequence as in step 10. Addition of acid will turn the positive control and positive samples from blue to yellow.
- Measure the color intensity of the solution in each well, as described in section I.5, at 450nm filter (reading) and at 620-630nm (background subtraction, mandatory), blanking the instrument on A1.

Important notes:

- Ensure that no finger prints are present on the bottom of the microwell before reading. Finger prints could generate false positive results on reading.
- Reading has to be carried out just after the addition of the Stop Solution and anyway not any longer than 20 minutes after its addition. Some self oxidation of the chromogen can occur leading to high background.

N. ASSAY SCHEME

Controls&calibrator (*)	100 ul
Samples diluted 1:101	100 ul
1st incubation	60 min
Temperature	+37°C
Washing	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
Immunocomplex	100 ul
2nd incubation	60 min
Temperature	+37°C
Washing	n° 5 cycles with 20" of soaking OR n° 6 cycles without soaking
TMB/H2O2 mix	100 ul
3rd incubation	20 min
Temperature	r.t.
Sulphuric Acid	100 ul
Reading OD	450nm / 620-630nm

(*) Important Notes:

- The Calibrator (CAL) does not affect the Cut Off calculation, therefore it does not affect the test's results calculation.
- The Calibrator (CAL) used only if a laboratory internal quality control is required by the Management.

An example of dispensation scheme is reported below:

Microplate												
	1	2	3	4	5	6	7	8	9	10	11	12
A	BLK	S3										
B	NC	S4										
C	NC	S5										
D	CAL(*)	S6										
E	CAL(*)	S7										
F	PC	S8										
G	S1	S9										
H	S2	S10										

Legenda: BLK = Blank NC = Negative Control
CAL(*) = Calibrator-Not Mandatory PC = Positive Control S = Sample

O. INTERNAL QUALITY CONTROL

A validation check is carried out on the controls any time the kit is used in order to verify whether the performances of the assay are as qualified.

Control that the following data are matched:

Parameter	Requirements
Blank well	< 0.05 OD450nm value
Negative Control mean value (NC)	< 0.200 OD450nm value after blanking coefficient of variation < 30%
Positive Control	> 1.000 OD450nm

If the results of the test match the requirements stated above, proceed to the next section.

If they do not, do not proceed any further and perform the following checks:

Problem	Check
Blank well > 0.05 OD450nm	1. that the Chromogen/Substrate solution has not become contaminated during the assay
Negative Control (NC) > 0.200 OD450nm after blanking coefficient of variation > 30%	1. that the washing procedure and the washer settings are as validated in the pre qualification study; 2. that the proper washing solution has been used and the washer has been primed with it before use; 3. that no mistake has been done in the assay procedure (dispensation of positive control instead of negative control); 4. that no contamination of the negative control or of the wells where the control was dispensed has occurred due to positive samples, to spills or to the enzyme conjugate; 5. that micropipettes have not become contaminated with positive samples or with the enzyme conjugate 6. that the washer needles are not blocked or partially obstructed.
Positive Control < 1.000 OD450nm	1. that the procedure has been correctly performed; 2. that no mistake has occurred during the distribution of the control (dispensation of negative control instead of positive control). 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the positive control has occurred.

If any of the above problems have occurred, report the problem to the supervisor for further actions.

**** Important Notes:**

The analysis must be done proceeding as the reading step described in the section M, point 12.

If the Calibrator has used, verify the following data:

Check	Requirements
Calibrator	S/Co > 1.2

If the results of the test doesn't match the requirements stated above, operate as follows:

Problem	Check
Calibrator S/Co < 1.2	1. that the procedure has been correctly performed; 2. that no mistake has occurred during its distribution (e.g.: dispensation of negative control instead) 3. that the washing procedure and the washer settings are as validated in the pre qualification study; 4. that no external contamination of the calibrator has occurred.

Anyway, if all other parameters (Blank, Negative Control, Positive Control), match the established requirements, the test may be considered valid.

P. CALCULATION OF THE CUT-OFF

The test results are calculated by means of the mean OD450nm/620-630nm value of the Negative Control (NC) and a mathematical calculation, in order to define the following cut-off formulation:

$$\text{Cut-Off} = \text{NC} + 0.250$$

The value found for the test is used for the interpretation of results as described in the next paragraph.

Important note: When the calculation of results is performed by the operating system of an ELISA automated work station, ensure that the proper formulation is used to calculate the cut-off value and generate the correct interpretation of results.

Q. INTERPRETATION OF RESULTS

Test results are interpreted as a ratio of the sample OD450nm/620-630nm and the Cut-Off value (or S/Co) according to the following table:

S/Co	Interpretation
< 1.0	Negative
1.0 - 1.2	Equivocal
> 1.2	Positive

A negative result indicates that the patient is not undergoing an acute infection of Herpes Simplex Virus type 2.

Any patient showing an equivocal result, should be re-tested by examining a second sample taken from the patient after 1-2 weeks from first testing.

A positive result is indicative of a Herpes Simplex Virus type 2 infection.

An example of calculation is reported below (data obtained proceeding as the reading step described in the section M, point 12).

Important Note: The following data must not be used instead of real figures obtained by the user.

Negative Control: 0.090 – 0.110 – 0.070 OD450nm

Mean Value: 0.100 OD450nm

Lower than 0.200 – Accepted

Positive Control: 1.850 OD450nm

Higher than 1.000 – Accepted

Cut-Off = 0.100+0.250 = 0.350

Calibrator: 0.900 – 1.100 OD450nm

Mean value: 1.000 OD450nm S/Co = 2.8

S/Co higher than 1.2 – Accepted

Sample 1: 0.070 OD450nm

Sample 2: 1.690 OD450nm

Sample 1 S/Co < 1 = negative

Sample 2 S/Co > 1.2 = positive

Important notes:

1. Interpretation of results should be done under the supervision of the laboratory supervisor to reduce the risk of judgment errors and misinterpretations.
2. Particular attention in the interpretation of results has to be used in the follow-up of pregnancy for an infection of HSV due to the risk of severe neonatal malformations.
3. In pregnancy monitoring, it is strongly recommended that any positive result is confirmed first with the procedure described below and secondly with a different device for HSV IgM detection, before taking any preventive medical action.
4. Any positive sample should be submitted to the Confirmation Test reported in section T before giving a result of positivity. By carrying out this test, false reactions, leading to a misinterpretation of the analytical result, can be revealed and then ruled out.
5. When test results are transmitted from the laboratory to another facility, attention must be paid to avoid erroneous data transfer.
6. Diagnosis of infection has to be taken and released to the patient by a suitably qualified medical doctor.

R. PERFORMANCE CHARACTERISTICS

1. Limit of detection

No international standard for HSV1&2 IgM Antibody detection has been defined so far by the European Community.

In its absence, an Internal Gold Standard (or IGS), calibrated on the preparation named "Accurun – Anti HSV2 IgM plasma" produced by Boston Biomedica Inc., USA, code 9106072, has been defined in order to provide the device with a constant and excellent sensitivity..

The limit of detection of the assay has been therefore calculated on the IGS. A limiting dilution curve was prepared in Negative Control (NC).

Results of Quality Control are given in the following table:

OD450nm values

IGS	HSV2M.CE Lot # RD1	HSV2M.CE Lot # RD2	HSV2M.CE Lot # RD3
1X	0.560	0.572	0.590
2X	0.343	0.324	0.348
4X	0.239	0.218	0.225
NC	0.145	0.132	0.139

2. Diagnostic sensitivity:

The diagnostic sensitivity has been tested in a clinical trial on panels of 40 samples classified positive by a kit US FDA approved. The value obtained from the analysis was > 98%.

3. Diagnostic specificity:

The diagnostic specificity has been determined in a performance evaluation study on panels of more than 300 specimens, negative with the reference kit, derived from normal individuals of European origin.

Both plasma, derived with different standard techniques of preparation (citrate, EDTA and heparin), and sera have been used to determine the specificity. No false reactivity due to the method of specimen preparation has been observed.

Frozen specimens have also been tested to check whether this interferes with the performance of the test. No interference was observed on clean and particle free samples.

A study conducted on more than 60 potentially cross-reactive samples has not revealed any interference in the system.

No cross reaction were observed.

The Performance Evaluation has provided a value > 98%.

False positive reactions may be anyway pointed out and then ruled out in the interpretation of results with the procedure reported in section T, able to verify whether or not a positive result is real.

4. Precision:

Results are reported as follows:

HSV2M.CE: lot # RD1

Negative (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.092	0.113	0.097	0.101
Std.Deviation	0.011	0.019	0.010	0.013
CV %	12.25	16.83	10.24	13.11

Low reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.451	0.471	0.435	0.452
Std.Deviation	0.018	0.000	0.033	0.017
CV %	3.92	0.00	7.48	3.8

High reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	1.530	1.574	1.527	1.543
Std.Deviation	0.023	0.052	0.006	0.027
CV %	1.48	3.33	0.37	1.73

HSV2M.CE: lot # RD2

Negative (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.095	0.101	0.097	0.098
Std.Deviation	0.006	0.008	0.005	0.006
CV %	6.30	7.92	5.15	6.45

Low reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.431	0.428	0.453	0.437
Std.Deviation	0.023	0.018	0.023	0.021
CV %	5.3	4.2	5.10	4.9

High reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	1.558	1.552	1.541	1.550
Std.Deviation	0.031	0.025	0.039	0.032
CV %	1.98	1.61	2.53	2.04

HSV2M.CE: lot # RD3

Negative (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.104	0.108	0.099	0.104
Std.Deviation	0.015	0.010	0.011	0.012
CV %	14.4	9.2	11.11	11.57

Low reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	0.425	0.436	0.440	0.434
Std.Deviation	0.008	0.006	0.009	0.008
CV %	1.8	1.4	2.0	1.7

High reactive (N = 16)

Mean values	1st run	2nd run	3 rd run	Average value
OD 450nm	1.571	1.562	1.558	1.564
Std.Deviation	0.040	0.034	0.024	0.033
CV %	2.54	2.17	1.54	2.08

Important note:

The performance data have been obtained proceeding as the reading step described in the section M, point 12.

S. LIMITATIONS

Frozen samples containing fibrin particles or aggregates may generate false positive results.

Bacterial contamination or heat inactivation of the specimen may affect the absorbance values of the samples with consequent alteration of the level of the analyte.

This test is suitable only for testing single samples and not pooled ones.

Diagnosis of an infectious disease should not be established on the basis of a single test result. The patient's clinical history, symptomatology, as well as other diagnostic data should be considered.

T. CONFIRMATION TEST

In order to provide the medical doctor with the best accuracy in the follow-up of pregnancy, where a false positive result could lead to an operation of abortion, a confirmation test is reported. The confirmation test has to be carried out on any positive sample before a diagnosis of primary infection of HSV is released to the doctor.

Proceed for confirmation as follows:

1. Prepare the Antigen/Conjugate Complex as described in the proper section. This reagent is called Solution A.
2. Then 25 µl concentrated Enzymatic Conjugate are diluted in 500 µl Antigen Diluent and mixed gently on vortex. Do not use any lyophilized antigen vial for this procedure ! This solution is called Solution B.
3. The well A1 of the strip is left empty for blanking.
4. The Negative Control is dispensed in the strip in positions B1+C1. This is used for the calculation of the cut-off and S/Co values.
5. The positive sample to be confirmed, diluted 1:101, is dispensed in the strip in position D1+E1.
6. The strip is incubated for 60 min at +37°C.
7. After washing, the blank well A1 is left empty.
8. 100 µl of Solution A are dispensed in wells B1+C1+D1.
9. Then 100 µl of Solution B are added to well E1.
10. The strip is incubated for 60 min at +37°C.
11. After washing, 100 µl Chromogen/Substrate are added to all the wells and the strip is incubated for 20 min at r.t.
12. 100 µl Sulphuric Acid are added to all the wells and then their color intensity is measured at 450nm (reading filter) and at 620-630nm (background subtraction), blanking the instrument on A1.

Interpretation of results is carried out as follows:

1. If the sample in position D1 shows a S/Co value lower than 1.0 a problem of dispensation or contamination in the first test is likely to be occurred. The Assay Procedure in Section M has to be repeated to double check the analysis.
2. If the sample in position D1 shows a S/Co value higher than 1.2 and in position E1 shows a S/Co value still higher than 1.2 the sample is considered a **false positive**. The reactivity of the sample is in fact not dependent on the specific presence of HSV2 and a crossreaction with enzymatic conjugate has occurred.
3. If the sample in position D1 shows a S/Co value higher than 1.2 and in position E1 shows a S/Co value lower than 1.0 the sample is considered a **true positive**. The reactivity of the sample is in fact dependent on the specific presence of HSV and not due to any crossreaction.

The following table is reported for the interpretation of results

Well	S/Co		
D1	< 1.0	> 1.2	> 1.2
E1	< 1.0	> 1.2	< 1.0
Interpretation	Problem of contam.	False positive	True positive

REFERENCES

1. Engvall E. and Perlmann P.. J.Immunochimistry 8: 871-874, 1971
2. Engvall E. and Perlmann P.. J.Immunol.. 109: 129-135, 1971
3. Remington J.S. and Klein J.O.. (1996) In "Infectious diseases of fetus and newborn infant". Sanders, Philadelphia, London, Toronto.
4. Volk W.A. (1982) In "essential of Medical Microbiology". 2nd ed., pp 729, G.B. Lippincott Co. Philadelphia, New York, S.Josè, Toronto.
5. Leinikki P.O. et al.. J.Clin.Microbiol.. 8:418, 1978
6. Piroid E. et al.. Révue Méd.Vet.. 131:25, 1980.
7. Vaheri A. et al.. J.Med.Virol.. 5:171, 1980.
8. Vejtorp M. et al.. Acta Path.Microbiol.Scand.. 88:349, 1980.
9. Voller A. et al.. Brit.J.Exp.Pathol.. 56:338, 1975

All the IVD Products manufactured by the company are under the control of a certified Quality Management System in compliance with ISO 13485 rule. Each lot is submitted to a quality control and released into the market only if conforming with the EC technical specifications and acceptance criteria.

Manufacturer:
Dia.Pro Diagnostic Bioprobes S.r.l.
Via G. Carducci n° 27 – Sesto San Giovanni (MI) – Italy



HSV2 IgM

**Ensayo inmunoenzimático (ELISA) de
“captura” para la determinación de
anticuerpos IgM al Virus
Herpes Simplex tipo 2
en plasma y suero humanos**

- Uso exclusivo para diagnóstico “in vitro” -



DIA.PRO

**Diagnostic Bioprobes Srl
Via G. Carducci n° 27
20099 Sesto San Giovanni
(Milán) - Italia**

Teléfono +39 02 27007161

Fax +39 02 44386771

e-mail: info@diapro.it

HSV2 IgM

A. OBJETIVO DEL ESTUCHE.

Ensayo inmunoenzimático (ELISA) para la determinación de anticuerpos IgM al Virus Herpes Simplex tipo 2, en plasma y suero humanos, mediante un sistema de "captura". El estuche ha sido concebido para el seguimiento de pacientes infectados con HSV y para el monitoreo de la infección durante el embarazo, causa de riesgo de malformaciones en el neonato. Uso exclusivo para diagnóstico "in vitro".

B. INTRODUCCIÓN.

Los Virus del Herpes Simplex tipos 1 (HSV1) y 2 (HSV2) son grandes y complejos virus ADN que inducen la síntesis de diversas proteínas durante la infección, poseen un alto número de determinantes de reactividad cruzada y pocas secuencias tipo específicas. La mayor parte de las infecciones herpéticas primarias y recurrentes son causadas por HSV2, mientras que aquellas infecciones no asociadas a los genitales son causadas fundamentalmente por HSV1.

La detección de anticuerpos IgG e IgM específicos al virus, es importante en el diagnóstico de las infecciones agudas/primarias, así como en las reactivaciones de una infección latente, en ausencia de síntomas clínicos evidentes.

En individuos aparentemente sanos y durante el embarazo, pueden aparecer infecciones asintomáticas debidas a HSV. En pacientes inmunocomprometidos se pueden presentar severas infecciones herpéticas, donde la enfermedad evoluciona hacia patologías clínicas.

La determinación de anticuerpos específicos al virus constituye un elemento importante para el seguimiento de pacientes en grupos de riesgo, así como para el monitoreo de las infecciones severas y agudas.

C. PRINCIPIOS DEL ENSAYO.

El ensayo se basa en el principio de "captura de IgM", donde los anticuerpos de esta clase presentes en la muestra, son capturados por la fase sólida recubierta con un anticuerpo anti-IgM humano.

Luego del lavado, que elimina el resto de los componentes de la muestra en particular los anticuerpos IgG, se adiciona una preparación purificada de HSV 2, inactivado y marcado con un anticuerpo específico conjugado con Peroxidasa (HRP), lo cual permite detectar los anticuerpos IgM inmovilizados en la fase sólida. Posteriormente a la incubación, los pocillos se lavan para eliminar cualquier traza de conjugado en exceso y se añade el sustrato cromogénico. En presencia del conjugado, el sustrato es hidrolizado generándose una señal coloreada proporcional a la cantidad de anticuerpos IgM al HSV 2, presentes en la muestra.

La Prueba de Confirmación controla la ocurrencia de falsos positivos, lo cual permite a los clínicos una correcta interpretación de los resultados.

D. COMPONENTES.

Cada estuche contiene reactivos suficientes para realizar 96 pruebas.

1. Microplaca: MICROPLATE

12 tiras de 8 pocillos recubiertos con anticuerpos de cabra anti-IgM humano, purificados por afinidad, en presencia de proteínas de bovino.

Las placas están en una bolsa sellada con desecante. Se deben poner las mismas a temperatura ambiente antes de abrirlas, sellar las tiras sobrantes en la bolsa con el desecante y almacenar entre 2 y 8°C.

2. Control Negativo: CONTROL

1x4.0 ml/vial. Listo para el uso. Contiene 1% de proteínas del suero humano, 2% de caseína, tampón Tris 10 mM pH 6.0 +/- 0.1, 0.1% de Tween 20, además de azida sódica 0.09% y ProClin 300 0.045% como preservativos. El control negativo está codificado con el color amarillo pálido.

3. Control Positivo: CONTROL +

1x4.0 ml/vial. Listo para el uso. Contiene 1% de suero humano positivo a IgM HSV2, 2% de caseína, tampón Tris 10 mM pH 6.0 +/- 0.1, 0.1% de Tween 20, además de azida sódica 0.09% y ProClin 300 0.045% como preservativos. El control positivo está codificado con el color verde.

4. Calibrador: CAL ...ml

n° 1 vial. Liofilizado. Para disolver en agua calidad EIA como se indica en la etiqueta. Contiene anticuerpos IgM a HSV2, suero fetal bovino, además de sulfato de gentamicina 0.2 mg/ml y ProClin 300 0.045% como preservativos.

Nota: El volumen necesario para disolver el contenido del frasco varía en cada lote. Se recomienda usar el volumen indicado en la etiqueta.

5. Antígenos liofilizados HSV2 Ag: AG HSV2

N° 6 viales liofilizados. Contienen antígenos de HSV2 en un tampón proteico, inactivados por radiaciones gamma, 2% de proteínas de bovino, tampón Tris HCl 10 mM pH 6.8 +/- 0.1 además de 0.2 mg/ml de sulfato de gentamicina y 0.045% de ProClin 300. Debe disolverse con 1.9 ml de Diluyente de Antígeno, según se indica más adelante.

6. Tampón de Lavado Concentrado: WASHBUF 20X

1x60ml/botella. Solución concentrada 20x. Una vez diluida, la solución de lavado contiene tampón fosfato 10 mM a pH 7.0 +/- 0.2, Tween 20 al 0.05% y ProClin 300 al 0.045%.

7. Conjugado: CONJ 20X

1x0.8 ml/vial. Solución concentrada 20x. Contiene un anticuerpo específico anti-HSV2 conjugado con peroxidasa (HRP) diluido en un tampón proteico, tampón Tris 10mM a pH 6.8 +/- 0.1, 2% de BSA, además de 0.2 mg/ml de sulfato de gentamicina y ProClin 300 0.045% como preservativos.

8. Diluyente de Antígeno: AG DIL

n° 1 vial de 16 ml. Solución tamponada proteica para la preparación del inmunocomplejo. Contiene tampón Tris 10mM a pH 6.8 +/- 0.1, 2% de BSA, además de 0.2 mg/ml de sulfato de gentamicina y ProClin 300 0.045% como preservativos. El reactivo está codificado con el color rojo (0.01% de colorante rojo).

9. Diluyente de muestras : DILSPE

2x60ml/vial. Solución tamponada proteica para la dilución de las muestras. Contiene 2% de caseína, tampón Tris 10 mM a pH 6.0 +/- 0.1, 0.2% de Tween 20, además de azida sódica al 0.09% y 0.045 de ProClin 300 como preservativos. El reactivo está codificado con el color azul (0.01% de colorante azul).

10. Cromógeno/Substrato: SUBS TMB

1x16ml/vial. Contiene una solución tamponada citrato-fosfato 50mM pH 3.5-3.8, tetra-metil-benzidina (TMB) 0.03% y peróxido de hidrógeno (H₂O₂) 0.02% así como dimetilsulfóxido al 4%.

Nota: Evitar la exposición a la luz, la sustancia es fotosensible.

11. Ácido Sulfúrico: H₂SO₄ 0.3M

1x15ml/vial. Contiene solución de H₂SO₄ 0.3M
Atención: Irritante (H315, H319; P280, P302+P352, P332+P313, P305+P351+P338, P337+P313, P362+P363).

12. Sellador adhesivo, n° 2

13. Manual de instrucciones, n° 1

E. MATERIALES NECESARIOS NO SUMINISTRADOS.

1. Micropipetas calibradas (1000 ul, 100 ul and 10 ul) y puntas plásticas desechables.
2. Agua de calidad EIA (bidestilada o desionizada, tratada con carbón para remover químicos oxidantes usados como desinfectantes).
3. Timer con un rango de 60 minutos como mínimo.
4. Papel absorbente.
5. Incubador termostático de microplacas ELISA, calibrado (en seco o húmedo) fijo a 37°C (+/-0.5°C tolerancia).
6. Lector calibrado de microplacas de ELISA con filtros de 450nm (lectura) y de 620-630 nm.
7. Lavador calibrado de microplacas ELISA.
8. Vórtex o similar.

F. ADVERTENCIAS Y PRECAUCIONES.

1. El estuche debe ser usado por personal técnico adecuadamente entrenado, bajo la supervisión de un doctor responsable del laboratorio.
2. Todas las personas encargadas de la realización de las pruebas deben llevar las ropas protectoras adecuadas de laboratorio, guantes y gafas. Evitar el uso de objetos cortantes (cuchillas) o punzantes (agujas). El personal debe ser adiestrado en procedimientos de bioseguridad, según ha sido recomendado por el Centro de Control de Enfermedades de Atlanta, Estados Unidos, y publicado por el Instituto Nacional de Salud: "Biosafety in Microbiological and Biomedical Laboratories", ed.1984.
3. Todo el personal involucrado en el manejo de muestras debe estar vacunado contra HBV y HAV, para lo cual existen vacunas disponibles, seguras y eficaces.
4. Se debe controlar el ambiente del laboratorio para evitar la contaminación de los componentes con polvo o agentes microbianos cuando se abran los estuches, así como durante la realización del ensayo. Evitar la exposición del sustrato a la luz y las vibraciones de la mesa de trabajo durante el ensayo.
5. Conservar el estuche a temperaturas entre 2-8 °C, en un refrigerador con temperatura regulada o en cámara fría.
6. No intercambiar reactivos de diferentes lotes ni tampoco de diferentes estuches.
7. Comprobar que los reactivos no contienen precipitados ni agregados en el momento del uso. De darse el caso, informar al supervisor para realizar el procedimiento pertinente y reemplazar el estuche.
8. Evitar contaminación cruzada entre muestras de suero/plasma usando puntas desechables y cambiándolas luego de cada uso. No reutilizar puntas desechables.
9. Evitar contaminación cruzada entre los reactivos del estuche usando puntas desechables y cambiándolas luego de cada uso. No reutilizar puntas desechables.
10. No usar el producto después de la fecha de caducidad indicada en el estuche e internamente en los reactivos. Según estudios realizados, no se ha detectado pérdida relevante de actividad en estuches abiertos, en uso por un período de hasta 3 meses.
11. Tratar todas las muestras como potencialmente infecciosas. Las muestras de suero humano deben ser manipuladas al nivel 2 de bioseguridad, según ha sido recomendado por el Centro de Control de Enfermedades de Atlanta, Estados Unidos y publicado por el Instituto Nacional de Salud: "Biosafety in Microbiological and Biomedical Laboratories", ed.1984.
12. Se recomienda el uso de material plástico desechable para la preparación de las soluciones de lavado y para la transferencia de los reactivos a los diferentes equipos automatizados a fin de evitar contaminaciones.

13. Los desechos producidos durante el uso del estuche deben ser eliminados según lo establecido por las directivas nacionales y las leyes relacionadas con el tratamiento de los residuos químicos y biológicos de laboratorio. En particular, los desechos líquidos provenientes del proceso de lavado deben ser tratados como potencialmente infecciosos y deben ser inactivados. Se recomienda la inactivación con lejía al 10% de 16 a 18 horas o el uso de la autoclave a 121°C por 20 minutos.
14. En caso de derrame accidental de algún producto, se debe utilizar papel absorbente embebido en lejía y posteriormente en agua. El papel debe eliminarse en contenedores designados para este fin en hospitales y laboratorios.
15. El ácido sulfúrico es irritante. En caso de derrame, se debe lavar la superficie con abundante agua.
16. Otros materiales de desecho generados durante la utilización del estuche (por ejemplo: puntas usadas en la manipulación de las muestras y controles, microplacas usadas) deben ser manipuladas como fuentes potenciales de infección de acuerdo a las directivas nacionales y leyes para el tratamiento de residuos de laboratorio.

G. MUESTRA: PREPARACIÓN Y RECOMENDACIONES.

1. Extraer la sangre asépticamente por punción venosa y preparar el suero o plasma según las técnicas estándar de los laboratorios de análisis clínico. No se ha detectado que el tratamiento con citrato, EDTA o heparina afecte las muestras.
2. Las muestras deben estar identificadas claramente mediante código de barras o nombres, a fin de evitar errores en los resultados. Se recomienda el uso del código de barras.
3. Las muestras hemolizadas (color rojo) o hiperlipémicas (aspecto lechoso) deben ser descartadas para evitar falsos resultados, al igual que aquellas donde se observe la presencia de precipitados, restos de fibrina o filamentos microbianos.
4. El suero y el plasma pueden conservarse a una temperatura entre +2° y +8°C en tubos de recolección principales hasta cinco días después de la extracción. No congelar tubos de recolección principales. Para periodos de almacenamiento más prolongados, las muestras de plasma o suero, retiradas cuidadosamente del tubo de extracción principal, pueden almacenarse congeladas a -20°C durante al menos 12 meses. Evitar congelar/descongelar cada muestra más de una vez, ya que pueden generarse partículas que podrían afectar al resultado de la prueba.
5. Si hay presencia de agregados, la muestra se puede aclarar mediante centrifugación a 2000 rpm durante 20 minutos o por filtración con un filtro de 0,2-0,8 micras.

H. PREPARACIÓN DE LOS COMPONENTES Y PRECAUCIONES.

Estudios de estabilidad realizados en estuches en uso (hasta 6 veces) no han arrojado pérdida de actividad significativa en un período de 3 meses.

Microplacas:

Dejar la microplaca a temperatura ambiente (aprox. 1 hora) antes de abrir el envase. Compruebe que el desecante no esté de un color verde oscuro, lo que indicaría un defecto de conservación. De ser así, debe solicitar el servicio de Dia.Pro: atención al cliente.

Las tiras de pocillos no utilizadas, deben guardarse herméticamente cerradas en la bolsa de aluminio con el desecante a 2-8°C. Una vez abierto el envase, las tiras sobrantes, se mantienen estables hasta que el indicador de humedad dentro de la bolsa del desecante cambie de amarillo a verde.

Control Negativo:

Listo para el uso. Mezclar bien con la ayuda de un vórtex, antes de usar.

Control Positivo:

Listo para el uso. Mezclar bien con la ayuda de un vórtex, antes de usar.

Calibrador:

Añadir al polvo liofilizado, el volumen de agua de calidad ELISA indicado en la etiqueta. Dejar disolver completamente y luego mezclar cuidadosamente con el vórtex antes de usar.

Nota: Para preservar la reactividad se recomienda mantenerla congelada en alícuotas a -20°C . No recongelar.

Solución de Lavado Concentrada:

Todo el contenido de la solución concentrada debe diluirse 20x con agua bidestilada y mezclarse suavemente antes de usarse. Durante la preparación evitar la formación de espuma y burbujas, lo que podría influir en la eficiencia de los ciclos de lavado.

Nota: Una vez diluida, la solución es estable por una semana a temperaturas entre $+2$ y 8°C .

Inmunocomplejo Ag/Ab:

Proceder cuidadosamente según se indica:

1. Disolver el contenido de un vial liofilizado utilizando 1.9 ml de Diluyente Antígeno. Dejar disolver completamente y luego mezclar cuidadosamente con el vórtex.
2. Mezclar el Conjugado concentrado con ayuda del vórtex. Añadir luego 0.1 ml del mismo al vial del Ag HSV2 disuelto y mezclar suavemente en el vórtex.

Notas Importantes:

1. Disolver y preparar solamente los viales necesarios para la prueba. El inmunocomplejo obtenido no es estable. Almacenar la solución sobrante en alícuotas a -20°C .
2. La preparación del inmunocomplejo debe realizarse **justo antes** de dispensar las muestras y los controles en la placa. Mezclar nuevamente en vórtex justo antes de usar.

Diluyente de muestras :

Listo para el uso. Mezclar bien con un vórtex antes de usar.

Cromógeno/ Substrato:

Listo para el uso. Mezclar bien con un vórtex antes de usar. Evitar posible contaminación del líquido con oxidantes químicos, polvo o microbios. Evitar la exposición a la luz, agentes oxidantes y superficies metálicas. En caso de que deba transferirse el reactivo, usar contenedores de plástico, estériles y desechables, siempre que sea posible.

Ácido Sulfúrico:

Listo para el uso. Mezclar bien con un vórtex antes de usar.

Leyenda:

Indicación de peligro, **Frases H**

H315 – Provoca irritación cutánea.

H319 – Provoca irritación ocular grave.

Consejo de prudencia, **Frases P**

P280 – Llevar guantes/prendas/gafas/máscara de protección.

P302 + P352 – EN CASO DE CONTACTO CON LA PIEL: Lavar con agua y jabón abundantes.

P332 + P313 – En caso de irritación cutánea: Consultar a un médico.

P305 + P351 + P338 – EN CASO DE CONTACTO CON LOS OJOS: Aclarar cuidadosamente con agua durante varios minutos. Quitar las lentes de contacto, si lleva y resulta fácil. Seguir aclarando.

P337 + P313 – Si persiste la irritación ocular: Consultar a un médico.

P362 + P363 – Quitarse las prendas contaminadas y lavarlas antes de volver a usarlas.

I. INSTRUMENTOS Y EQUIPAMIENTO UTILIZADOS EN COMBINACIÓN CON EL ESTUCHE.

1. Las micropipetas deben ser calibradas para dispensar correctamente el volumen requerido en el ensayo y sometidas a una descontaminación periódica de las partes que pudieran entrar accidentalmente en contacto con la muestra o los reactivos (acohol 70%, lejía 10%, de calidad de los desinfectantes hospitalarios). Deben además, ser regularmente revisadas para mantener una precisión del 1% y una confiabilidad de $\pm 2\%$. Deben descontaminarse periódicamente los residuos de los componentes del estuche.
2. La incubadora de ELISA debe ser ajustada a 37°C ($\pm 0.5^{\circ}\text{C}$ de tolerancia) y controlada periódicamente para mantener la temperatura correcta. Pueden emplearse incubadoras secas o baños de agua siempre que estén validados para la incubación de pruebas de ELISA.
3. El **lavador ELISA** es extremadamente importante para el rendimiento global del ensayo. El lavador debe ser validado de forma minuciosa previamente, revisado para comprobar que suministra el volumen de dispensación correcto y enviado regularmente a mantenimiento de acuerdo con las instrucciones de uso del fabricante. En particular, deben lavarse minuciosamente las sales con agua desionizada del lavador al final de la carga de trabajo diaria. Antes del uso, debe suministrarse extensivamente solución de lavado diluida al lavador. Debe enviarse el instrumento semanalmente a descontaminación según se indica en su manual (se recomienda descontaminación con NaOH 0.1 M). Para asegurar que el ensayo se realiza conforme a los rendimientos declarados, basta con 5 ciclos de lavado (aspiración + dispensado de 350 μl /pocillo de solución de lavado + 20 segundos de remojo = 1 ciclo). Si no es posible remojar, añadir un ciclo de lavado adicional. Un ciclo de lavado incorrecto o agujas obstruidas con sal son las principales causas de falsas reacciones positivas.
4. Los tiempos de incubación deben tener un margen de $\pm 5\%$.
5. El lector de microplacas ELISA debe estar provisto de un filtro de lectura de 450nm y de un segundo filtro de 620-630nm, obligatorio para reducir interferencias en la lectura. El procedimiento estándar debe contemplar: a) Ancho de banda $\leq 10\text{nm}$ b) Rango de absorbancia de 0 a ≥ 2.0 , c) Linealidad ≥ 2.0 , reproducibilidad $\geq 1\%$. El blanco se prueba en el pocillo indicado en la sección "Control de calidad interno". El sistema óptico del lector debe ser calibrado periódicamente para garantizar la correcta medición de la densidad óptica, según las normas del fabricante.
6. En caso de usar un sistema automatizado de ELISA, los pasos críticos (dispensado, incubación, lavado, lectura, agitación y procesamiento de datos) deben ser cuidadosamente fijados, calibrados, controlados y periódicamente ajustados, para garantizar los valores indicados en la sección "Control de calidad interno". El protocolo del ensayo debe ser instalado en el sistema operativo de la unidad y validado tanto para el lavador como para el lector. Por otro lado, la parte del sistema que maneja los líquidos (dispensado y lavado) debe ser validada y fijada correctamente. Debe prestarse particular atención a evitar el arrastre por las agujas de dispensación y de lavado, a fin de minimizar la posibilidad de ocurrencia de falsos positivos por contaminación de los pocillos adyacentes por muestras fuertemente reactivas. Se recomienda el uso de sistemas automatizados para el pesquaje en unidades de sangre y cuando la cantidad de muestras supera las 20-30 unidades por ensayo.
7. El servicio de atención al cliente de Dia.Pro, ofrece apoyo al usuario para calibrar, ajustar e instalar los equipos a usar en combinación con el estuche, con el propósito de asegurar el cumplimiento de los requerimientos descritos.

L. OPERACIONES Y CONTROLES PREVIOS AL ENSAYO.

1. Compruebe la fecha de caducidad indicada en la parte externa del estuche (envase primario). No usar si ha caducado.
2. Compruebe que los componentes líquidos no están contaminados con partículas o agregados visibles. Asegúrese de que el cromógeno (TMB) es incoloro o azul pálido, aspirando un pequeño volumen del mismo con una pipeta estéril de plástico. Compruebe que no han ocurrido rupturas ni derrames de líquido dentro de la caja (envase primario) durante el transporte. Asegurarse de que la bolsa de aluminio que contiene la microplaca no esté rota o dañada.
3. Diluir totalmente la Solución de Lavado Concentrada 20X, como se ha descrito anteriormente.
4. Disolver el Calibrador como se ha descrito anteriormente y mezclar suavemente.
5. Dejar los componentes restantes alcanzar la temperatura ambiente (aprox. 1 hora), mezclar luego suavemente en el vórtex todos los reactivos líquidos.
6. Ajustar la incubadora de ELISA a 37°C y cebar el lavador de ELISA utilizando la solución de lavado, según las instrucciones del fabricante. Fijar el número de ciclos de lavado lavado según se indica en la sección específica.
7. Comprobar que el lector de ELISA esté encendido al menos 20 minutos antes de realizar la lectura.
8. En caso de trabajar automáticamente, encender el equipo y comprobar que los protocolos estén correctamente programados.
9. Comprobar que las micropipetas estén fijadas en el volumen requerido.
10. Asegurarse de que el equipamiento a usar esté en perfecto estado, disponible y listo para el uso.
11. En caso de surgir algún problema, se debe detener el ensayo y avisar al supervisor.

M. PROCEDIMIENTO DEL ENSAYO.

El ensayo debe realizarse según las instrucciones que siguen a continuación, es importante mantener en todas las muestras el mismo tiempo de incubación.

M.1 Ensayo automatizado:

En el caso de que el ensayo se realice de manera automatizada con un sistema ELISA, se recomienda programar al equipo para aspirar 1000µl de Diluyente de Muestras, y posteriormente 10µl de muestra (factor de dilución 1:101).

La mezcla debe ser dispensada cuidadosamente en un tubo de dilución. Antes de aspirar la muestra siguiente, las agujas deben lavarse debidamente para evitar cualquier contaminación cruzada entre las muestras. Cuando todas las muestras han sido diluidas, programar el equipo para dispensar 100 µl de las mismas en los pocillos correspondientes.

Este procedimiento puede realizarse en dos pasos de dilución de 1:10 cada uno (90 µl de Diluyente de Muestras + 10 µl de muestra) en una segunda plataforma de dilución. Programar el equipo para aspirar primeramente 100 µl de Diluyente de Muestras, luego 10 µl de la primera dilución en la plataforma y finalmente dispensar todo el contenido en los pocillos apropiados de la microplaca.

No diluir el Calibrador ni los controles, ya que están listos para el uso.

Dispensar 100ul de controles/calibrador en los pocillos correspondientes.

Para las operaciones siguientes, consulte las instrucciones que aparecen debajo para el Ensayo Manual.

Es muy importante comprobar que el tiempo entre el dispensado de la primera y la última muestra sea calculado por el instrumento y considerado para los lavados.

M.2 Ensayo Manual.

1. Diluir las muestras 1:101 dispensando primeramente 10 µl de muestra y luego 1 ml de Diluyente de Muestra en un tubo de dilución, mezclar bien con vórtex.
2. Poner el número de tiras necesarias en el soporte plástico. Dejar el pocillo A1 vacío para el blanco.
3. Dispensar 100 µl del Control Negativo y 100µl del Calibrador por duplicado. Luego dispensar 100µl del Control Positivo (sencillo) en los respectivos pocillos. No diluir los controles ni el calibrador ya que están listos para el uso.
4. Dispensar 100 µl de las muestras diluidas en los pocillos correspondientes y chequear luego que estos pocillos son de color azul y que los controles y el calibrador han sido añadidos.
5. Incubar la microplaca **60 min a +37°C**.

Nota importante: Las tiras se deben sellar con el adhesivo suministrado solo cuando se hace el test manualmente. No sellar cuando se emplean equipos automatizados de ELISA.

6. Lavar la microplaca según se indica en la sección I.3.
7. Dispensar 100uL del **Inmunocomplejo Ag/Ab** en todos los pocillos, excepto en el A1 y cubrir con el sellador. Compruebe que este reactivo de color rojo ha sido añadido en todos los pocillos excepto el A1.

Nota importante: Tener cuidado de no tocar la pared interna del pocillo con la punta de la pipeta al dispensar el **Inmunocomplejo Ag/Ab**. Podría producirse contaminación.

8. Incubar la microplaca **60 min a +37°C**.
9. Lavar la microplaca, de igual forma que en el paso 6.
10. Dispensar 100µl del Cromógeno/Substrato en todos los pocillos, incluido el A1. Incubar la microplaca a **temperatura ambiente (18-24°C) durante 20 minutos**.

Nota importante: No exponer directamente a fuerte iluminación, de lo contrario se generan interferencias.

11. Dispensar 100µl de Ácido Sulfúrico en todos los pocillos para detener la reacción enzimática, usar la misma secuencia que en el paso 10. La adición del ácido cambia el color de los controles positivos y las muestras positivas de azul a amarillo.
12. Medir la intensidad del color con el lector, según se describe en la sección I.5, utilizando un filtro de 450 nm (lectura) y otro de 620-630 nm (substracción del fondo, obligatorio), calibrando el instrumento con el pocillo A1 (blanco).

Notas generales importantes:

1. Segurarse de que no hay impresiones digitales en el fondo de los pocillos antes de leer. Podrían generarse falsos positivos en la lectura.
2. La lectura debe hacerse inmediatamente después de añadir la solución de parada y, en cualquier caso, nunca transcurridos 20 minutos después de su adición. Se podría producir auto oxidación del cromógeno causando un elevado fondo.

N. ESQUEMA DEL ENSAYO.

Controles&Calibrador (*)	100 ul
Muestras diluidas 1:101	100 ul
1^{ra} incubación	60 min
Temperatura	+37°C
Lavado	5 ciclos con 20"de remojo o 6 ciclos sin remojo
Inmunocomplejo	100 ul
2^{da} incubación	60 min

Temperatura	+37°C
Lavado	5 ciclos con 20"de remojo o 6 ciclos sin remojo
Mezcla TMB/H ₂ O ₂	100 ul
3 ^{ra} incubación	20 min
Temperatura	t.a.*
Ácido Sulfúrico	100 ul
Lectura D.O.	450nm / 620-630nm

t.a.*temperatura ambiente

(*) Notas importantes:

- El calibrador (CAL) no afecta al cálculo del valor de corte y, por lo tanto, no afecta al cálculo de los resultados de la prueba.
- El calibrador (CAL) se usa solo si la gestión requiere un control interno de calidad del laboratorio.

A continuación se describe un ejemplo del esquema de dispensado:

Microplaca												
	1	2	3	4	5	6	7	8	9	10	11	12
A	BL	M 3										
B	CN	M 4										
C	CN	M 5										
D	CAL(*)	M 6										
E	CAL(*)	M 7										
F	CP	M 8										
G	M 1	M 9										
H	M 2	M10										

Leyenda: BL = Blanco CN = Control Negativo
(*) CAL = Calibrador - No Obligatorio CP = Control Positivo
M = Muestra

O. CONTROL DE CALIDAD INTERNO.

Se realiza una validación sobre los controles y el calibrador cada vez que se usa el estuche, para verificar si el performance del ensayo es el esperado.

Asegurar el cumplimiento de los siguientes parámetros:

Parámetro	Exigencia
Pocillo Blanco	< 0.050 DO450nm
Control Negativo, valor medio (CN)	< 0.200 DO450nm valor después de leer el blanco Coeficiente de variación < 30%
Control Positivo	> 1000 DO450nm

Si los resultados del ensayo coinciden con lo establecido anteriormente, pase a la siguiente sección.

En caso contrario, detenga el ensayo y compruebe:

Problema	Compruebe que
Pocillo blanco > 0.050DO450nm	la solución cromógeno/substrato no se ha contaminado durante el ensayo.
Control Negativo (CN) > 0.200 DO450nm después de leer el blanco Coeficiente de variación > 30%	1. el proceso de lavado y los parámetros del lavador estén validados según los estudios previos de calificación. 2. se ha usado la solución de lavado apropiada y que el lavador ha sido cebado con la misma antes del uso. 3. no se han cometido errores en el procedimiento (dispensar el control positivo en lugar del negativo). 4. no ha existido contaminación del control negativo o de sus pocillos debido a muestras

	positivas derramadas, o al conjugado. 5. las micropipetas no se han contaminado con muestras positivas o con el conjugado. 6. las agujas del lavador no estén parcial o totalmente obstruidas.
Control Positivo < 1000 DO450nm	1. el procedimiento ha sido realizado correctamente. 2. no se han cometido errores en el procedimiento (dispensar el control negativo en lugar del positivo). 3. el proceso de lavado y los parámetros del lavador estén validados según los estudios previos de calificación. 4. no ha ocurrido contaminación externa del control positivo.

Si ocurre alguno de los problemas anteriores, luego de comprobar, informe al supervisor para tomar las medidas pertinentes.

** Notas importantes:

El análisis debe seguir el paso de lectura descrito en la sección M, punto 12.

Si se ha usado el Calibrador, comprobar los siguientes datos:

Parámetro	Exigencia
Calibrador	M/Co > 1.2

Si los resultados de la prueba no se corresponden con los requisitos indicados anteriormente, proceder del siguiente modo:

Problema	Compruebe que
Calibrador M/Co < 1.2	1. el procedimiento ha sido realizado correctamente. 2. no ha habido errores durante su distribución (dispensar el control negativo en lugar del calibrador). 3. el proceso de lavado y los parámetros del lavador estén validados según los estudios previos de calificación. 4. no ha ocurrido contaminación externa del calibrador.

En cualquier caso, si todos los demás parámetros (blanco, control negativo, control positivo) se corresponden con los requisitos establecidos, la prueba puede considerarse válida.

P. CÁLCULO DEL VALOR DE CORTE.

Los resultados de la prueba se calculan a partir de un valor medio de DO450nm/620-630nm del control Negativo (CN), mediante un valor de corte (Co) hallado con la siguiente fórmula:

$$\text{Valor de corte} = \text{CN} + 0.250$$

El valor encontrado en la prueba es utilizado para la interpretación de los resultados, según se describe a continuación.

Nota Importante: Cuando el cálculo de los resultados se halla mediante el sistema operativo de un equipo de ELISA automático, asegurarse de que la formulación usada para el cálculo del valor de corte, y para la interpretación de los resultados sea correcta.

Q. INTERPRETACIÓN DE LOS RESULTADOS.

La interpretación de los resultados se realiza mediante la razón entre las DO a 450nm de las muestras (M) y el Valor de corte (Co).

Los resultados se interpretan según la siguiente tabla:

(M/Co)	Interpretación
< 1.0	Negativo
1.0 – 1.2	Equívoco
> 1.2	Positivo

Un resultado negativo indica que el paciente no está padeciendo infección aguda por el Virus Herpes Simplex tipo 2. Cualquier paciente, cuya muestra resulte equívoca debe someterse a una nueva prueba con una segunda muestra de sangre colectada 1 ó 2 semanas después de la inicial.

Un resultado positivo es indicativo de infección por el Virus Herpes Simplex tipo 2.

A continuación, un ejemplo de los cálculos a realizar (datos obtenidos siguiendo el paso de lectura descrito en la sección M, punto 12).

Los siguientes datos no deben usarse en lugar de los valores reales obtenidos en el laboratorio.

Control Negativo: 0.090 – 0.110 – 0.070 DO 450nm

Valor medio: 0.100 DO 450nm

Menor de 0.150 – Válido

Control Positivo: 1.850 DO 450nm

Mayor de 1000 – Válido

Valor de corte = 0.100+0.250 = 0.350

Calibrador: 0.900 – 1.100 DO 450nm

Valor medio: 1.000 DO 450nm *M/Co = 2.8*

M/Co Mayor de 1.2 – Válido

Muestra 1: 0.070 DO 450nm

Muestra 2: 1.690 DO 450nm

Muestra 1 M/Co < 1 = negativa

Muestra 2 M/Co > 1.2 = positiva

Notas importantes:

- La interpretación de los resultados debe hacerse bajo la vigilancia del supervisor del laboratorio para reducir el riesgo de errores de juicio y de interpretación.
- Debe ponerse particular atención a la interpretación de los resultados ante sospecha de infección primaria por HSV en el embarazo, debido a las posibilidades de malformaciones del neonato.
- En el monitoreo de infección por HSV durante el embarazo, se recomienda, antes de tomar cualquier decisión médica preventiva, confirmar cualquier resultado positivo, primero con el procedimiento descrito y luego con un sistema de detección de IgM anti-HSV.
- Antes de emitir un resultado positivo, cada muestra reactiva debe someterse al examen de confirmación reportado en la sección T, lo cual permite una correcta interpretación de los resultados ya que descarta los falsos positivos.
- Cuando se transmiten los resultados de la prueba, del laboratorio a otras instalaciones, debe ponerse mucha atención para evitar el traslado de datos erróneos.
- El diagnóstico de infección debe ser evaluado y comunicado al paciente por un médico calificado.

R. PERFORMANCES.

1. Límite de detección.

Hasta el momento no ha sido definido por la Comunidad Europea, un estándar internacional para la detección de

anticuerpos IgM a HSV1&2. En ausencia del mismo y para garantizar una óptima sensibilidad, el límite de detección del ensayo ha sido calculado por medio de un Gold Standard Interno (IGS), a partir de una preparación "Accurun–Anti HSV 2 IgM Plasma", producida por Boston Biomedica Inc., Estados Unidos, código 9106072. Se construyó una curva de dilución limitante utilizando el Control Negativo (CN).

La siguiente tabla muestra los resultados del Control de Calidad:

Valores DO 450nm

IGS	HSV2M.CE Lote # RD1	HSV2M.CE Lote # RD2	HSV2M.CE Lote # RD3
1X	0.560	0.572	0.590
2X	0.343	0.324	0.348
4X	0.239	0.218	0.225
CN	0.145	0.132	0.139

2. Sensibilidad diagnóstica :

La sensibilidad diagnóstica se ha estudiado en un ensayo clínico utilizando paneles de 40 muestras, clasificadas como positivas mediante un estuche aprobado US FDA. El valor obtenido del análisis fue > 98%.

3. Especificidad diagnóstica :

La especificidad diagnóstica ha sido determinada en un ensayo clínico, utilizando paneles de más de 300 muestras provenientes de individuos sanos de origen europeo, clasificadas como negativas mediante un estuche de referencia. Se emplearon además plasma sometido a métodos de tratamiento estándar (citrato, EDTA y heparina) y suero humano para determinar la especificidad. No se ha observado falsa reactividad debida a los métodos de tratamiento de muestras.

Las muestras congeladas han sido probadas para comprobar si la colección y el almacenamiento interfiere con el procedimiento del ensayo. No se ha observado interferencia a partir de muestras limpias y libres de agregados.

Se realizó un estudio con más de 60 muestras que pudieran introducir reacción cruzada y no se observó interferencia alguna en el sistema. No se detectó reacción cruzada.

El estudio para evaluar el performance reveló un valor > 98%.

El procedimiento reportado en la sección T, permite verificar los resultados falsos positivos y de esta forma lograr una correcta interpretación de los resultados.

4. Precisión :

Ha sido calculada a partir de tres muestras, una negativa, una débilmente positiva y una positiva, examinadas en 16 réplicas en tres corridas separadas.

Los resultados son los siguientes:

HSV2M.CE: lote # RD1

Negativa (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	0.092	0.113	0.097	0.101
Desviación estándar	0.011	0.019	0.010	0.013
CV %	12.25	16.83	10.24	13.11

Débil reactiva (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	0.451	0.471	0.435	0.452
Desviación estándar	0.018	0.000	0.033	0.017
CV %	3.92	0.00	7.48	3.8

Altamente reactiva (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	1.530	1.574	1.527	1.543
Desviación estándar	0.023	0.052	0.006	0.027
CV %	1.48	3.33	0.37	1.73

HSV2M.CE: lote # RD2

Negativa (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	0.095	0.101	0.097	0.098
Desviación estándar	0.006	0.008	0.005	0.006
CV %	6.30	7.92	5.15	6.45

Débil reactiva (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	0.431	0.428	0.453	0.437
Desviación estándar	0.023	0.018	0.023	0.021
CV %	5.3	4.2	5.10	4.9

Altamente reactiva (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	1.558	1.552	1.541	1.550
Desviación estándar	0.031	0.025	0.039	0.032
CV %	1.98	1.61	2.53	2.04

HSV2M.CE: lote # RD3

Negativa (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	0.104	0.108	0.099	0.104
Desviación estándar	0.015	0.010	0.011	0.012
CV %	14.4	9.2	11.11	11.57

Débil reactiva (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	0.425	0.436	0.440	0.434
Desviación estándar	0.008	0.006	0.009	0.008
CV %	1.8	1.4	2.0	1.7

Altamente reactiva (N = 16)

Valores medios	1ª corrida	2ª corrida	3ª corrida	Valor promedio
DO 450nm	1.571	1.562	1.558	1.564
Desviación estándar	0.040	0.034	0.024	0.033
CV %	2.54	2.17	1.54	2.08

Nota importante:

Los datos de rendimiento se obtuvieron siguiendo el paso de lectura descrito en la sección M, punto 12.

S. LIMITACIONES.

La contaminación bacteriana o la inactivación por calor de la muestra pueden afectar los valores de DO y por tanto alterar los niveles del analito.

Las muestras que luego de ser descongeladas presentan partículas de fibrina o partículas agregadas, generan algunos resultados falsos positivos.

El ensayo es útil solo para probar muestras independientes y no mezclas.

El diagnóstico de una enfermedad infecciosa no debe establecerse en base a un solo resultado, sino que deben tenerse en consideración la historia clínica del paciente, la sintomatología, así como otros datos diagnósticos.

T. PRUEBA DE CONFIRMACIÓN.

Se ejecuta esta prueba con el propósito de garantizar la mayor precisión del ensayo en el seguimiento del embarazo, donde un resultado falso positivo puede conducir a un aborto. La misma debe realizarse a cada una de las muestras positivas, antes de emitir un diagnóstico de infección por HSV.

Proceder para la confirmación como sigue:

1. Preparar el complejo Antígeno/Conjugado como se describe anteriormente. Este reactivo se denomina Solución A.
2. Diluir el Conjugado concentrado, 1:20 en el Diluyente de Antígeno (ej: 25 µl de Conjugado concentrado en 500 µl de Diluyente de Antígeno) y mezclar suavemente con ayuda del vórtex. No usar ningún vial de Ag liofilizado para este procedimiento! Este reactivo se denomina Solución B.
3. Dejar vacío el pocillo A1 para el blanco.
4. Dispensar el Control Negativo en las posiciones B1+C1, se utiliza para calcular el valor de corte y los valores M/Co.
5. Diluir 1:101 la muestra positiva para confirmar y dispensarla en las posiciones D1+E1.
6. Incubar la tira 60 minutos a +37°C.
7. Luego del lavado, el pocillo A1 para el blanco queda vacío.
8. Dispensar 100 µl de la Solución A en los pocillos B1+C1+D1.
9. Dispensar 100 µl de la Solución B en el pocillo E1.
10. Incubar la tira 60 minutos a +37°C.
11. Luego del lavado, adicionar 100 µl del Cromógeno/Substrato en todos los pocillos e incubar la tira 20 minutos a temperatura ambiente.
12. Dispensar 100µl del Acido Sulfúrico en todos los pocillos y medir la intensidad del color con el lector, según se describe en la sección I.5, utilizando un filtro de 450 nm (lectura) y otro de 620-630 nm (substracción del fondo, recomendado), calibrando el instrumento con el pocillo A1 (blanco).

La interpretación de los resultados se realiza de la siguiente forma:

1. Si la muestra en posición D1 tiene un valor de M/Co menor de 1.0, probablemente en el primer ensayo haya ocurrido un error en el dispensado o alguna contaminación. Debe repetirse el Procedimiento del Ensayo, sección M.
2. Si la muestra en posición D1 tiene un valor de M/Co mayor de 1.2 y en posición E1 el valor de M/Co es todavía mayor de 1.2, la muestra se considera un **falso positivo**. La reactividad de la muestra, en este caso, no depende de la presencia específica de HSV2, por lo tanto ha ocurrido una reacción cruzada con el conjugado.
3. Si la muestra en posición D1 tiene un valor de M/Co mayor de 1.2 y en la posición E1 el valor M/Co es menor de 1.0 se considera **realmente positiva**. La reactividad de la muestra, en este caso se debe a la presencia específica de HSV y no a reacciones cruzadas.

En la siguiente tabla se muestra la interpretación de los resultados:

Pocillo	M/Co		
D1	< 1.0	> 1.2	> 1.2
E1	< 1.0	> 1.2	< 1.0
Interpretación	Probl. de contam.	Falso positivo	Realmente positivo

BIBLIOGRAFÍA.

1. Engvall E. and Perlmann P.. J.Immunochemistry 8: 871-874, 1971
2. Engvall E. and Perlmann P.. J.Immunol.. 109: 129-135, 1971
3. Remington J.S. and Klein J.O.. (1996) In "Infectious diseases of fetus and newborn infant". Sanders, Philadelphia, London, Toronto.
4. Volk W.A. (1982) In "essential of Medical Microbiology". 2nd ed., pp 729, G.B. Lippincott Co. Philadelphia, New York, S.Josè, Toronto.
5. Leinikki P.O. et al.. J.Clin.Microbiol.. 8:418, 1978
6. Piroid E. et al.. Révue Méd.Vet.. 131:25, 1980.
7. Vaheri A. et al.. J.Med.Virol.. 5:171, 1980.
8. Vejtorp M. et al.. Acta Path.Microbiol.Scand.. 88:349, 1980.
9. Voller A. et al.. Brit.J.Exp.Pathol.. 56:338, 1975

Todos los productos de diagnóstico in vitro fabricados por la empresa son controlados por un sistema certificado de control de calidad conforme a la norma ISO 13485. Cada lote se somete a un control de calidad y se libera al mercado únicamente si se ajusta a las especificaciones técnicas y criterios de aceptación de la CE.

Fabricante:
Dia.Pro Diagnostic Bioprobes S.r.l.
Via G. Carducci n° 27 – Sesto San Giovanni (Milán) – Italia





Dia.Pro
Diagnostic
Bio**Probes**

EC DECLARATION OF CONFORMITY

MANUFACTURER	DIA.PRO DIAGNOSTIC BIOPROBES S.R.L. VIA G. CARDUCCI N° 27 – 20099 SESTO SAN GIOVANNI (MILANO) – ITALY
PRODUCT	HSV2 IgM CODE: HSV2M.CE (96 tests)
CLASSIFICATION	GENERAL IVD
CONFORMITY ASSESSMENT ROUTE	SELF CERTIFICATION

**WE HEREBY DECLARE THAT THE ABOVE MENTIONED PRODUCT MEETS
THE PROVISIONS OF THE COUNCIL DIRECTIVE 98/79/EC
FOR IN VITRO DIAGNOSTIC DEVICES.**

ISO CERTIFICATE	UNE EN ISO 13485 N° 2013 11 0039 EN, RELEASED BY AEMPS (AGENCIA ESPAÑOLA DE MEDICAMENTOS Y PRODUCTOS SANITARIOS)
------------------------	---

PLACE & DATE OF FIRST ISSUE	MILANO – OCTOBER 2004
PLACE & DATE OF CURRENT ISSUE	SESTO SAN GIOVANNI (MI) – MARCH 2019
SIGNATURE Legal Representative Dr.ssa Fiorenza Scozzesi	

Rev: 05/2018