


DW-T8 Color Ultrasonic Diagnostic Apparatus



DW-T8 is configured with a new imaging engine, which can significantly optimize imaging performance. It is a comprehensive imaging system engineered to meet today's most demanding needs, from deep abdominal, vascular to superficial small parts.



Specifications

- Windows Embedded operation system
- 21.5" LED & free arm design
- Multi-angle adjustable LED screen meets different clinic demands
- 13.3" LED touch screen
- Brings doctors convenient operation and the time of examination is greatly shortened
- Stereo audio system
- Newly designed ergonomic console
- The fully height-adjustable control panel enables optional positioning with just one touch
- Front and rear handles
- Handles on both front and rear make T8 easy to be transported
- 4 Active transducers ports
- Improve your flexibility in moving through a wide range of applications
- 8G RAM, 120G SSD+1T HDD
- 4 Wheels with locks
- Highly maneuverable, highly mobile four-wheeled platform makes portable exams easy

Performance

- Tissue Doppler Imaging (TDI)
- Auto IMT (Intima-Media Thickness)
- Free hand Elastography
- 3D/4D Imaging
- Panoramic Imaging
- Trapezoid Imaging
- Anatomic M Mode
- Auto-Adaptive Imaging Processing

www.daweiultrasound.com

Technical Specification

1:	Summary of main specifications and system of cart type color Doppler ultrasound
1.1	Trolley type all digital color Doppler ultrasonic mainframe
1.2	Ultrasonic host operating system: Windows operating system
1.3	<p>Applications:</p> <p>Abdomen, obstetrics, gynecology, heart, urinary system, small organs, superficial, blood vessels, pediatrics, newborns, musculoskeletal</p>
1.4	Probes:

	Convex probe, Tran-vaginal probe, Linear probe, Micro-convex probe, Cardiac probe, 4D Volume probe
1.5	Applications and report: Abdominal, OB, GYN, Cardiac, Urinary, Small Parts, Superficial, Vascular, Pediatrics, Advanced measurement software packages, report software packages, case management software packages, etc.
☆1.6	carotid artery intima measurement thickness (IMT)
☆1.7	Automatic spectral envelope measurement
1.8	Full digital transmission and reception of beam synthesizer
1.9	Color Doppler imaging (C)
1.10	Pulse Doppler Imaging (PW)
☆1.12	Coherent Contrast imaging (CCI)
1.12	Continuous wave Doppler imaging (CW)
☆1.13	B/C/D Real-time three synchronous imaging
☆1.14	Power Doppler imaging (PDI)
☆1.15	Direct power Doppler imaging (DPDI)
1.16	M mode imaging
☆1.17	Anatomic M mode imaging
☆1.18	Color Doppler M mode imaging
☆1.19	Elastography



☆1.20	Tissue Doppler imaging (TDI)
☆1.21	Strain rate imaging (SRI)
1.22	Tissue harmonic imaging (THI)
1.23	Fusion harmonic imaging (FHI)
1.24	Speckle Reduce imaging (SRI)
☆1.25	Panoramic imaging
☆1.26	Deflection imaging
☆1.27	Trapezoidal imaging
1.28	Adaptive velocity optimization
☆1.29	Free hand 3D
1.30	Real time 3D imaging (3D/4D)
1.31	DICOM3.0
1.32	Monitor: ≥21.5 inch, high definition ultrasonic display
1.33	≥13.3 inch touch screen
1.34	Physical clipboard: save the image on the left side of the screen, which can be directly saved or deleted.
1.35	The system has the function of on-the-spot upgrade
1.36	Presupposition: for different inspection of the viscera, preset the inspection conditions for the best image, reduce the adjustment of the operation, and the commonly used external adjustment and combination regulation.

1.37	Probe interface: 4
1.38	Chinese and English System, Chinese and English input, optional
1.39	Depth: ≥360mm;
1.40	Extended imaging
2:	Probes
2.1 Convex probe	<p>Fundamental Frequency: 2.0MHz/2.3MHz/2.5MHz/3.0MHz/3.5MHz/4.0MHz/4.6MHz/5.0MHz/5.4MHz,</p> <p>Harmonic Frequency: 4.0MHz/4.6MHz/5.0MHz,</p>
2.2 Linear probe	<p>Fundamental Frequency: 4.0MHz/4.6MHz/5.0MHz/6.0MHz/7.0MHz/8.0MHz/9.2MHz/10.0MHz/12.0MHz/13.3MHz,</p> <p>Harmonic Frequency: 8.0MHz/9.2MHz/10.0MHz,</p>
2.3 Trans-vagi nal probe	<p>Fundamental Frequency: 3.0MHz/3.5MHz/4.0MHz/5.0MHz/5.4MHz/6.0MHz/7.0MHz/8.0MHz/10.0MHz,</p> <p>Harmonic Frequency: 6.0MHz/7.0MHz/8.0MHz,</p>
2.4 Micro-con vex probe	<p>Fundamental Frequency: 3.0MHz/3.5MHz/4.0MHz/5.0MHz/5.4MHz/6.0MHz/7.0MHz/8.0MHz,</p> <p>Harmonic Frequency:</p>

	6.0MHz/7.0MHz/8.0MHz,
2.5 Cardiac probe	<p>Fundamental Frequency: 1.7MHz/1.9MHz/2.1MHz/2.5MHz/3.0MHz/3.4MHz/3.8MHz/4.2MHz/5.0MHz,</p> <p>Harmonic Frequency: 3.4MHz/3.8MHz/4.2MHz,</p>
2.6 4D Volume probe	<p>Fundamental Frequency: 2.0MHz/2.5MHz/3.0MHz/3.3MHz/3.7MHz/4.0MHz/5.0MHz/6.0MHz,</p> <p>Harmonic Frequency: 4.0MHz/5.0MHz/6.0MHz,</p>
3:	2D imaging mode
3.1	Gain: 0 - 100, Step 2 adjustable
3.2	TGC: 8 segment adjustable
3.3	Maximum focus point: ≥ 7 , which can be moved throughout the whole process.
3.4	Speckle reduction: 0-5, 5 level
3.5	Space Synthesis: 0-2, 2 level (Liner probe: 3 level, cardiac probe:0)
3.6	Dynamic: 30-180, 35 level, step 5 adjustable
3.7	Line density: low、middle、 high, 3 level
3.8	Frame correlation: 0-4,4 level
3.9	Noise reduction: 0-5, 5 level
3.10	Edge Enhancement: 0-5, 5 level

3.11	Sound power: 2-10, 9 level
3.12	Grey scale: 0-67, 67 level
3.13	False color: 0-67, 67 level
3.14	Image style: Soft-Comparison, 2 level
	The screen has real-time display of voice power, probe frequency, dynamic range, pseudo color, gray scale and other 11 parameters can be adjusted
4:	Color Doppler imaging mode
4.1	Blood gain: 0-100, Step 2
4.2	Parameter display: Velocity、Variance
4.3	B-Restrain (B/W restrain) : 0-7, 7 level
4.4	Speed Through: 0-8, 8 level
4.5	Sampling number: 6-24, 7 level
4.6	Blood flow preferred: 0-8, 8 level
4.7	Filtering: 1-6, 6 level
4.8	Sound power: 2-6, 4 level
4.9	Noise reduction: 0-4, 4 level
4.10	Smooth treatment: 0-4, 4 level
4.11	Frame correlation: 0-6, 6 level
4.12	Chromatography (Blood flow graph) : 0-37, 37 level
4.13	Line density: Low-Middle-High, 3 level

4.14	Frequency: 4 level adjustable
	Velocity: Minimum 0.4K, Maximum 40.5K
	Convex probe: 0.4K-4.3K-38.5K
	Linear probe: 0.4K-14.7K-39.0K
4.15	Trans-vaginal probe: 0.4K-7.8K-39.7K
	Volume probe: 0.4K-4.2K-34.8K
	Micro-convex probe: 0.4K-10.3K-40.5K
	cardiac probe: 0.4K-7.8K-39.7K
	PS: The frequency of the probe changes and the frequency value changes
	PS: Frame rate changes with speed
5:	Pulse wave Doppler (PW)
1	Gain: 0-100, Step 2
5.2	Spectrum envelope function: real time automatic spectrum envelope, manual spectrum envelope, and other modes. The system automatically analyses and displays various data such as PSV, EDV, RI, PI, S/D, ACC, HR and so on. Can wake up or close
5.3	Sample volume: 0.5mm~30mm
5.4	Blood angel: -75—75 degree, Step 5
5.5	False color: 0-67, 67 level
5.6	Dynamic range: 20-40, 4 level
5.7	Filter: 0-9, 9 level



5.8	Smooth treatment: 1-4, 4 level
5.9	Sound power: 2-5, 4 level
5.10	Volume: 0-100, 10 level, Step 10
5.11	Audio filtering: 0-4, 4 level
5.12	Base line: -1.0~1.0,
5.13	Grey map: 0-67, 67 level
5.14	Scan velocity: 100-500, 6 level
5.15	<p>PRF:Minimum 0.5K, Maximum 87.5K</p> <p>Convex probe: 0.5K-4.3K-63.3K</p> <p>Linear probe: 0.5K-14.5K-78.4K</p> <p>Trans-vaginal probe: 0.5K-8.1K-78.4K</p> <p>Volume probe: 0.5K-4.2K-53.8K</p> <p>Micro-convex probe: 0.5K-10.3K-81.1K</p> <p>cardiac probe: 0.5K-4.3K-87.5K</p>
5.16	Frequency: 4 level
	PS: The frequency of the probe changes and the PRF value changes
	PS: The frequency of the probe changes and the frequency value changes
6:	Continuous Wave Doppler (CW)
6.1	Support probe: Cardiac probe
6.2	Adjustment of B mode parameters is switchable

6.3	Gain: 0-100, Step 2
6.4	Sampling line position is adjustable
6.5	PRF:0.9K~36.1K
6.6	Baseline: -1.0~1.0
6.7	Blood angel: -75~75 degree
6.8	Grey map: 0-67
6.9	Scan velocity: 100-300
6.10	False color: 0-67
6.11	Dynamic range: 20-40
6.12	Filtering: 0-9, 9 level
6.13	Smooth treatment: 1-4
6.14	Frequency: 2.0MHz/2.3MHz/2.5MHz/3.0MHz, 4 level adjustable
6.15	Sound power: 2-5
6.16	Volume: 0-100
6.17	Audio Filtering: 0-4
☆7:	Anatomical M imaging
7.1	Support probe: Convex probe, Linear probe, Cardiac probe
7.2	Adjustment of B mode parameters is switchable
7.3	Gain: 0-100, Step 2
7.4	M Sampling line angel is adjustable

7.5	M Sampling line length is adjustable
7.6	Sampling line: 3, Can be displayed or hidden separately
☆8:	Blood flow M mode (MC)
8.1	Adjustment of B mode parameters is switchable
8.2	Gain: 0-100, Step2
8.3	MC Sampling line angel is adjustable
8.4	MC Sampling line length is adjustable
8.5	Frequency: 4 level
8.6	Sampling number: 6-24
8.7	Speed through: 0-8, 8 level
8.8	Scan velocity: 150-500
8.9	Frame correlation: 0-6, 6 level
8.10	Filtering: 1-6, 6 level
8.11	Blood flow preferred: 0-8, 8 level
8.12	Smooth treatment: 0-4,4 level
8.13	Map: 0-37, 37 level
☆9:	Elastography
9.1	Adjustment of B mode parameters is switchable
9.2	Gain: 0-100, Step 2
9.3	B/E, Double real-time display on the same screen



9.4	Probe displacement curve display: Up/Down
9.5	Pressure indicator bar display
9.6	Frequency: 8-9 level, Adjustable; According to the probe display
9.7	Noise reduction: 0-2, 2 level
9.8	Frame correlation: 0-3, 3 level
9.9	Comparison: 0-13, 13 level
9.10	False color: 0-3, 3 level
9.11	Don' t support cardiac probe
☆10:	Tissue Doppler imaging (TDI)
10.1	Support probe: Cardiac probe
10.2	Adjustment of B mode parameters is switchable
10.3	Gain: 0-100, step 2
10.4	ROI area adjustable
10.5	Sampling number: 6-24
10.6	Velocity: 0.4K-8.0K
10.7	Frame correlation: 0-6, 6 level
10.8	Tissue preferred: 0-7, 7 level
10.9	Frequency: 2.0MHz/2.3MHz/2.5MHz/3.0MHz
10.10	Support color reversal
☆11:	Strain rate imaging

11.1	Support probe: Cardiac probe
11.2	Adjustment of B mode parameters is switchable
11.3	ROI area adjustable
11.4	Gain: 0-100, Step 2
11.5	Sampling number: 6-24, 6 level
11.6	Axial average: 1-4, 4 level
11.7	Velocity: 0.4K-8K
11.8	Frame correlation: 0-6, 6 level
11.9	Tissue optimization: 0-7, 7 level
☆12:	Panoramic imaging
12.1	Support probe: Linear probe
12.2	Speckle Reduction: 0-5, 5 level
☆13:	Deflection imaging
13.1	Support probe: Linear probe
13.2	Adjustment of B mode parameters is switchable
13.3	Deflection angel: 8 level
13.4	Speckle reduction: 0-5, 5 level
13.5	Dynamic rate: 30-180, Step 5
13.6	Line density: low-middle-high, 3 level
13.7	Frame Correlation: 0-4, 4 level

13.8	False color: 0-67, 67 level
13.9	Image style: Soft-Comparison, 2 level
13.10	Noise reduction: 0-5, 5 level
13.11	Edge Enhancement: 0-5, 5 level
13.12	Sound power: 2-10, 8 level
13.13	Grey map: 0-67, 67 level
☆14:	Trapezoidal imaging
14.1	Probe support: linear probe
14.2	Adjustment of B mode parameters is switchable
14.3	Deflection angel: 8 level
14.4	Speckle reduction: 0-5, 5 level
14.5	Dynamic rate: 30-180, Step 5
14.6	Line density: low-middle-high, 3 level
14.7	Frame Correlation: 0-4, 4 level
14.8	False color: 0-67, 67 level
14.9	Image style: Soft-Comparison, 2 level
14.10	Noise reduction: 0-5, 5 level
14.11	Edge Enhancement: 0-5, 5 level
14.12	Sound power: 2-10, 8 level
14.13	Grey map: 0-67, 67 level

14.14	Space Synthesis: 0-2, 2 level
☆15	Freehand 3D imaging
15.1	Support probe: convex probe, linear probe
15.2	Display model: 4 pictures
15.3	Image Rotation X/Y/Z Axis
15.4	Multi-slice Visibility
16	Real-time 4D imaging
16.1	Support probe: 4D volume probe
16.2	Adjustment of B mode parameters is switchable
16.3	Gain: 0-100, Step 2
16.4	Display model: one image、two images、four images
16.5	Image Rotation: X/Y/Z Axis
16.6	Multi-slice Visibility
16.7	Light&Shade inversion
16.8	Smooth: 0-4, 4 level
16.9	Threshold level: 0-129, Step 3
16.10	Transparency: 1-509, Step 10
16.11	Render type: 4 kinds, Surface、maximum、minimum、perspective
17:	Extended Imaging
17.1	Gain: 0 - 100, Step 2







17.2	TGC: 8 segment adjustable
17.3	Maximum focus point: ≥ 7 , which can be moved throughout the whole process.
17.4	Speckle reduction: 0-5, 5 level
17.5	Space Compound: 0-2, 2 level (Linear probe: 3 level, don't support cardiac probe)
17.6	Dynamic range: 30-180, 35 level, Step 5
17.7	Line density: Low、Middle、High, 3 level
17.8	Frame correlation: 0-4, 4 level
17.9	Noise reduction: 0-5, 5 level
17.10	Edge enhancement: 0-5, 5 level
17.11	Sound power: 2-10, 9 level
17.12	Grey map: 0-67, 67 level
17.13	False color: 0-67, 67 level
17.14	Image style: Soft-Comparison, 2 level
17.15	<p>Extended level: Maximum 72 level</p> <p>Convex probe: 9 level</p> <p>Trans-vaginal probe: 72 level</p> <p>Micro-convex probe: 29 level</p> <p>Cardiac probe: 40 level</p> <p>4D Volume probe: 17 level</p>
	PS: The screen has real-time display of voice power, probe frequency, dynamic range, pseudo color,

	gray scale and other 11 parameters can be adjusted
	PS: When the probe scan range reaches the maximum, the space synthesized is 0.
18:	Measurement and analysis function:
18.1	General measurement: Distance, area, ellipse, cross line, angle, distance ratio, volume, Volume (ellipse), area ratio, diameter, joint angle
18.2	Cardiac: Automatic spectrum envelope, LV, Main Pulmonary artery diameter, RVEDd, RVEDs, LVM, LAV, HR, MVF, AO, AR, LVOT, TVF, Pulmonic valve, Pulmonary vein, RV, Doppler fetal heart sound, LVET, LVM, LVMI, AV
18.3	Vascular: carotid intima (IMT), length stenosis ratio, area stenosis ratio, IMT (back wall), IMT (front wall)
18.4	OB: Fetal routine, AFI, TW, GS, CRI, OFD, HL, ulna, NT, Fibula, Nbonel, Radial, Tibia
18.5	GYN: uterus, cervix, corpus uteri/cervix uterus, left ovarian vein, right ovarian vein, dominant follicle, intima thickness
18.6	Urology: prostate, residual urine, left kidney, right kidney, left suprarenal vein, right suprarenal, left testis, right testis, left seminal vesicle, right seminal vesicle
18.7	Abdomen: liver, CHD, portal vein diameter, cholecyst, CBD, pancreas, spleen, Internal diameter of abdominal aorta, kidney
18.8	Small parts: Thyroid
18.9	Software package: Measurement package, Software package, Medical records management software package

19:	Graphic and text management system
19.1	Host build in 2 hard disk (SSD 120+1T), Start fast and stable, RAM: 8GB
19.2	Movie playback: ≥ 1200 frames
19.3	Internal file information management system: can record patient number, name, check number, check date and so on, and can be searched and managed by numbering, checking number, name and so on.
19.4	Type of report is 16
19.5	One key fast report graphic and text management
20:	Interface
20.1	USB interface: 4
20.2	HDMI interface: 1
20.3	RJ-45 interface: 1
20.4	Grounding wire interface: 1
20.5	DVD RW: 1
21:	Configuration
21.1	Trolley type full digital color Doppler ultrasound diagnostic system
21.2	Probe: convex array probe (standard), linear probe (optional), Trans-vaginal probe (optional), cardiac probe (optional), 4D volume probe (optional)
21.3	≥ 13 quick adjusting knobs
22:	Technology, after-sales service and other requirements
22.1	After acceptance, the warranty is free for two years (Provide manufacturer warranty certificate)

22.2	Manufacturer has ISO13485 certification and EU CE certification.

Image Gallery



**COLOR DOPPLER
ULTRASOUND SERIES**

Specially For *Cardiology, Gyn&Ob, Musculoskeletal, etc.*



Support *4D/3D, CW/PW/TDI, Elastography/Panoramic, Auto IMT/M, etc.*

**BLACK & WHITE
ULTRASOUND SERIES**

Based on *Windows system, 128 elements, Printer connecting directly, etc.*



Rich measurement software package, with *PW* function, support *User Defined, etc.*

**USB PROBE
ULTRASOUND SERIES**




Support *Mobile phone, Pad, Computer*. Support both *Color Doppler and B/W*.
It is convenient to check *whenever and wherever possible.*



THE NEXT CHAPTER, COMING SOON ...



ZERTIFIKAT CERTIFICATE CERTIFICADO CERTIFICAT

EC Certificate
Full Quality Assurance System
Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
(Directive in Cases (a), (b) or (c))
No. G1 089075 0007 Rev. 00

Manufacturer: Dawei Medical (Jiangsu) Co., Ltd.
21 Jinqiao Road
Economic and Technological Development Zone
221004 Xuzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Facility(ies): Dawei Medical (Jiangsu) Co., Ltd.
21 Jinqiao Road, Economic and Technological Development
Zone, 221004 Xuzhou, Jiangsu, PEOPLE'S REPUBLIC OF
CHINA

Product Category(ies): Ultrasonic Diagnostic System,
Color Ultrasonic Diagnostic Apparatus


The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodic surveillance. For marketing of class II devices an additional Annex II (4) certificate is mandatory. See also notes overleaf.

Report No.: SH19878EXT01
Valid from: 2019-07-25
Valid until: 2024-06-26

Date: 2019-07-25

S. Punnit
Saratn Punnit
Head of Certification/Notified Body

Page 1 of 1
TÜV SÜD Product Service GmbH is Notified Body with Identification no. 0123
TÜV SÜD Product Service GmbH - Certification Body - Rosenstraße 65 - 80352 Munich - Germany



SGS No: QIP-AS1 1241265
Made-in-China.com
Connecting Buyers with China Suppliers

Section 2: Photos of Company

Description: Company Gate	Description: Office
Description: Workshop	Description: Workshop
Description: Products Sample	Description: Products Sample

Page No.: 11 of 12

Ada Lau

Regional Sales Manager

International Trade Dept | Dawei Medical (Jiangsu) Co., LTD

For Love, Image the world.

+86 18052251703

ada@dwultrasound.com

www.dwultrasound.com

28 Jinqiao Road Economic and Technological Development Zone, Xuzhou, Jiangsu, China