



America


CERTIFICATE

No. U8 058067 0090 Rev. 00

Holder of Certificate: **Hamilton Medical AG**

Via Crusch 8
7402 Bonaduz
SWITZERLAND

Certification Mark:



Product: **Breathing therapy
Lung Ventilator**

Model(s): **HAMILTON-C1**

Parameters:	Rated Input voltage:	100-240 VAC
	Rated frequency:	50/60 Hz
	Rated Input Current:	150 VA
	Protection class:	II or internally powered
	IP Protection:	IP22
	Applied Part Type:	B
	SpO ₂ , CO ₂	BF

Remarks:
The certificate is valid for USA and Canada.
Other countries might have other requirements,
which were not part of this certification.

Tested according to: CAN/CSA-C22.2 No. 60601-1:2014
ANSI/AAMI ES60601-1:2005/(R)2012

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713196652

Date, 2021-02-25

(Stefan Reis)