



## Certificate of Conformity

No.: PD/0357/2024

Applicant: Shenzhen Gangzhu Medical Technology Co., Ltd.  
Address: 12th Floor, North Zone of Building 2#B, SkyworthInnovation Valley, No.8 Tangtou No.1 Road, ShiyanStreet, Bao'an District, Shenzhen, China  
Manufacturer: Shenzhen Gangzhu Medical Technology Co., Ltd.  
Address: 12th Floor, North Zone of Building 2#B, SkyworthInnovation Valley, No.8 Tangtou No.1 Road, ShiyanStreet, Bao'an District, Shenzhen, China  
Product: Real-time Fluorescence PCR Instrument  
Type / Models: GZ-8 Plus  
Tested Standard(s): (1) EN 61010-1:2010  
EN 61010-1:2010/A1:2019/AC:2019-04  
EN 61010-1:2010/A1:2019  
(2) EN 61326-1:2013  
(3) EN IEC 61326-2-6:2021

The product meets the requirements of the above-mentioned standards, which are harmonized with Low Voltage Directive 2014/35/EU (1) and Electromagnetic Compatibility Directive 2014/30/EU (2).

The assessment of conformity was based on:

- evaluation of the technical documentation of the product as mentioned above,
- results contained in test report no: 20230318G02429X-Y, 20230318G02429X-E

This certificate was issued first time at 22/04/2024 This certificate remains valid until **21/04/2029**.

This certificate applies only to product samples with the same characteristics and complying with the above-mentioned requirements as the samples submitted for assessment.

This certificate has been issued in accordance with voluntary certification programme DPCW-01 (type 1a by EN ISO/IEC 17067).

Chief Executive Officer

Ben Xu



Warsaw, 22/04/2024

