
 <b>Agilent</b>	<b>DECLARATION OF CONFORMITY</b> According to EN ISO/IEC 17050-1 :2010	
--	---	---

**Manufacturer's Name:** Agilent Technologies, Inc  
**Manufacturer's Address:** 2850 Centerville Road  
 Wilmington, Delaware 19808-1644 USA

**Declares under sole responsibility that the product as originally delivered**

**Product Name:** 5977 Series Mass Spectroscopy System  
**Product Number:** 5977x Series Mass Spectroscopy System (where x = any letter other than A)  
**Product Options:** All options, accessories and customized products based on the above listed models

**complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:**

Low Voltage Directive	2014/35/EU
EMC Directive	2014/30/EU
RoHS Directive	2011/65/EU as amended by (EU) 2015/863

**and conforms with the following product standards:**

<b>EMC:</b>	<b>Standard</b>	EN 61326-1:2013 EN 55011:2009+A1:2010
	IEC 61326-1:2012 CISPR 11:2009+A1:2010	
	Canada: ICES-001 Issue 4 Australia/New Zealand: IEC 61326-1:2012	
<b>Safety:</b>	IEC/EN 61010-1:2010 IEC/EN 61010-2-010:2014	IEC 61010-2-081:2015
	Canada: CSA C22.2 No. 61010-1-12	CAN/CSA-C22.2 No. 61010-2-081:15
	USA: ANSI/UL 61010-1 Third Edition	UL 61010-2-081 Ed. 2. (2015)
<b>Other:</b>	EN IEC 63000:2018	

**Supplementary Information:**

The product was tested in a typical configuration with Agilent Technologies test systems.

**This DoC applies to above-listed product placed on the EU market after:**

28 May 2021

Date



Kevin Draper  
Quality Manager

For further information, please contact your local Agilent Technologies sales office, agent or distributor, or Agilent Technologies Deutschland GmbH, Hewlett-Packard-Strasse 8, 76337 Waldbronn, Germany.