



Bio-Rad
Laboratories

Diagnostics Group
9500 Jeronimo Road
Irvine, California 92618-2017
Telephone: (949) 598-1200

CERTIFICATE-OF-ANALYSIS

Product Name: Lyphochek® Tumor Marker Plus Control

Master Batch Number: 54680

Product Base: Human Serum

Physical Form: Lyophilized

Date of Manufacture: 2019-05

Description	Material Number	Batch Number	Expiration Date
Level 1	367	54681	2022-08-31
Level 2	368	54682	2022-08-31
Level 3	369	54683	2022-08-31
Trilevel MiniPak	368X	54680	2022-08-31

Testing Results:

Each human donor unit used to manufacture this product was tested and found non-reactive at the donor level per current applicable FDA requirements using FDA-accepted methods including testing for HIV-1/HIV-2 antibody, HBsAg, HCV Antibody, HIV-1 and HCV by a nucleic acid test (NAT).

The bovine source material(s) of this product is collected in USDA licensed establishments. These animals received ante and post mortem inspections at the abattoir by a US veterinary service inspector.

Water Activity: Specification: <0.200 aw	Level 1 0.050 aw	Level 2 0.049 aw	Level 3 0.050 aw
--	---------------------	---------------------	---------------------

Homogeneity Claim:


"Testing has been conducted to verify sufficient homogeneity in accordance with established requirements."

ADDITIONAL TESTING:


This product has been manufactured under applicable guidelines/standards and meets all established Bio-Rad quality requirements.

Approved By:

Date:



 (QC Supervisor/Manager)



 (QA Supervisor/Manager)

