

S/N: 001211

No.: **ICR Polska/VC/HZ201105**

**Name and address of Applicant** Shandong Chengwu Sainuo Medical Devices Co., Ltd.  
Beicheng Development Zone, North Section of Deshang Road, Chengwu County (west side, 1000 meters to the North of Paper Mill), Heze City, Shandong Province, China

**Name and address of manufacturer:** Shandong Chengwu Sainuo Medical Devices Co., Ltd.  
Beicheng Development Zone, North Section of Deshang Road, Chengwu County (west side, 1000 meters to the North of Paper Mill), Heze City, Shandong Province, China

**Product name:** Scalp Vein Set

**Product types:** Butterfly Needle, 0.4(27G), 0.45(26G), 0.5(25G), 0.55(24G), 0.6(23G), 0.7(22G), 0.8(21G), 0.9(20G), 1.1(19G), 1.2(18G)

**Product trademark:** 宽荣

This document confirms that the product sample meets the requirements of the following standards:

- Related with Directive 93/42/EEC:  
EN ISO 10993-1:2009+AC:2010  
EN ISO 15223-1:2016  
EN 1707:1997

The assessment process has been carried out in accordance with individual rules and conditions agreed with the applicant. Evaluation has been carried out in accordance with:

**Test report:** STA-(20)-A-1123317-MDD  
**Test conducted by:** STA International (UK) Testing Ltd.

**Issue date:** 24.11.2020  
**Expiration date:** 23.11.2025

#### Remarks

This Verification of conformity refers to the above mentioned product and its conformity in regards of above mentioned standard(s) was proven on test sample.

This Verification of conformity was issued on voluntary basis and does not imply meeting all essential requirements, assessment of the series-production or any other restricted Notified Bodies conformity assessment procedure appropriate for the product.



marking remarks:  
mark is not sanctioned by the following verification of conformity mark given here as reference, can be only use by the manufacturer after applying all essential requirements from relevant directives

document status can be checked: <https://cert.icrpolska.com/>

Director: Rafal Kalinowski

Warsaw, 24.11.2020

